

LA121345


SFAM01.1

INSTRUMENT ID: CV2

Linear

μ Abs.:

39449



A= 0.0000e+000

B= 2.5450e-004 *slope*

C= -1.5077e-001 *y-intercept*

Rho= 0.9997691

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.095	-0.095	218	0.000	218					=
0.05	0.050					185	0				17
0.20	0.200	0.234	0.034	1510	0.0 %	1510					0
2.50	2.500	2.499	-0.001	10413	0.0 %	10413					2
5.00	5.000	5.093	0.093	20606	0.0 %	20606					1
7.50	7.500	7.580	0.080	30378	0.0 %	30378					1
10.0	10.000	9.889	-0.111	39449	0.0 %	39449					-1

LB121345

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	218	-		0 SFAM01.1	PPB	8/3/2022 18:28	Std
	0.2 S01	1510	-		0.2 SFAM01.1	PPB	8/3/2022 18:31	Std
	2.5 S02	10413	-		2.5 SFAM01.1	PPB	8/3/2022 18:33	Std
	5 S03	20606	-		5 SFAM01.1	PPB	8/3/2022 18:35	Std
	7.5 S04	30378	-		7.5 SFAM01.1	PPB	8/3/2022 18:40	Std
	10 S05	39449	-		10 SFAM01.1	PPB	8/3/2022 18:47	Std
ICV065	ICV065	17042	4.1863 -		SFAM01.1	PPB	8/3/2022 18:50	SMPL
ICB065	ICB065	242	-0.0892 -		SFAM01.1	PPB	8/3/2022 18:55	SMPL
CCV046	CCV046	20706	5.1188 -		SFAM01.1	PPB	8/3/2022 18:57	SMPL
CCB046	CCB046	247	-0.0879 -		SFAM01.1	PPB	8/3/2022 18:59	SMPL
PB146693BL	PBW693	207	-0.0981 -		SFAM01.1	PPB	8/3/2022 19:04	SMPL
PB146693TB	LEB693	269	-0.0823 -		SFAM01.1	PPB	8/3/2022 19:09	SMPL
N3946-01	MH1A01	495	-0.0248 -		SFAM01.1	PPB	8/3/2022 19:11	SMPL
N3946-03	MH1A01D	273	-0.0813 -		SFAM01.1	PPB	8/3/2022 19:13	SMPL
N3946-04	MH1A01S	4414	0.9726 -		SFAM01.1	PPB	8/3/2022 19:15	SMPL
CCV047	CCV047	20467	5.058 -		SFAM01.1	PPB	8/3/2022 19:18	SMPL
CCB047	CCB047	237	-0.0905 -		SFAM01.1	PPB	8/3/2022 19:20	SMPL