

LB121351


SFAM01.1

INSTRUMENT ID: CV2

Linear

μ Abs.:

41375



A= 0.0000e+000

B= 2.4903e-004 *slope*

C= -1.5695e-001 *y-intercept*

Rho= 0.9990297

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.067	-0.067	362	0.000	362					1
0.05	0.050					185	0				2
0.20	0.200	0.241	0.041	1597	0.0 %	1597					3
2.50	2.500	2.583	0.083	11003	0.0 %	11003					3
5.00	5.000	5.125	0.125	21211	0.0 %	21211					-4
7.50	7.500	7.171	-0.329	29426	0.0 %	29426					1
10.0	10.000	10.147	0.147	41375	0.0 %	41375					

LB121351

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	362	-		0 SFAM01.1	PPB	8/4/2022 10:18	Std
	0.2 S01	1597	-		0.2 SFAM01.1	PPB	8/4/2022 10:20	Std
	2.5 S02	11003	-		2.5 SFAM01.1	PPB	8/4/2022 10:25	Std
	5 S03	21211	-		5 SFAM01.1	PPB	8/4/2022 10:30	Std
	7.5 S04	29426	-		7.5 SFAM01.1	PPB	8/4/2022 10:32	Std
	10 S05	41375	-		10 SFAM01.1	PPB	8/4/2022 10:42	Std
ICV066	ICV066	16010	3.83 -		SFAM01.1	PPB	8/4/2022 10:44	SMPL
ICB066	ICB066	129	-0.1248 -		SFAM01.1	PPB	8/4/2022 10:47	SMPL
CCV048	CCV048	21042	5.0831 -		SFAM01.1	PPB	8/4/2022 10:49	SMPL
CCB048	CCB048	132	-0.1241 -		SFAM01.1	PPB	8/4/2022 10:51	SMPL
PB146694BL	PBW694	95	-0.1333 -		SFAM01.1	PPB	8/4/2022 10:56	SMPL
PB146683TB	LEB683	308	-0.0803 -		SFAM01.1	PPB	8/4/2022 10:58	SMPL
N3954-01	MH1A02	180	-0.1121 -		SFAM01.1	PPB	8/4/2022 11:01	SMPL
N3954-02	MH1A02D	162	-0.1166 -		SFAM01.1	PPB	8/4/2022 11:03	SMPL
N3954-03	MH1A02S	4200	0.889 -		SFAM01.1	PPB	8/4/2022 11:05	SMPL
CCV049	CCV049	21413	5.1755 -		SFAM01.1	PPB	8/4/2022 11:07	SMPL
CCB049	CCB049	148	-0.1201 -		SFAM01.1	PPB	8/4/2022 11:10	SMPL