

LB121358


245.1

INSTRUMENT ID: CV4

Linear

 μ Abs.:

34240



A= 0.0000e+000

B= 2.9194e-004

C= -2.7716e-002

Rho= 0.9999560

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/o
0.0	0.000	-0.006	-0.006	76	0.000	76					
0.05	0.050	0.056	0.006	286	0.0 %	286					12
0.2	0.200	0.180	-0.020	713	0.0 %	713	0				-10
2.5	2.500	2.537	0.037	8785	0.0 %	8785					1
5.0	5.000	4.954	-0.046	17065	0.0 %	17065					-1
7.5	7.500	7.560	0.060	25990	0.0 %	25990					1
10.0	10.000	9.968	-0.032	34240	0.0 %	34240					0

LB121353 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 04 Aug 2022 12:30:01

Sample ID	Extended ID	μ Abs	Conc	Strd Conc	Method	Units	Date	Type	Type
0.0 - 1	50	76	-	0.0000	245.1	PPB	04 Aug 2022 13:01:07	S	Std
0.05 - 1	50.05	286	-	0.0500	245.1	PPB	04 Aug 2022 13:06:07	S	Std
0.2 - 1	50.2	713	-	0.2000	245.1	PPB	04 Aug 2022 13:08:24	S	Std
2.5 - 1	52.5	8785	-	2.5000	245.1	PPB	04 Aug 2022 13:10:41	S	Std
5.0 - 1	55	17065	-	5.0000	245.1	PPB	04 Aug 2022 13:12:57	S	Std
7.5 - 1	57.5	25990	-	7.5000	245.1	PPB	04 Aug 2022 13:15:13	S	Std
10.0 - 1	51.8	34240	-	10.0000	245.1	PPB	04 Aug 2022 13:20:04	S	Std
ICV67 - 1	ICV67	14080	4.0828	-245.1		PPB	04 Aug 2022 13:25:19	U	SMPL
ICB67 - 1	ICB67	4	-0.0265	-245.1		PPB	04 Aug 2022 13:27:34	U	SMPL
CCV50 - 1	CCV50	17241	5.0056	-245.1		PPB	04 Aug 2022 13:29:50	U	SMPL
CCB50 - 1	CCB50	8	-0.0301	-245.1		PPB	04 Aug 2022 13:32:05	U	SMPL
CRA - 1	CRA	289	0.0567	-245.1		PPB	04 Aug 2022 13:34:22	U	SMPL
HighStd - 1	HighStd	33861	9.8577	-245.1		PPB	04 Aug 2022 13:36:37	U	SMPL
ChkStd - 1	ChkStd	24472	7.1166	-245.1		PPB	04 Aug 2022 13:38:53	U	SMPL
PB146757BL - 1	PBW	-7	-0.0298	-245.1		PPB	04 Aug 2022 13:41:10	U	SMPL
PB146757BS - 1	LCSW	12670	3.6712	-245.1		PPB	04 Aug 2022 13:45:43	U	SMPL
N3969-01 - 1	CHAMBERS-STREET-A3-9065X-MH	586	0.1434	-245.1		PPB	04 Aug 2022 13:47:59	U	SMPL
N3969-01DUP - 1	CHAMBERS-STREET-A3-9065X-MHDUP	597	0.1466	-245.1		PPB	04 Aug 2022 13:50:16	U	SMPL
N3969-01MS - 1	CHAMBERS-STREET-A3-9065X-MHMS	15294	4.4372	-245.1		PPB	04 Aug 2022 13:54:39	U	SMPL
N3969-01MSD - 1	CHAMBERS-STREET-A3-9065X-MHMSD	14206	4.1196	-245.1		PPB	04 Aug 2022 13:56:55	U	SMPL
CCV51 - 1	CCV51	16934	4.9160	-245.1		PPB	04 Aug 2022 13:59:11	U	SMPL
CCB51 - 1	CCB51	7	-0.0257	-245.1		PPB	04 Aug 2022 14:01:27	U	SMPL
N3969-01LX5 - 1		184	0.0260	-245.1		PPB	04 Aug 2022 14:03:43	U	SMPL
N3969-01A - 1		14291	4.1444	-245.1		PPB	04 Aug 2022 14:05:58	U	SMPL
CCV52 - 1	CCV52	17225	5.0009	-245.1		PPB	04 Aug 2022 14:08:14	U	SMPL
CCB52 - 1	CCB52	5	-0.0263	-245.1		PPB	04 Aug 2022 14:10:30	U	SMPL