

LB121356


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

41955



A= 0.0000e+000

B= 2.3592e-004

C= -2.5766e-003

Rho= 0.9997916

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.001	0.001	15	0.000	15					-
0.2	0.200	0.199	-0.001	854	0.0 %	854					-1
2.5	2.500	2.430	-0.070	10312	0.0 %	10312					-3
5.0	5.000	5.049	0.049	21412	0.0 %	21412					1
7.5	7.500	7.626	0.126	32335	0.0 %	32335					2
10.0	10.000	9.895	-0.105	41955	0.0 %	41955					-1

LB121356 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 04 Aug 2022 14:57:52

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	15	-	0.0000	7471B	PPB	04 Aug 2022 15:32:44	S	Std
0.2 - 1	S0.2	854	-	0.2000	7471B	PPB	04 Aug 2022 15:35:01	S	Std
2.5 - 1	S2.5	10312	-	2.5000	7471B	PPB	04 Aug 2022 15:37:17	S	Std
5.0 - 1	S5	21412	-	5.0000	7471B	PPB	04 Aug 2022 15:39:34	S	Std
7.5 - 1	S7.5	32335	-	7.5000	7471B	PPB	04 Aug 2022 15:47:17	S	Std
10.0 - 1	S10	41955	-	10.0000	7471B	PPB	04 Aug 2022 15:49:32	S	Std
ICV68 - 1	ICV68	18103	4.2682	-	7471B	PPB	04 Aug 2022 15:54:51	U	SMPL
ICB68 - 1	ICB68	3	-0.0019	-	7471B	PPB	04 Aug 2022 15:59:36	U	SMPL
CCV53 - 1	CCV53	22676	5.3471	-	7471B	PPB	04 Aug 2022 16:01:51	U	SMPL
CCB53 - 1	CCB53	-85	-0.0226	-	7471B	PPB	04 Aug 2022 16:04:06	U	SMPL
CRA - 1	CRA	852	0.1984	-	7471B	PPB	04 Aug 2022 16:06:23	U	SMPL
HighStd - 1	HighStd	43517	10.2638	-	7471B	PPB	04 Aug 2022 16:11:15	U	SMPL
ChkStd - 1	ChkStd	30152	7.1108	-	7471B	PPB	04 Aug 2022 16:13:30	U	SMPL
PB146759BL - 1	PBS	375	0.0859	-	7471B	PPB	04 Aug 2022 16:15:48	U	SMPL
PB146759BS - 1	LCSS	17762	4.1878	-	7471B	PPB	04 Aug 2022 16:23:01	U	SMPL
N3960-01 - 1	RVE-149	2725	0.6403	-	7471B	PPB	04 Aug 2022 16:29:55	U	SMPL
N3960-01DUP - 1	RVE-149DUP	2675	0.6285	-	7471B	PPB	04 Aug 2022 16:32:11	U	SMPL
N3960-01MS - 1	RVE-149MS	19736	4.6535	-	7471B	PPB	04 Aug 2022 16:39:24	U	SMPL
N3960-01MSD - 1	RVE-149MSD	18516	4.3657	-	7471B	PPB	04 Aug 2022 16:41:41	U	SMPL
N3960-02 - 1	RVE-150	2284	0.5363	-	7471B	PPB	04 Aug 2022 16:43:58	U	SMPL
CCV54 - 1	CCV54	22420	5.2867	-	7471B	PPB	04 Aug 2022 16:46:14	U	SMPL
CCB54 - 1	CCB54	-84	-0.0224	-	7471B	PPB	04 Aug 2022 16:48:32	U	SMPL
N3960-03 - 1	RVE-151	2884	0.6778	-	7471B	PPB	04 Aug 2022 16:50:49	U	SMPL
N3960-04 - 1	RVE-152	1144	0.2673	-	7471B	PPB	04 Aug 2022 16:53:05	U	SMPL
N3960-05 - 1	RVE-153	1869	0.4384	-	7471B	PPB	04 Aug 2022 16:55:20	U	SMPL
N3960-06 - 1	RVE-154	1887	0.4426	-	7471B	PPB	04 Aug 2022 16:57:36	U	SMPL
N3960-07 - 1	RVE-155	2710	0.6368	-	7471B	PPB	04 Aug 2022 16:59:52	U	SMPL
N3960-08 - 1	RVE-156	4394	1.0340	-	7471B	PPB	04 Aug 2022 17:02:08	U	SMPL
N3960-09 - 1	RVE-157	2881	0.6771	-	7471B	PPB	04 Aug 2022 17:04:24	U	SMPL
N3960-10 - 1	RVE-158	2007	0.4709	-	7471B	PPB	04 Aug 2022 17:06:41	U	SMPL
N3960-11 - 1	RVE-159	2746	0.6452	-	7471B	PPB	04 Aug 2022 17:08:58	U	SMPL
N3960-12 - 1	RVE-159D	3657	0.8602	-	7471B	PPB	04 Aug 2022 17:11:14	U	SMPL
CCV55 - 1	CCV55	22737	5.3615	-	7471B	PPB	04 Aug 2022 17:19:39	U	SMPL
CCB55 - 1	CCB55	-65	-0.0179	-	7471B	PPB	04 Aug 2022 17:21:55	U	SMPL
N3960-13 - 1	RVE-160	2757	0.6478	-	7471B	PPB	04 Aug 2022 17:24:12	U	SMPL
N3960-14 - 1	RVE-161	2487	0.5841	-	7471B	PPB	04 Aug 2022 17:26:27	U	SMPL
N3960-15 - 1	RVE-162	3581	0.8422	-	7471B	PPB	04 Aug 2022 17:28:43	U	SMPL
N3960-16 - 1	RVE-163	2214	0.5197	-	7471B	PPB	04 Aug 2022 17:30:58	U	SMPL
N3960-17 - 1	RVE-164	2256	0.5297	-	7471B	PPB	04 Aug 2022 17:33:14	U	SMPL
N3960-18 - 1	RVE-165	682	0.1583	-	7471B	PPB	04 Aug 2022 17:35:31	U	SMPL
N3960-19 - 1	RVE-166	4177	0.9828	-	7471B	PPB	04 Aug 2022 17:37:47	U	SMPL
N3960-20 - 1	RVE-167	2963	0.6964	-	7471B	PPB	04 Aug 2022 17:40:04	U	SMPL
PB146760BL - 1	PBS	-2	-0.0030	-	7471B	PPB	04 Aug 2022 17:42:22	U	SMPL
PB146760BS - 1	LCSS	17517	4.1300	-	7471B	PPB	04 Aug 2022 17:49:23	U	SMPL
CCV56 - 1	CCV56	23220	5.4754	-	7471B	PPB	04 Aug 2022 17:53:52	U	SMPL
CCB56 - 1	CCB56	-63	-0.0174	-	7471B	PPB	04 Aug 2022 17:56:07	U	SMPL
N3961-01 - 1	RVE-168	3482	0.8189	-	7471B	PPB	04 Aug 2022 18:03:22	U	SMPL
N3961-01DUP - 1	RVE-168DUP	3444	0.8099	-	7471B	PPB	04 Aug 2022 18:05:37	U	SMPL
N3961-01MS - 1	RVE-168MS	21188	4.9960	-	7471B	PPB	04 Aug 2022 18:10:07	U	SMPL
N3961-01MSD - 1	RVE-168MSD	21030	4.9587	-	7471B	PPB	04 Aug 2022 18:12:22	U	SMPL
N3961-02 - 1	RVE-169	2191	0.5143	-	7471B	PPB	04 Aug 2022 18:14:39	U	SMPL
N3961-03 - 1	RVE-170	2290	0.5377	-	7471B	PPB	04 Aug 2022 18:16:56	U	SMPL
N3961-04 - 1	RVE-171	8897	2.0964	-	7471B	PPB	04 Aug 2022 18:19:12	U	SMPL
N3961-05 - 1	RVE-172	2279	0.5351	-	7471B	PPB	04 Aug 2022 18:21:29	U	SMPL
N3961-06 - 1	RVE-173	2489	0.5846	-	7471B	PPB	04 Aug 2022 18:23:46	U	SMPL
N3961-07 - 1	RVE-174	3004	0.7061	-	7471B	PPB	04 Aug 2022 18:26:02	U	SMPL
CCV57 - 1	CCV57	23083	5.4431	-	7471B	PPB	04 Aug 2022 18:30:45	U	SMPL
CCB57 - 1	CCB57	-54	-0.0153	-	7471B	PPB	04 Aug 2022 18:33:00	U	SMPL
N3961-08 - 1	RVE-175	2979	0.7002	-	7471B	PPB	04 Aug 2022 18:35:17	U	SMPL
N3961-09 - 1	RVE-176	2688	0.6316	-	7471B	PPB	04 Aug 2022 18:37:32	U	SMPL
N3961-10 - 1	RVE-177	771	0.1793	-	7471B	PPB	04 Aug 2022 18:39:48	U	SMPL
N3961-11 - 1	RVE-178	3551	0.8352	-	7471B	PPB	04 Aug 2022 18:42:04	U	SMPL
N3961-12 - 1	RVE-179	2312	0.5429	-	7471B	PPB	04 Aug 2022 18:44:19	U	SMPL
N3961-13 - 1	RVE-179D	2360	0.5542	-	7471B	PPB	04 Aug 2022 18:46:35	U	SMPL
N3961-14 - 1	RVE-180	2589	0.6082	-	7471B	PPB	04 Aug 2022 18:48:52	U	SMPL
N3961-15 - 1	RVE-181	5545	1.3056	-	7471B	PPB	04 Aug 2022 18:51:09	U	SMPL
N3961-16 - 1	RVE-182	3614	0.8500	-	7471B	PPB	04 Aug 2022 18:53:26	U	SMPL
N3961-17 - 1	RVE-183	2808	0.6599	-	7471B	PPB	04 Aug 2022 18:55:41	U	SMPL
CCV58 - 1	CCV58	21575	5.0873	-	7471B	PPB	04 Aug 2022 18:57:58	U	SMPL
CCB58 - 1	CCB58	-59	-0.0165	-	7471B	PPB	04 Aug 2022 19:00:14	U	SMPL
N3961-18 - 1	RVE-184	5369	1.2641	-	7471B	PPB	04 Aug 2022 19:02:30	U	SMPL
N3961-19 - 1	RVE-185	2343	0.5502	-	7471B	PPB	04 Aug 2022 19:04:46	U	SMPL
N3961-20 - 1	RVE-186	3443	0.8097	-	7471B	PPB	04 Aug 2022 19:07:03	U	SMPL

LB121356 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 04 Aug 2022 14:57:52

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
N3960-01LX5 - 1		684	0.1588		-7471B	PPB	04 Aug 2022 19:09:19	U	SMPL
N3960-01A - 1		21580	5.0885		-7471B	PPB	04 Aug 2022 19:14:11	U	SMPL
N3961-01LX5 - 1		840	0.1956		-7471B	PPB	04 Aug 2022 19:20:52	U	SMPL
N3961-01A - 1		21156	4.9885		-7471B	PPB	04 Aug 2022 19:23:08	U	SMPL
CCV59 - 1	CCV59	20464	4.8252		-7471B	PPB	04 Aug 2022 19:28:01	U	SMPL
CCB59 - 1	CCB59	-37	-0.0113		-7471B	PPB	04 Aug 2022 19:30:18	U	SMPL