

LB121369


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

47532



A= 0.0000e+000

B= 2.1106e-004

C= 5.1868e-002

Rho= 0.9998710

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.054	0.054	8	0.000	8					-
0.2	0.200	0.206	0.006	730	0.0 %	730					3
2.5	2.500	2.502	0.002	11609	0.0 %	11609					0
5.0	5.000	4.910	-0.090	23019	0.0 %	23019					-2
7.5	7.500	7.445	-0.055	35027	0.0 %	35027					-1
10.0	10.000	10.084	0.084	47532	0.0 %	47532					1

LB121369 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 05 Aug 2022 12:06:55				
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0.0 - 1	50	8	-	0.0000	7470A	PPB	05 Aug 2022 12:36:00	S Std
0.2 - 1	50.2	730	-	0.2000	7470A	PPB	05 Aug 2022 12:38:17	S Std
2.5 - 1	52.5	11609	-	2.5000	7470A	PPB	05 Aug 2022 12:40:34	S Std
5.0 - 1	55	23019	-	5.0000	7470A	PPB	05 Aug 2022 12:47:50	S Std
7.5 - 1	57.5	35027	-	7.5000	7470A	PPB	05 Aug 2022 12:50:07	S Std
10.0 - 1	511	47532	-	10.0000	7470A	PPB	05 Aug 2022 12:52:23	S Std
ICV69 - 1	ICV69	18606	3.9788	-	7470A	PPB	05 Aug 2022 12:55:05	U SMPL
ICB69 - 1	ICB69	-41	0.0432	-	7470A	PPB	05 Aug 2022 12:57:20	U SMPL
CCV60 - 1	CCV60	24014	5.1202	-	7470A	PPB	05 Aug 2022 12:59:35	U SMPL
CCB60 - 1	CCB60	-56	0.0400	-	7470A	PPB	05 Aug 2022 13:01:50	U SMPL
CRA - 1	CRA	730	0.2059	-	7470A	PPB	05 Aug 2022 13:04:06	U SMPL
HighStd - 1	HighStd	47726	10.1247	-	7470A	PPB	05 Aug 2022 13:06:20	U SMPL
ChkStd - 1	ChkStd	33868	7.1999	-	7470A	PPB	05 Aug 2022 13:08:36	U SMPL
PB146729BL - 1	PBW	-77	0.0356	-	7470A	PPB	05 Aug 2022 13:10:53	U SMPL
PB146729BS - 1	LCSW	18842	4.0286	-	7470A	PPB	05 Aug 2022 13:13:10	U SMPL
N4016-03 - 1	MH-DS	167	0.0871	-	7470A	PPB	05 Aug 2022 13:25:10	U SMPL
N4016-06 - 1	MH-ES	242	0.1029	-	7470A	PPB	05 Aug 2022 13:27:26	U SMPL
N4016-09 - 1	MH-ED	136	0.0806	-	7470A	PPB	05 Aug 2022 13:29:42	U SMPL
N4016-09DUP - 1	MH-EDDUP	112	0.0755	-	7470A	PPB	05 Aug 2022 13:31:59	U SMPL
N4016-09MS - 1	MH-EDMS	18499	3.9562	-	7470A	PPB	05 Aug 2022 13:34:15	U SMPL
CCV61 - 1	CCV61	23593	5.0313	-	7470A	PPB	05 Aug 2022 13:36:30	U SMPL
CCB61 - 1	CCB61	-39	0.0436	-	7470A	PPB	05 Aug 2022 13:38:45	U SMPL
N4016-09MSD - 1	MH-EDMSD	18649	3.9879	-	7470A	PPB	05 Aug 2022 13:41:00	U SMPL
PB146730BL - 1	PBW	-41	0.0432	-	7470A	PPB	05 Aug 2022 13:43:16	U SMPL
PB146730BS - 1	LCSW	18832	4.0265	-	7470A	PPB	05 Aug 2022 13:45:31	U SMPL
N3997-15 - 1	1986	286	0.1122	-	7470A	PPB	05 Aug 2022 13:50:18	U SMPL
N3997-15DUP - 1	1986DUP	219	0.0981	-	7470A	PPB	05 Aug 2022 13:52:34	U SMPL
N3997-15MS - 1	1986MS	14984	3.2143	-	7470A	PPB	05 Aug 2022 13:54:50	U SMPL
N3997-15MSD - 1	1986MSD	16239	3.4792	-	7470A	PPB	05 Aug 2022 13:59:41	U SMPL
CCV62 - 1	CCV62	24432	5.2084	-	7470A	PPB	05 Aug 2022 14:01:57	U SMPL
CCB62 - 1	CCB62	-61	0.0390	-	7470A	PPB	05 Aug 2022 14:04:14	U SMPL
PB146686TB - 1	PB146686TB	110	0.0751	-	7470A	PPB	05 Aug 2022 14:09:08	U SMPL
PB146691TB - 1	PB146691TB	-5	0.0508	-	7470A	PPB	05 Aug 2022 14:11:23	U SMPL
N3997-11 - 1	COMP1-0802	47	0.0618	-	7470A	PPB	05 Aug 2022 14:13:39	U SMPL
N3997-12 - 1	COMP2-0802	39	0.0601	-	7470A	PPB	05 Aug 2022 14:15:55	U SMPL
N3997-13 - 1	COMP3-0802	40	0.0603	-	7470A	PPB	05 Aug 2022 14:18:11	U SMPL
N3997-14 - 1	1987	55	0.0635	-	7470A	PPB	05 Aug 2022 14:20:26	U SMPL
N4016-09LX5 - 1		13	0.0546	-	7470A	PPB	05 Aug 2022 14:22:42	U SMPL
N4016-09A - 1		15907	3.4091	-	7470A	PPB	05 Aug 2022 14:24:58	U SMPL
N3997-15LX5 - 1		-48	0.0417	-	7470A	PPB	05 Aug 2022 14:27:15	U SMPL
N3997-15A - 1		15451	3.3129	-	7470A	PPB	05 Aug 2022 14:29:31	U SMPL
CCV63 - 1	CCV63	23668	5.0471	-	7470A	PPB	05 Aug 2022 14:34:43	U SMPL
CCB63 - 1	CCB63	-18	0.0481	-	7470A	PPB	05 Aug 2022 14:36:59	U SMPL