

LB121377


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

51788



A= 0.0000e+000

B= 1.9307e-004

C= 2.7344e-002

Rho= 0.9999587

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.038	0.038	56	0.000	56					-
0.2	0.200	0.189	-0.011	836	0.0 %	836					-6
2.5	2.500	2.507	0.007	12842	0.0 %	12842					0
5.0	5.000	4.934	-0.066	25412	0.0 %	25412					-1
7.5	7.500	7.507	0.007	38739	0.0 %	38739					0
10.0	10.000	10.026	0.026	51788	0.0 %	51788					0

LB121377 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 06 Aug 2022 12:59:30

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	56	-	0.0000	7471B	PPB	06 Aug 2022 13:10:58	S	Std
0.2 - 1	50.2	836	-	0.2000	7471B	PPB	06 Aug 2022 13:16:46	S	Std
2.5 - 1	52.5	12842	-	2.5000	7471B	PPB	06 Aug 2022 13:19:03	S	Std
5.0 - 1	55	25412	-	5.0000	7471B	PPB	06 Aug 2022 13:21:20	S	Std
7.5 - 1	57.5	38739	-	7.5000	7471B	PPB	06 Aug 2022 13:23:38	S	Std
10.0 - 1	510	51788	-	10.0000	7471B	PPB	06 Aug 2022 13:33:42	S	Std
ICV70 - 1	ICV70	18827	3.6623	-	7471B	PPB	06 Aug 2022 13:39:23	U	SMPL
ICB70 - 1	ICB70	-235	-0.0180	-	7471B	PPB	06 Aug 2022 13:41:37	U	SMPL
CCV65 - 1	CCV65	24739	4.8037	-	7471B	PPB	06 Aug 2022 13:43:55	U	SMPL
CCB65 - 1	CCB65	-69	0.0140	-	7471B	PPB	06 Aug 2022 13:46:11	U	SMPL
CRA - 1	CRA	774	0.1768	-	7471B	PPB	06 Aug 2022 13:48:28	U	SMPL
HighStd - 1	HighStd	51725	10.0139	-	7471B	PPB	06 Aug 2022 13:50:43	U	SMPL
ChkStd - 1	ChkStd	33612	6.5168	-	7471B	PPB	06 Aug 2022 13:55:40	U	SMPL
PB146788BL - 1	PBS	-196	-0.0105	-	7471B	PPB	06 Aug 2022 13:57:57	U	SMPL
PB146788BS - 1	LCSS	21885	4.2527	-	7471B	PPB	06 Aug 2022 14:02:46	U	SMPL
N4023-01 - 1	TP1-1-6	23846	4.6313	-	7471B	PPB	06 Aug 2022 14:05:02	U	SMPL
N4023-04 - 1	TP2-1-6	23189	4.5044	-	7471B	PPB	06 Aug 2022 14:07:19	U	SMPL
N4023-07 - 1	TP3-1-6	47124	9.1256	-	7471B	PPB	06 Aug 2022 14:09:37	U	SMPL
N4032-01 - 1	SU-03-080322	3840	0.7687	-	7471B	PPB	06 Aug 2022 14:11:55	U	SMPL
N4032-01DUP - 1	SU-03-080322DUP	4571	0.9098	-	7471B	PPB	06 Aug 2022 14:14:12	U	SMPL
CCV66 - 1	CCV66	24777	4.8110	-	7471B	PPB	06 Aug 2022 14:16:27	U	SMPL
CCB66 - 1	CCB66	-59	0.0160	-	7471B	PPB	06 Aug 2022 14:18:43	U	SMPL
N4032-01MS - 1	SU-03-080322MS	24166	4.6931	-	7471B	PPB	06 Aug 2022 14:23:03	U	SMPL
N4032-01MSD - 1	SU-03-080322MSD	26825	5.2064	-	7471B	PPB	06 Aug 2022 14:25:17	U	SMPL
N4036-01 - 1	MH-DD	323	0.0897	-	7471B	PPB	06 Aug 2022 14:27:34	U	SMPL
N4044-01 - 1	VNJ-230	2091	0.4311	-	7471B	PPB	06 Aug 2022 14:29:51	U	SMPL
N4044-03 - 1	VNJ-241	986	0.2177	-	7471B	PPB	06 Aug 2022 14:32:07	U	SMPL
N4054-01 - 1	OR-02-080422	1110	0.2417	-	7471B	PPB	06 Aug 2022 14:34:23	U	SMPL
N4060-01 - 1	JERSEY-CITY-SP	41785	8.0948	-	7471B	PPB	06 Aug 2022 14:36:39	U	SMPL
N4061-01 - 1	ETGI-328	107608	20.8032	-	7471B	PPB	06 Aug 2022 14:38:56	U	SMPL
N4061-03 - 1	COMP1-0804	411194	HIGH	-	7471B	PPB	06 Aug 2022 14:41:15	X	
N4061-05 - 1	COMP2-0804	64651	12.5095	-	7471B	PPB	06 Aug 2022 14:43:09	U	SMPL
CCV67 - 1	CCV67	25052	4.8641	-	7471B	PPB	06 Aug 2022 14:45:28	U	SMPL
CCB67 - 1	CCB67	-90	0.0100	-	7471B	PPB	06 Aug 2022 14:47:45	U	SMPL
N4061-07 - 1	COMP3-0804	101066	19.5401	-	7471B	PPB	06 Aug 2022 14:50:02	U	SMPL
N4062-01 - 1	OK-02-080422	545	0.1326	-	7471B	PPB	06 Aug 2022 14:52:18	U	SMPL
PB146789BL - 1	PBS	19	0.0310	-	7471B	PPB	06 Aug 2022 14:54:36	U	SMPL
PB146789BS - 1	LCSS	19820	3.8540	-	7471B	PPB	06 Aug 2022 14:56:52	U	SMPL
N4046-01 - 1	BEACONPOINT-BACKFILL-01	135	0.0534	-	7471B	PPB	06 Aug 2022 14:59:07	U	SMPL
N4046-01DUP - 1	BEACONPOINT-BACKFILL-01DUP	116	0.0497	-	7471B	PPB	06 Aug 2022 15:01:24	U	SMPL
N4046-01MS - 1	BEACONPOINT-BACKFILL-01MS	19844	3.8586	-	7471B	PPB	06 Aug 2022 15:03:41	U	SMPL
N4046-01MSD - 1	BEACONPOINT-BACKFILL-01MSD	19948	3.8787	-	7471B	PPB	06 Aug 2022 15:05:57	U	SMPL
N4046-03 - 1	BEACONPOINT-BACKFILL-02	94	0.0455	-	7471B	PPB	06 Aug 2022 15:08:16	U	SMPL
N4046-05 - 1	BEACONPOINT-BACKFILL-03	139	0.0542	-	7471B	PPB	06 Aug 2022 15:10:34	U	SMPL
CCV68 - 1	CCV68	25202	4.8931	-	7471B	PPB	06 Aug 2022 15:12:49	U	SMPL
CCB68 - 1	CCB68	-52	0.0173	-	7471B	PPB	06 Aug 2022 15:15:05	U	SMPL
N4046-07 - 1	BEACONPOINT-BACKFILL-04	95	0.0457	-	7471B	PPB	06 Aug 2022 15:17:22	U	SMPL
N4046-09 - 1	BEACONPOINT-BACKFILL-05	60	0.0389	-	7471B	PPB	06 Aug 2022 15:19:37	U	SMPL
N4046-11 - 1	BEACONPOINT-BACKFILL-06	62	0.0393	-	7471B	PPB	06 Aug 2022 15:21:53	U	SMPL
N4046-13 - 1	BEACONPOINT-BACKFILL-07	57	0.0383	-	7471B	PPB	06 Aug 2022 15:24:08	U	SMPL
N4046-15 - 1	BEACONPOINT-BACKFILL-08	83	0.0434	-	7471B	PPB	06 Aug 2022 15:26:24	U	SMPL
N4046-17 - 1	BEACONPOINT-BACKFILL-09	80	0.0428	-	7471B	PPB	06 Aug 2022 15:28:40	U	SMPL
N4046-19 - 1	BEACONPOINT-BACKFILL-10	71	0.0411	-	7471B	PPB	06 Aug 2022 15:30:57	U	SMPL
N4046-21 - 1	BEACONPOINT-BACKFILL-11	61	0.0391	-	7471B	PPB	06 Aug 2022 15:33:14	U	SMPL
PB146790BL - 1	PBS	17	0.0306	-	7471B	PPB	06 Aug 2022 15:35:30	U	SMPL
PB146790BS - 1	LCSS	19733	3.8372	-	7471B	PPB	06 Aug 2022 15:37:46	U	SMPL
CCV69 - 1	CCV69	26644	5.1715	-	7471B	PPB	06 Aug 2022 15:40:02	U	SMPL
CCB69 - 1	CCB69	-55	0.0167	-	7471B	PPB	06 Aug 2022 15:42:18	U	SMPL
N3962-01 - 1	RVE-187	1871	0.3886	-	7471B	PPB	06 Aug 2022 15:44:36	U	SMPL
N3962-01DUP - 1	RVE-187DUP	1762	0.3675	-	7471B	PPB	06 Aug 2022 15:46:51	U	SMPL
N3962-01MS - 1	RVE-187MS	18902	3.6767	-	7471B	PPB	06 Aug 2022 15:49:07	U	SMPL
N3962-01MSD - 1	RVE-187MSD	19893	3.6681	-	7471B	PPB	06 Aug 2022 15:51:23	U	SMPL
N3962-02 - 1	RVE-188	1536	0.3239	-	7471B	PPB	06 Aug 2022 15:53:39	U	SMPL
N3962-03 - 1	RVE-189	7018	1.3823	-	7471B	PPB	06 Aug 2022 15:55:55	U	SMPL
N3962-04 - 1	RVE-190	1493	0.3156	-	7471B	PPB	06 Aug 2022 16:01:37	U	SMPL
N3962-05 - 1	RVE-191	2098	0.4324	-	7471B	PPB	06 Aug 2022 16:03:53	U	SMPL
N3962-06 - 1	RVE-192	2582	0.5259	-	7471B	PPB	06 Aug 2022 16:06:10	U	SMPL
N3962-07 - 1	RVE-193	2387	0.4882	-	7471B	PPB	06 Aug 2022 16:08:26	U	SMPL
CCV70 - 1	CCV70	26821	5.2057	-	7471B	PPB	06 Aug 2022 16:10:44	U	SMPL
CCB70 - 1	CCB70	-51	0.0175	-	7471B	PPB	06 Aug 2022 16:13:00	U	SMPL
N3962-08 - 1	RVE-194	2827	0.5732	-	7471B	PPB	06 Aug 2022 16:15:16	U	SMPL
N3962-09 - 1	RVE-195	1706	0.3567	-	7471B	PPB	06 Aug 2022 16:17:33	U	SMPL
N3962-10 - 1	RVE-196	97	0.0461	-	7471B	PPB	06 Aug 2022 16:19:50	U	SMPL
N3962-11 - 1	RVE-197	1735	0.3623	-	7471B	PPB	06 Aug 2022 16:22:06	U	SMPL
N3962-12 - 1	RVE-198	3007	0.6079	-	7471B	PPB	06 Aug 2022 16:24:22	U	SMPL
N3962-13 - 1	RVE-199	2335	0.4782	-	7471B	PPB	06 Aug 2022 16:26:39	U	SMPL
N3962-14 - 1	RVE-199D	3123	0.6303	-	7471B	PPB	06 Aug 2022 16:28:56	U	SMPL
N3962-15 - 1	RVE-200	3423	0.6882	-	7471B	PPB	06 Aug 2022 16:31:13	U	SMPL
N3962-16 - 1	RVE-201	1850	0.3845	-	7471B	PPB	06 Aug 2022 16:33:32	U	SMPL
N3962-17 - 1	RVE-202	2221	0.4562	-	7471B	PPB	06 Aug 2022 16:35:49	U	SMPL
CCV71 - 1	CCV71	26867	5.2145	-	7471B	PPB	06 Aug 2022 16:38:05	U	SMPL
CCB71 - 1	CCB71	-50	0.0177	-	7471B	PPB	06 Aug 2022 16:40:22	U	SMPL
N3962-18 - 1	RVE-203	2955	0.5979	-	7471B	PPB	06 Aug 2022 16:42:38	U	SMPL
N3962-19 - 1	RVE-204	2163	0.4450	-	7471B	PPB	06 Aug 2022 16:44:56	U	SMPL
N3962-20 - 1	RVE-205	2405	0.4917	-	7471B	PPB	06 Aug 2022 16:47:12	U	SMPL
PB146791BL - 1	PBS	9	0.0291	-	7471B	PPB	06 Aug 2022 16:49:28	U	SMPL
PB146791BS - 1	LCSS	20048	3.8980	-	7471B	PPB	06 Aug 2022 16:51:44	U	SMPL
N3963-01 - 1	RVE-206	1782	0.3714	-	7471B	PPB	06 Aug 2022 16:54:01	U	SMPL
N3963-01DUP - 1	RVE-206DUP	1701	0.3558	-	7471B	PPB	06 Aug 2022 16:56:17	U	SMPL
N3963-01MS - 1	RVE-206MS	22282	4.3293	-	7471B	PPB	06 Aug 2022 16:58:34	U	SMPL
N3963-01MSD - 1	RVE-206MSD	21876	4.2509	-	7471B	PPB	06 Aug 2022 17:00:51	U	SMPL
N3963-02 - 1	RVE-207	4480	0.8923	-	7471B	PPB	06 Aug 2022 17:03:08	U	SMPL
CCV72 - 1	CCV72	25897	5.0273	-	7471B	PPB	06 Aug 2022 17:05:25	U	SMPL
					7471B	PPB	06 Aug 2022 17:07:43	U	SMPL

LB121377 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 06 Aug 2022 12:59:30

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
CCB72 - 1	CCB72	-52	0.0173	-7471B	PPB		06 Aug 2022 17:09:59	U	SMPL
N3963-03 - 1	RVE-208	3430	0.6896	-7471B	PPB		06 Aug 2022 17:12:17	U	SMPL
N3963-04 - 1	RVE-209	2234	0.4587	-7471B	PPB		06 Aug 2022 17:14:33	U	SMPL
N3963-05 - 1	RVE-210	3578	0.7181	-7471B	PPB		06 Aug 2022 17:16:49	U	SMPL
N3963-06 - 1	RVE-211	2249	0.4616	-7471B	PPB		06 Aug 2022 17:19:05	U	SMPL
N3963-07 - 1	RVE-212	1416	0.3007	-7471B	PPB		06 Aug 2022 17:21:22	U	SMPL
N3963-08 - 1	RVE-213	3617	0.7257	-7471B	PPB		06 Aug 2022 17:23:39	U	SMPL
N3963-09 - 1	RVE-214	3133	0.6322	-7471B	PPB		06 Aug 2022 17:25:55	U	SMPL
N3963-10 - 1	RVE-215	1931	0.4002	-7471B	PPB		06 Aug 2022 17:28:12	U	SMPL
N3963-11 - 1	RVE-216	726	0.1675	-7471B	PPB		06 Aug 2022 17:30:29	U	SMPL
N3963-12 - 1	RVE-217	1772	0.3695	-7471B	PPB		06 Aug 2022 17:32:46	U	SMPL
CCV73 - 1	CCV73	25176	4.8881	-7471B	PPB		06 Aug 2022 17:35:02	U	SMPL
CCB73 - 1	CCB73	-47	0.0183	-7471B	PPB		06 Aug 2022 17:37:19	U	SMPL
N3963-13 - 1	RVE-218	2129	0.4384	-7471B	PPB		06 Aug 2022 17:39:36	U	SMPL
N3963-14 - 1	RVE-219	993	0.2191	-7471B	PPB		06 Aug 2022 17:41:53	U	SMPL
N3963-15 - 1	RVE-219D	1313	0.2808	-7471B	PPB		06 Aug 2022 17:44:10	U	SMPL
N3963-16 - 1	RVE-220	2665	0.5419	-7471B	PPB		06 Aug 2022 17:46:26	U	SMPL
N3963-17 - 1	RVE-221	1410	0.2996	-7471B	PPB		06 Aug 2022 17:48:43	U	SMPL
N3963-18 - 1	RVE-222	1761	0.3673	-7471B	PPB		06 Aug 2022 17:50:59	U	SMPL
N3963-19 - 1	RVE-223	3632	0.7286	-7471B	PPB		06 Aug 2022 17:53:16	U	SMPL
N3963-20 - 1	RVE-224	2188	0.4498	-7471B	PPB		06 Aug 2022 17:55:32	U	SMPL
N4032-01LX5 - 1		767	0.1754	-7471B	PPB		06 Aug 2022 17:57:49	U	SMPL
N4032-01A - 1		24230	4.7054	-7471B	PPB		06 Aug 2022 18:00:06	U	SMPL
CCV74 - 1	CCV74	25367	4.9249	-7471B	PPB		06 Aug 2022 18:02:23	U	SMPL
CCB74 - 1	CCB74	-72	0.0134	-7471B	PPB		06 Aug 2022 18:04:41	U	SMPL
N4046-01LX5 - 1		14	0.0300	-7471B	PPB		06 Aug 2022 18:07:00	U	SMPL
N4046-01A - 1		19936	3.8764	-7471B	PPB		06 Aug 2022 18:09:17	U	SMPL
N3962-01LX5 - 1		251	0.0758	-7471B	PPB		06 Aug 2022 18:11:33	U	SMPL
N3962-01A - 1		23715	4.6060	-7471B	PPB		06 Aug 2022 18:13:51	U	SMPL
N3963-01LX5 - 1		298	0.0849	-7471B	PPB		06 Aug 2022 18:16:07	U	SMPL
N3963-01A - 1		22208	4.3152	-7471B	PPB		06 Aug 2022 18:18:25	U	SMPL
N4061-01DLX5 - 1		17962	3.4953	-7471B	PPB		06 Aug 2022 18:24:23	U	SMPL
N4061-05DLX5 - 1		12917	2.5212	-7471B	PPB		06 Aug 2022 18:28:59	U	SMPL
N4061-07DLX5 - 1		20876	4.0579	-7471B	PPB		06 Aug 2022 18:31:18	U	SMPL
N4061-03DLX25 - 1		21766	4.2297	-7471B	PPB		06 Aug 2022 18:33:35	U	SMPL
CCV75 - 1	CCV75	25294	4.9109	-7471B	PPB		06 Aug 2022 18:35:53	U	SMPL
CCB75 - 1	CCB75	-101	0.0078	-7471B	PPB		06 Aug 2022 18:38:12	U	SMPL