

LB121378


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

45264



A= 0.0000e+000

B= 2.2137e-004 slope

C= 2.3649e-002 intercept

Rho= 0.9999608

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.028	0.028	19	0.000	19					=
0.05	0.050					185	0				12
0.20	0.200	0.224	0.024	905	0.0 %	905					-2
2.50	2.500	2.460	-0.040	11005	0.0 %	11005					-1
5.00	5.000	4.973	-0.027	22358	0.0 %	22358					0
7.50	7.500	7.472	-0.028	33646	0.0 %	33646					0
10.0	10.000	10.044	0.044	45264	0.0 %	45264					0

LB121378

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	19	-		0 SFAM01.1	PPB	8/6/2022 18:53	Std
	0.2 S01	905	-		0.2 SFAM01.1	PPB	8/6/2022 18:55	Std
	2.5 S02	11005	-		2.5 SFAM01.1	PPB	8/6/2022 18:57	Std
	5 S03	22358	-		5 SFAM01.1	PPB	8/6/2022 19:00	Std
	7.5 S04	33646	-		7.5 SFAM01.1	PPB	8/6/2022 19:02	Std
	10 S05	45264	-		10 SFAM01.1	PPB	8/6/2022 19:04	Std
ICV071	ICV071	17980	4.0038 -		SFAM01.1	PPB	8/6/2022 19:07	SMPL
ICB071	ICB071	-28	0.0175 -		SFAM01.1	PPB	8/6/2022 19:09	SMPL
CCV076	CCV076	22618	5.0305 -		SFAM01.1	PPB	8/6/2022 19:12	SMPL
CCB076	CCB076	-24	0.0183 -		SFAM01.1	PPB	8/6/2022 19:14	SMPL
PB146803BL	PBS803	28	0.0298 -		SFAM01.1	PPB	8/6/2022 19:16	SMPL
N3803-01	GBHS9	903	0.2235 -		SFAM01.1	PPB	8/6/2022 19:18	SMPL
N3803-02	GBHS9D	859	0.2138 -		SFAM01.1	PPB	8/6/2022 19:21	SMPL
N3803-03	GBHS9S	12249	2.7352 -		SFAM01.1	PPB	8/6/2022 19:23	SMPL
N3803-04	GBHT0	3017	0.6915 -		SFAM01.1	PPB	8/6/2022 19:25	SMPL
N3803-05	GBHT1	3757	0.8553 -		SFAM01.1	PPB	8/6/2022 19:28	SMPL
N3803-06	GBHT2	2505	0.5782 -		SFAM01.1	PPB	8/6/2022 19:30	SMPL
N3803-07	GBHT3	2410	0.5571 -		SFAM01.1	PPB	8/6/2022 19:32	SMPL
N3803-08	GBHT4	570	0.1498 -		SFAM01.1	PPB	8/6/2022 19:34	SMPL
N3803-09	GBHT5	2298	0.5324 -		SFAM01.1	PPB	8/6/2022 19:37	SMPL
N3803-10	GBHT6	2538	0.5855 -		SFAM01.1	PPB	8/6/2022 19:39	SMPL
N3803-11	GBHT7	19610	4.3647 -		SFAM01.1	PPB	8/6/2022 19:41	SMPL
PB146804BL	PBS804	23	0.0287 -		SFAM01.1	PPB	8/6/2022 19:43	SMPL
N3807-01	GBHW1	754	0.1906 -		SFAM01.1	PPB	8/6/2022 19:46	SMPL
N3807-02	GBHW2	1775	0.4166 -		SFAM01.1	PPB	8/6/2022 19:48	SMPL
N3807-03	GBHW2D	1553	0.3674 -		SFAM01.1	PPB	8/6/2022 19:50	SMPL
N3807-04	GBHW2S	12229	2.7307 -		SFAM01.1	PPB	8/6/2022 19:52	SMPL
N3807-05	GBHW3	932	0.23 -		SFAM01.1	PPB	8/6/2022 19:55	SMPL
N3807-06	GBHW4	1453	0.3453 -		SFAM01.1	PPB	8/6/2022 19:57	SMPL
CCV077	CCV077	22711	5.0511 -		SFAM01.1	PPB	8/6/2022 20:02	SMPL
CCB077	CCB077	-15	0.0203 -		SFAM01.1	PPB	8/6/2022 20:04	SMPL
N3807-07	GBHW7	20376	4.5342 -		SFAM01.1	PPB	8/6/2022 20:15	SMPL
CCV078	CCV078	22555	5.0166 -		SFAM01.1	PPB	8/6/2022 20:20	SMPL
CCB078	CCB078	-29	0.0172 -		SFAM01.1	PPB	8/6/2022 20:22	SMPL