

LB121379


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

42077



A= 0.0000e+000

B= 2.3832e-004 slope

C= -3.8080e-002 y-intercept

Rho= 0.9993352

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.006	-0.006	134	0.000	134					1
0.05	0.050					185	0				28
0.20	0.200	0.256	0.056	1234	0.0 %	1234					3
2.50	2.500	2.581	0.081	10990	0.0 %	10990					5
5.00	5.000	4.726	-0.274	19990	0.0 %	19990					2
7.50	7.500	7.654	0.154	32275	0.0 %	32275					0
10.0	10.000	9.990	-0.010	42077	0.0 %	42077					

LB121379

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	134	-		0 SFAM01.1	PPB	8/7/2022 13:25	Std
	0.2 S01	1234	-		0.2 SFAM01.1	PPB	8/7/2022 13:27	Std
	2.5 S02	10990	-		2.5 SFAM01.1	PPB	8/7/2022 13:30	Std
	5 S03	19990	-		5 SFAM01.1	PPB	8/7/2022 13:32	Std
	7.5 S04	32275	-		7.5 SFAM01.1	PPB	8/7/2022 13:34	Std
	10 S05	42077	-		10 SFAM01.1	PPB	8/7/2022 13:39	Std
ICV072	ICV072	17261	4.0755 -		SFAM01.1	PPB	8/7/2022 13:42	SMPL
ICB072	ICB072	-6	-0.0395 -		SFAM01.1	PPB	8/7/2022 13:44	SMPL
CCV079	CCV079	21221	5.0193 -		SFAM01.1	PPB	8/7/2022 13:46	SMPL
CCB079	CCB079	-32	-0.0457 -		SFAM01.1	PPB	8/7/2022 13:49	SMPL
PB146807BL	PBW807	32	-0.0305 -		SFAM01.1	PPB	8/7/2022 13:51	SMPL
N3804-01	GBHT8	33	-0.0302 -		SFAM01.1	PPB	8/7/2022 13:53	SMPL
N3804-02	GBHT8D	28	-0.0314 -		SFAM01.1	PPB	8/7/2022 13:55	SMPL
N3804-03	GBHT8S	4089	0.9364 -		SFAM01.1	PPB	8/7/2022 13:58	SMPL
N3804-05	GBHW0	21204	5.0152 -		SFAM01.1	PPB	8/7/2022 14:02	SMPL
PB146808BL	PBW808	-14	-0.0414 -		SFAM01.1	PPB	8/7/2022 14:09	SMPL
N3808-01	GBHW5	26	-0.0319 -		SFAM01.1	PPB	8/7/2022 14:11	SMPL
N3808-02	GBHW5D	31	-0.0307 -		SFAM01.1	PPB	8/7/2022 14:14	SMPL
N3808-03	GBHW5S	4213	0.966 -		SFAM01.1	PPB	8/7/2022 14:16	SMPL
CCV080	CCV080	21717	5.1375 -		SFAM01.1	PPB	8/7/2022 14:18	SMPL
CCB080	CCB080	-11	-0.0407 -		SFAM01.1	PPB	8/7/2022 14:20	SMPL