



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB121385

Analyst By : JIGNESH

Supervisor Review By : Sohil

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	08/08/2022	08:30
2	CAL2	1	Water	NA	NA	20.3	7.00	08/08/2022	08:31
3	CAL3	1	Water	NA	NA	20.4	10.03	08/08/2022	08:32
4	ICV	1	Water	NA	NA	20.4	6.99	08/08/2022	08:35
5	CCV1	1	Water	NA	NA	20.4	2.01	08/08/2022	08:40
6	N4018-01	1	Water	NA	NA	20.2	4.89	08/08/2022	08:45
7	N4018-02	1	Water	NA	NA	20.2	5.34	08/08/2022	08:46
8	N4020-01	1	Water	NA	NA	20.2	5.66	08/08/2022	08:50
9	N4020-03	1	Water	NA	NA	20.1	8.81	08/08/2022	09:00
10	N4053-01	1	Water	NA	NA	20.4	4.49	08/08/2022	09:03
11	N4056-01	1	Water	NA	NA	20.2	4.90	08/08/2022	09:05
12	N4056-01DUP	1	Water	NA	NA	20.3	4.91	08/08/2022	09:06
13	CCV2	1	Water	NA	NA	20.5	12.03	08/08/2022	09:10

VB 121385

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph n4053 w WorkList ID : 162061 Department : Wet-Chemistry Date : 08-08-2022 08:22:38

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/03/2022	Water	N4018-01	pH	Cool 4 deg C	PSEG03	J11	33773	08/03/2022	9040C
08/03/2022	Water	N4018-02	pH	Cool 4 deg C	PSEG03	J11	33774	08/03/2022	9040C
08/03/2022	Water	N4020-01	pH	Cool 4 deg C	PSEG03	J11	34083	08/03/2022	9040C
08/03/2022	Water	N4020-03	pH	Cool 4 deg C	PSEG03	J11	34029	08/03/2022	9040C
08/03/2022	Water	N4053-01	pH	Cool 4 deg C	STER04	J11	POOL-WATER	08/03/2022	9040C
08/04/2022	Water	N4056-01	pH	Cool 4 deg C	PSEG03	J11	TOTE	08/04/2022	9040C

Date/Time 08/08/22 08:25
 Raw Sample Received by: JH/2011
 Raw Sample Relinquished by: CP Sm

Date/Time 08/08/22 12:00
 Raw Sample Received by: CP Sm
 Raw Sample Relinquished by: JH/2011