



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB121386

Analyst By : JIGNESH

Supervisor Review By : Sohil

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	08/08/2022	08:30
2	CAL2	1	Water	NA	NA	20.3	7.00	08/08/2022	08:31
3	CAL3	1	Water	NA	NA	20.4	10.03	08/08/2022	08:32
4	ICV	1	Water	NA	NA	20.4	6.99	08/08/2022	08:35
5	CCV1	1	Water	NA	NA	20.4	2.01	08/08/2022	09:15
6	N4074-03	1	Water	NA	NA	20.4	6.42	08/08/2022	09:19
7	N4074-04	1	Water	NA	NA	20.3	6.40	08/08/2022	09:20
8	N4077-03	1	Water	NA	NA	20.6	6.49	08/08/2022	09:22
9	N4090-01	1	Water	NA	NA	20.7	6.83	08/08/2022	09:24
10	N4090-01DUP	1	Water	NA	NA	20.8	6.84	08/08/2022	09:25
11	CCV2	1	Water	NA	NA	20.7	12.03	08/08/2022	09:26

WORKLIST(Hardcopy Internal Chain)

183121386

WorkList Name : corrosivity n4053 w WorkList ID : 162066 Department : Wet-Chemistry Date : 08-08-2022 08:25:56

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/05/2022	Water	N4074-03	Corrosivity	Cool 4 deg C	PSEG03	J11	315	08/05/2022	9040C
08/05/2022	Water	N4074-04	Corrosivity	Cool 4 deg C	PSEG03	J11	317	08/05/2022	9040C
08/05/2022	Water	N4077-03	Corrosivity	Cool 4 deg C	PSEG03	J11	2360	08/05/2022	9040C
08/05/2022	Water	N4090-01	Corrosivity	Cool 4 deg C	PSEG03	J11	164-166-COMP	08/05/2022	9040C

Date/Time 08/08/22 08:25
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 08/08/22 12:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]