



PERCENT SOLID

Supervisor: amarnath

Analyst: Nilesch

Date: 8/18/2022

OVENTEMP IN Celsius(°C): 108

Time IN: 17:00

In Date: 08/08/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:20

Out Date: 08/09/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB121389

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N4004-01	EY139	1	0.91	8.94	9.85	8.88	89.1	
N4004-02	EY139MS	2	0.91	8.94	9.85	8.88	89.1	
N4004-03	EY139MSD	3	0.91	8.94	9.85	8.88	89.1	
N4004-04	EY140	4	0.92	8.75	9.67	8.91	91.3	
N4004-05	VHBLK001	5	1.00	1.00	2.00	2.00	100.0	VHBLK
N4004-06	EY134	6	0.92	9.02	9.94	7.8	76.3	
N4004-07	EY135	7	0.92	8.79	9.71	8.26	83.5	
N4004-08	EY136	8	0.91	8.77	9.68	7.92	79.9	
N4004-09	EY137	9	0.91	8.81	9.72	9.32	95.5	
N4004-10	EY138	10	0.92	8.73	9.65	7.84	79.3	
N4004-11	EY141	11	0.91	9.07	9.98	7.38	71.3	
N4004-12	EY142	12	0.91	8.85	9.76	7.6	75.6	
N4006-01	JRMM8	13	1.00	1.00	2.00	1.93	93.0	DUST SAMPLES
N4006-02	JRMM8D	14	1.00	1.00	2.00	1.93	93.0	DUST SAMPLES
N4006-03	JRMM8S	15	1.00	1.00	2.00	1.93	93.0	DUST SAMPLES
N4006-04	JRMN0	16	1.00	1.00	2.00	1.938	93.8	DUST SAMPLES
N4006-05	JRMN1	17	1.00	1.00	2.00	1.934	93.4	DUST SAMPLES
N4006-06	JRMN3	18	1.00	1.00	2.00	1.94	94.0	DUST SAMPLES
N4007-01	JRMM9	19	1.00	1.00	2.00	1.93	93.0	DUST SAMPLES
N4007-02	JRMM9D	20	1.00	1.00	2.00	1.93	93.0	DUST SAMPLES
N4007-03	JRMM9S	21	1.00	1.00	2.00	1.93	93.0	DUST SAMPLES
N4007-04	JRMN2	22	1.00	1.00	2.00	1.942	94.2	DUST SAMPLES

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$