



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB121392

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

True Value of CCV4 = 12.00 Control Limits[+/- 0.1].

True Value of CCV5 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.9	4.01	08/08/2022	09:00
2	CAL2	1	Water	NA	NA	20.9	7.00	08/08/2022	09:01
3	CAL3	1	Water	NA	NA	20.9	10.03	08/08/2022	09:03
4	ICV	1	Water	NA	NA	20.8	6.98	08/08/2022	09:05
5	CCV1	1	Water	NA	NA	20.8	2.01	08/08/2022	09:06
6	N4017-01	1	Solid	20.02	20	22.6	6.75	08/08/2022	09:10
7	N4017-01DUP	1	Solid	20.03	20	22.7	6.76	08/08/2022	09:11
8	N4017-02	1	Solid	20.04	20	22.4	7.06	08/08/2022	09:12
9	N4017-03	1	Solid	20.01	20	22.3	7.19	08/08/2022	09:15
10	N4017-04	1	Solid	20.04	20	22.4	7.29	08/08/2022	09:20
11	N4017-05	1	Solid	20.03	20	21.7	7.26	08/08/2022	09:22
12	N4017-06	1	Solid	20.04	20	22.4	7.29	08/08/2022	09:25
13	N4017-07	1	Solid	20.03	20	20.5	6.65	08/08/2022	09:30
14	N4017-08	1	Solid	20.04	20	21.0	6.50	08/08/2022	09:33
15	N4017-09	1	Solid	20.02	20	20.7	7.51	08/08/2022	09:36
16	CCV2	1	Water	NA	NA	20.5	12.03	08/08/2022	09:38
17	N4017-10	1	Solid	20.02	20	21.2	6.78	08/08/2022	09:40
18	N4023-02	1	Solid	20.01	20	24.5	8.13	08/08/2022	09:43
19	N4023-05	1	Solid	20.03	20	24.6	6.03	08/08/2022	09:45

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
20	N4023-08	1	Solid	20.04	20	24.4	6.25	08/08/2022	09:48
21	N4060-08	1	Solid	20.02	20	24.9	8.92	08/08/2022	09:50
22	N4061-02	1	Solid	20.03	20	23.5	6.06	08/08/2022	09:52
23	N4061-04	1	Solid	20.04	20	23.0	6.26	08/08/2022	10:00
24	N4061-06	1	Solid	20.03	20	23.3	6.99	08/08/2022	10:02
25	N4061-08	1	Solid	20.04	20	23.3	6.42	08/08/2022	10:05
26	N4064-01	1	Solid	20.02	20	24.1	8.57	08/08/2022	10:10
27	CCV3	1	Water	NA	NA	20.9	2.01	08/08/2022	10:13
28	N4064-02	1	Solid	20.03	20	24.3	8.61	08/08/2022	10:15
29	N4064-03	1	Solid	20.04	20	24.4	8.21	08/08/2022	10:20
30	N4064-04	1	Solid	20.03	20	24.4	7.30	08/08/2022	10:23
31	N4064-05	1	Solid	20.02	20	24.3	7.48	08/08/2022	10:25
32	N4064-06	1	Solid	20.01	20	23.8	5.51	08/08/2022	10:26
33	N4064-07	1	Solid	20.02	20	23.9	7.11	08/08/2022	10:30
34	N4064-08	1	Solid	20.03	20	23.9	6.44	08/08/2022	10:33
35	N4074-01	1	Solid	20.04	20	23.4	6.77	08/08/2022	10:36
36	N4077-01	1	Solid	20.02	20	23.7	6.06	08/08/2022	10:40
37	N4077-02	1	Solid	20.02	20	20.6	4.75	08/08/2022	10:43
38	CCV4	1	Water	NA	NA	20.8	12.03	08/08/2022	10:45
39	N4081-01	1	Solid	20.03	20	23.3	6.52	08/08/2022	10:48
40	N4081-01DUP	1	Solid	20.04	20	23.4	6.53	08/08/2022	10:49
41	CCV5	1	Water	NA	NA	20.6	2.01	08/08/2022	10:50