

LB121395


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

54290



A= 0.0000e+000

B= 1.8381e-004

C= -2.3580e-002

Rho= 0.9999618

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.010	-0.010	72	0.000	72					0
0.2	0.200	0.200	-0.000	1216	0.0 %	1216					0
2.5	2.500	2.472	-0.028	13579	0.0 %	13579					0
5.0	5.000	5.045	0.045	27572	0.0 %	27572					0
7.5	7.500	7.538	0.038	41136	0.0 %	41136					0
10.0	10.000	9.956	-0.044	54290	0.0 %	54290					0

LLB121395 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 08 Aug 2022 12:30:34

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	72	-	0.0000	7471B	PPB	08 Aug 2022 12:40:54	S	Std
0.2 - 1	S0.2	1216	-	0.2000	7471B	PPB	08 Aug 2022 12:43:11	S	Std
2.5 - 1	S2.5	13579	-	2.5000	7471B	PPB	08 Aug 2022 12:45:29	S	Std
5.0 - 1	S5	27572	-	5.0000	7471B	PPB	08 Aug 2022 12:47:46	S	Std
7.5 - 1	S7.5	41136	-	7.5000	7471B	PPB	08 Aug 2022 12:50:02	S	Std
10.0 - 1	S10	54290	-	10.0000	7471B	PPB	08 Aug 2022 12:52:20	S	Std
ICV74 - 1	ICV74	20567	3.7569	-	7471B	PPB	08 Aug 2022 12:55:03	U	SMPL
ICB74 - 1	ICB74	-58	-0.0342	-	7471B	PPB	08 Aug 2022 12:57:19	U	SMPL
CCV86 - 1	CCV86	27456	5.0232	-	7471B	PPB	08 Aug 2022 12:59:35	U	SMPL
CCB86 - 1	CCB86	-54	-0.0335	-	7471B	PPB	08 Aug 2022 13:01:51	U	SMPL
CRA - 1	CRA	1316	0.2183	-	7471B	PPB	08 Aug 2022 13:04:07	U	SMPL
HighStd - 1	HighStd	52049	9.5437	-	7471B	PPB	08 Aug 2022 13:06:22	U	SMPL
ChkStd - 1	ChkStd	36868	6.7533	-	7471B	PPB	08 Aug 2022 13:08:37	U	SMPL
PB146827BL - 1	PB146827BL	-89	-0.0399	-	7471B	PPB	08 Aug 2022 13:10:55	U	SMPL
N4068-01 - 1	BU-02-080522	1479	0.2483	-	7471B	PPB	08 Aug 2022 13:15:29	U	SMPL
N4073-01 - 1	CHRT-24133	27345	5.0028	-	7471B	PPB	08 Aug 2022 13:17:46	U	SMPL
N4086-01 - 1	RLA-4711-WC	3280	0.5793	-	7471B	PPB	08 Aug 2022 13:20:03	U	SMPL
N4088-01 - 1	PL-02-080522	2446	0.4260	-	7471B	PPB	08 Aug 2022 13:22:22	U	SMPL
N4088-01DUP - 1	PL-02-080522DUP	2134	0.3687	-	7471B	PPB	08 Aug 2022 13:24:37	U	SMPL
PB146827BS - 1	PB146827BS	19213	3.5080	-	7471B	PPB	08 Aug 2022 13:29:46	U	SMPL
CCV87 - 1	CCV87	25711	4.7025	-	7471B	PPB	08 Aug 2022 13:32:18	U	SMPL
CCB87 - 1	CCB87	-52	-0.0331	-	7471B	PPB	08 Aug 2022 13:34:35	U	SMPL
N4088-01MS - 1	PL-02-080522MS	21983	4.0172	-	7471B	PPB	08 Aug 2022 13:36:52	U	SMPL
N4088-01MSD - 1	PL-02-080522MSD	21414	3.9126	-	7471B	PPB	08 Aug 2022 13:39:07	U	SMPL
N4089-02 - 1	VNJ-228	4137	0.7369	-	7471B	PPB	08 Aug 2022 13:41:27	U	SMPL
N4091-01 - 1	PSC-164718	3293	0.5817	-	7471B	PPB	08 Aug 2022 13:43:43	U	SMPL
N4091-03 - 1	PSC-164717	6484	1.1683	-	7471B	PPB	08 Aug 2022 13:45:59	U	SMPL
N4091-05 - 1	PSC-124037	2796	0.4904	-	7471B	PPB	08 Aug 2022 13:48:16	U	SMPL
N4091-07 - 1	PSC-124042	1433	0.2398	-	7471B	PPB	08 Aug 2022 13:50:33	U	SMPL
N4091-09 - 1	RO-124019	1247	0.2056	-	7471B	PPB	08 Aug 2022 13:52:50	U	SMPL
N4088-01LX5 - 1		453	0.0597	-	7471B	PPB	08 Aug 2022 13:55:07	U	SMPL
N4088-01A - 1		23935	4.3760	-	7471B	PPB	08 Aug 2022 13:57:25	U	SMPL
CCV88 - 1	CCV88	30043	5.4987	-	7471B	PPB	08 Aug 2022 14:07:44	U	SMPL
CCB88 - 1	CCB88	-53	-0.0333	-	7471B	PPB	08 Aug 2022 14:09:59	U	SMPL