

23121416


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

39425



A= 0.0000e+000

B= 2.4635e-004

C= -7.6588e-002

Rho= 0.9979483

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.028	-0.028	199	0.000	199					
0.2	0.200	0.205	0.005	1143	0.0 %	1143					
2.5	2.500	2.369	-0.131	9927	0.0 %	9927					
5.0	5.000	5.100	0.100	21013	0.0 %	21013					
7.5	7.500	7.918	0.418	32450	0.0 %	32450					
10.0	10.000	9.636	-0.364	39425	0.0 %	39425					

9.1
-2
-5
2
6
-4

LB121416

Method: 7470A

Operator: Admin

Date of Analysis: 09 Aug 2022 12:46:19

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	199	-	0.0000	7470A	PPB	09 Aug 2022 13:42:22	S	Std
0.2 - 1	50.2	1143	-	0.2000	7470A	PPB	09 Aug 2022 13:44:41	S	Std
2.5 - 1	52.5	9927	-	2.5000	7470A	PPB	09 Aug 2022 13:46:58	S	Std
5.0 - 1	55	21013	-	5.0000	7470A	PPB	09 Aug 2022 13:49:15	S	Std
7.5 - 1	57.5	32450	-	7.5000	7470A	PPB	09 Aug 2022 13:51:32	S	Std
10.0 - 1	59	39425	-	10.0000	7470A	PPB	09 Aug 2022 13:56:24	S	Std
ICV75 - 1	ICV75	15052	3.6315	-	7470A	PPB	09 Aug 2022 13:59:39	U	SMPL
ICB75 - 1	ICB75	415	0.0256	-	7470A	PPB	09 Aug 2022 14:01:54	U	SMPL
CCV89 - 1	CCV89	22018	5.3476	-	7470A	PPB	09 Aug 2022 14:06:42	U	SMPL
CCB89 - 1	CCB89	390	0.0195	-	7470A	PPB	09 Aug 2022 14:08:57	U	SMPL
CRA - 1	CRA	1139	0.2040	-	7470A	PPB	09 Aug 2022 14:11:13	U	SMPL
HighStd - 1	HighStd	40073	9.7956	-	7470A	PPB	09 Aug 2022 14:13:28	U	SMPL
ChkStd - 1	ChkStd	28874	7.0366	-	7470A	PPB	09 Aug 2022 14:15:43	U	SMPL
PB146852BL - 1	PBW	19	-0.0719	-	7470A	PPB	09 Aug 2022 14:18:00	U	SMPL
PB146852BS - 1	LCSW	15906	3.8419	-	7470A	PPB	09 Aug 2022 14:22:47	U	SMPL
N4036-02 - 1	MH-DD	13	-0.0734	-	7470A	PPB	09 Aug 2022 14:25:03	U	SMPL
N4060-08 - 1	JERSEY-CITY-SP	86	-0.0554	-	7470A	PPB	09 Aug 2022 14:27:19	U	SMPL
N4061-02 - 1	ETGI-328	128	-0.0451	-	7470A	PPB	09 Aug 2022 14:29:36	U	SMPL
N4061-04 - 1	COMP1-0804	292	-0.0047	-	7470A	PPB	09 Aug 2022 14:31:53	U	SMPL
N4061-06 - 1	COMP2-0804	362	0.0126	-	7470A	PPB	09 Aug 2022 14:34:08	U	SMPL
CCV90 - 1	CCV90	19222	4.6588	-	7470A	PPB	09 Aug 2022 14:36:24	U	SMPL
CCB90 - 1	CCB90	408	0.0239	-	7470A	PPB	09 Aug 2022 14:38:39	U	SMPL
N4061-08 - 1	COMP3-0804	257	-0.0133	-	7470A	PPB	09 Aug 2022 14:40:55	U	SMPL
N4077-01 - 1	2361	279	-0.0079	-	7470A	PPB	09 Aug 2022 14:43:10	U	SMPL
N4077-02 - 1	2378	424	0.0279	-	7470A	PPB	09 Aug 2022 14:45:25	U	SMPL
N4080-01 - 1	149TH-ST-DRUM-PPE	227	-0.0207	-	7470A	PPB	09 Aug 2022 14:47:41	U	SMPL
N4080-02 - 1	149TH-ST-DRUM-POLY	128	-0.0451	-	7470A	PPB	09 Aug 2022 14:49:57	U	SMPL
N4081-01 - 1	2616	186	-0.0308	-	7470A	PPB	09 Aug 2022 14:52:13	U	SMPL
N4091-02 - 1	PSC-164718	229	-0.0202	-	7470A	PPB	09 Aug 2022 14:54:30	U	SMPL
N4091-04 - 1	PSC-164717	188	-0.0303	-	7470A	PPB	09 Aug 2022 14:56:47	U	SMPL
N4091-06 - 1	PSC-124037	198	-0.0278	-	7470A	PPB	09 Aug 2022 14:59:04	U	SMPL
N4091-08 - 1	PSC-124042	301	-0.0024	-	7470A	PPB	09 Aug 2022 15:01:19	U	SMPL
CCV91 - 1	CCV91	21732	5.2772	-	7470A	PPB	09 Aug 2022 15:03:35	U	SMPL
CCB91 - 1	CCB91	451	0.0345	-	7470A	PPB	09 Aug 2022 15:05:50	U	SMPL
N4091-10 - 1	RO-124019	115	-0.0483	-	7470A	PPB	09 Aug 2022 15:08:07	U	SMPL
N4107-06 - 1	17183	82	-0.0564	-	7470A	PPB	09 Aug 2022 15:10:22	U	SMPL
N4107-06DUP - 1	17183DUP	424	0.0279	-	7470A	PPB	09 Aug 2022 15:12:38	U	SMPL
N4107-06MS - 1	17183MS	15936	3.8493	-	7470A	PPB	09 Aug 2022 15:14:54	U	SMPL
N4107-06MSD - 1	17183MSD	14666	3.5364	-	7470A	PPB	09 Aug 2022 15:19:52	U	SMPL
PB146853BL - 1	PBW	4	-0.0756	-	7470A	PPB	09 Aug 2022 15:22:08	U	SMPL
PB146853BS - 1	LCSW	13840	3.3330	-	7470A	PPB	09 Aug 2022 15:24:25	U	SMPL
N4074-03 - 1	315	15	-0.0729	-	7470A	PPB	09 Aug 2022 15:26:42	U	SMPL
N4074-04 - 1	317	80	-0.0569	-	7470A	PPB	09 Aug 2022 15:29:00	U	SMPL
N4077-03 - 1	2360	104	-0.0510	-	7470A	PPB	09 Aug 2022 15:31:15	U	SMPL
CCV92 - 1	CCV92	20503	4.9744	-	7470A	PPB	09 Aug 2022 15:33:31	U	SMPL
CCB92 - 1	CCB92	408	0.0239	-	7470A	PPB	09 Aug 2022 15:35:47	U	SMPL
N4090-01 - 1	164-166-COMP	64	-0.0608	-	7470A	PPB	09 Aug 2022 15:38:03	U	SMPL
N4090-01DUP - 1	164-166-COMPDUP	108	-0.0500	-	7470A	PPB	09 Aug 2022 15:40:18	U	SMPL
N4090-01MS - 1	164-166-COMPMS	14698	3.5443	-	7470A	PPB	09 Aug 2022 15:42:34	U	SMPL
N4090-01MSD - 1	164-166-COMPMSD	13843	3.3337	-	7470A	PPB	09 Aug 2022 15:44:50	U	SMPL
N4107-04 - 1	COMP-2	45	-0.0655	-	7470A	PPB	09 Aug 2022 15:47:05	U	SMPL
N4107-14 - 1	COMP-1	762	0.1111	-	7470A	PPB	09 Aug 2022 15:49:21	U	SMPL
PB146858BL - 1	PBW	112	-0.0490	-	7470A	PPB	09 Aug 2022 15:54:43	U	SMPL
PB146858BS - 1	LCSW	16223	3.9200	-	7470A	PPB	09 Aug 2022 15:57:00	U	SMPL
N4104-01 - 1	MTL-1	61	-0.0616	-	7470A	PPB	09 Aug 2022 15:59:17	U	SMPL
N4104-01DUP - 1	MTL-1DUP	78	-0.0574	-	7470A	PPB	09 Aug 2022 16:01:33	U	SMPL
CCV93 - 1	CCV93	20489	4.9710	-	7470A	PPB	09 Aug 2022 16:03:48	U	SMPL
CCB93 - 1	CCB93	198	-0.0278	-	7470A	PPB	09 Aug 2022 16:06:04	U	SMPL
N4104-01MS - 1	MTL-1MS	15850	3.8281	-	7470A	PPB	09 Aug 2022 16:11:20	U	SMPL
N4104-01MSD - 1	MTL-1MSD	14087	3.3938	-	7470A	PPB	09 Aug 2022 16:13:36	U	SMPL
PB146826TB - 1	PB146826TB	33	-0.0685	-	7470A	PPB	09 Aug 2022 16:15:51	U	SMPL
PB146832TB - 1	PB146832TB	64	-0.0608	-	7470A	PPB	09 Aug 2022 16:18:08	U	SMPL
N4107-06LX5 - 1		11	-0.0739	-	7470A	PPB	09 Aug 2022 16:20:24	U	SMPL
N4107-06A - 1		13648	3.2857	-	7470A	PPB	09 Aug 2022 16:22:40	U	SMPL
N4090-01LX5 - 1		-69	-0.0936	-	7470A	PPB	09 Aug 2022 16:24:57	U	SMPL
N4090-01A - 1		13393	3.2228	-	7470A	PPB	09 Aug 2022 16:27:14	U	SMPL
N4104-01LX5 - 1		-74	-0.0948	-	7470A	PPB	09 Aug 2022 16:29:31	U	SMPL
N4104-01A - 1		17036	4.1203	-	7470A	PPB	09 Aug 2022 16:38:58	U	SMPL
CCV94 - 1	CCV94	18939	4.5891	-	7470A	PPB	09 Aug 2022 16:41:14	U	SMPL
CCB94 - 1	CCB94	375	0.0158	-	7470A	PPB	09 Aug 2022 16:43:32	U	SMPL