



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB121418

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	08/09/2022	12:15
2	CAL2	1	Water	NA	NA	20.7	7.00	08/09/2022	12:16
3	CAL3	1	Water	NA	NA	20.7	10.03	08/09/2022	12:19
4	ICV	1	Water	NA	NA	20.8	6.98	08/09/2022	12:20
5	CCV1	1	Water	NA	NA	20.8	2.01	08/09/2022	12:22
6	N4112-01	1	Water	NA	NA	25.5	4.53	08/09/2022	12:25
7	N4112-01DUP	1	Water	NA	NA	25.6	4.54	08/09/2022	12:26
8	N4113-01	1	Water	NA	NA	24.4	4.76	08/09/2022	12:30
9	N4113-02	1	Water	NA	NA	24.9	4.82	08/09/2022	12:32
10	N4113-03	1	Water	NA	NA	24.9	4.87	08/09/2022	12:37
11	CCV2	1	Water	NA	NA	20.8	12.03	08/09/2022	12:40

WORKLIST(Hardcopy Internal Chain)

VB 12/24/18

WorkList Name : ph n4112

WorkList ID : 162123

Department : Wet-Chemistry

Date : 08-09-2022 12:08:04

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/21/2022	Water	N4112-01	pH	Cool 4 deg C	STER10	N11	CELL-A-POOL-WATER(DE-IONIZED)	07/21/2022	9040C
08/02/2022	Water	N4113-01	pH	Cool 4 deg C	STER13	N11	CELL-A-METALS-ANAYLSIS	08/02/2022	9040C
08/02/2022	Water	N4113-02	pH	Cool 4 deg C	STER13	N11	CELL-B-METALS-ANAYLSIS	08/02/2022	9040C
08/02/2022	Water	N4113-03	pH	Cool 4 deg C	STER13	N11	CELL-C-METALS-ANAYLSIS	08/02/2022	9040C

Date/Time 08/09/22 12:10
 Raw Sample Received by: ACW
 Raw Sample Relinquished by: CP

Date/Time 08/09/22 13:30
 Raw Sample Received by: CP
 Raw Sample Relinquished by: JB