



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB121420

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.3	08/09/2022	11:50
2	N4056-01		1	12	100.00	>212.0	>212.0	08/09/2022	12:15
3	N4056-01DUP		1	12	100.00	>212.0	>212.0	08/09/2022	12:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB121420

WorkList Name : fp8822 WorkList ID : 162093 Department : Wet-Chemistry Date : 08-08-2022 16:45:57

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/09/2022	Water	N4056-01	Flash Point		Cool 4 deg C	J11	TOTE	08/04/2022	1010B

Date/Time 08/09/22 11:30
Raw Sample Received by: KS (WC)
Raw Sample Relinquished by: JACDC1

Date/Time 08/09/22 12:45
Raw Sample Received by: KS (WC)
Raw Sample Relinquished by: KS (WC)