



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB121424

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 98.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.9	4.01	08/09/2022	12:50
2	CAL2	1	Water	NA	NA	20.9	7.01	08/09/2022	12:51
3	CAL3	1	Water	NA	NA	20.9	10.03	08/09/2022	12:52
4	ICV	1	Water	NA	NA	20.8	6.99	08/09/2022	13:00
5	CCV1	1	Water	NA	NA	20.8	2.01	08/09/2022	13:05
6	N4107-04	1	Water	NA	NA	23.6	7.11	08/09/2022	13:10
7	N4107-04DUP	1	Water	NA	NA	23.6	7.12	08/09/2022	13:11
8	N4107-14	1	Water	NA	NA	23.3	2.57	08/09/2022	13:15
9	CCV2	1	Water	NA	NA	20.9	12.03	08/09/2022	13:20