

LB121429


SFAM01.1

INSTRUMENT ID: CYD

Linear

 μ Abs.:

43019



A= 0.0000e+000

B= 2.3434e-004 slope

C= 1.6118e-002 y-intercept

Rho= 0.9998572

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.033	0.033	73	0.000	73					
0.05	0.050					185	0				
0.20	0.200	0.224	0.024	885	0.0 %	885					
2.50	2.500	2.498	-0.002	10592	0.0 %	10592					
5.00	5.000	4.942	-0.058	21019	0.0 %	21019					
7.50	7.500	7.406	-0.094	31535	0.0 %	31535					
10.0	10.000	10.097	0.097	43019	0.0 %	43019					

11.0211

LB121429

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type
	0 S0	73	-		0 SFAM01.1	PPB	8/10/2022 6:58	Std
	0.2 S01	885	-		0.2 SFAM01.1	PPB	8/10/2022 7:00	Std
	2.5 S02	10592	-		2.5 SFAM01.1	PPB	8/10/2022 7:05	Std
	5 S03	21019	-		5 SFAM01.1	PPB	8/10/2022 7:07	Std
	7.5 S04	31535	-		7.5 SFAM01.1	PPB	8/10/2022 7:10	Std
	10 S05	43019	-		10 SFAM01.1	PPB	8/10/2022 7:12	Std
ICV076	ICV076	16291	3.8338 -		SFAM01.1	PPB	8/10/2022 7:15	SMPL
ICB076	ICB076	27	0.0224 -		SFAM01.1	PPB	8/10/2022 7:17	SMPL
CCV095	CCV095	21155	4.9736 -		SFAM01.1	PPB	8/10/2022 7:19	SMPL
CCB095	CCB095	21	0.021 -		SFAM01.1	PPB	8/10/2022 7:22	SMPL
N3853-01	YBB68	1111	0.2765 -		SFAM01.1	PPB	8/10/2022 7:26	SMPL
N3853-02	YBB69	1090	0.2715 -		SFAM01.1	PPB	8/10/2022 7:28	SMPL
N3853-03	YBB70	1665	0.4063 -		SFAM01.1	PPB	8/10/2022 7:31	SMPL
N3853-04	YBB71	1430	0.3512 -		SFAM01.1	PPB	8/10/2022 7:33	SMPL
N3853-05	YBB72	1171	0.2905 -		SFAM01.1	PPB	8/10/2022 7:35	SMPL
N3853-06	YBB73	1261	0.3116 -		SFAM01.1	PPB	8/10/2022 7:37	SMPL
N3853-07	YBB74	1414	0.3475 -		SFAM01.1	PPB	8/10/2022 7:40	SMPL
N3853-08	YBB75	1051	0.2624 -		SFAM01.1	PPB	8/10/2022 7:42	SMPL
N3853-09	YBB76	1236	0.3058 -		SFAM01.1	PPB	8/10/2022 7:44	SMPL
N3853-10	YBB77	1369	0.3369 -		SFAM01.1	PPB	8/10/2022 7:47	SMPL
N3853-11	YBB78	1306	0.3222 -		SFAM01.1	PPB	8/10/2022 7:49	SMPL
N3853-12	YBB85	1136	0.2823 -		SFAM01.1	PPB	8/10/2022 7:51	SMPL
N3853-13	YBB86	1028	0.257 -		SFAM01.1	PPB	8/10/2022 7:53	SMPL
N3853-14	YBB87	868	0.2195 -		SFAM01.1	PPB	8/10/2022 7:56	SMPL
N3853-15	YBBB2	1362	0.3353 -		SFAM01.1	PPB	8/10/2022 7:58	SMPL
N3853-16	YBBB3	1370	0.3372 -		SFAM01.1	PPB	8/10/2022 8:00	SMPL
N3853-17	YBBB4	1177	0.2919 -		SFAM01.1	PPB	8/10/2022 8:02	SMPL
N3853-18	YBB53	14061	3.3112 -		SFAM01.1	PPB	8/10/2022 8:05	SMPL
N3853-19	YBB53D	14757	3.4743 -		SFAM01.1	PPB	8/10/2022 8:07	SMPL
CCV096	CCV096	20698	4.8665 -		SFAM01.1	PPB	8/10/2022 8:09	SMPL
CCB096	CCB096	1	0.0164 -		SFAM01.1	PPB	8/10/2022 8:11	SMPL
N3853-20	YBB53S	24849	5.8393 -		SFAM01.1	PPB	8/10/2022 8:14	SMPL
N3853-21	YBBB9	572	0.1502 -		SFAM01.1	PPB	8/10/2022 8:16	SMPL
N3853-22	YBBCO	10585	2.4966 -		SFAM01.1	PPB	8/10/2022 8:18	SMPL
PB146889BL	PBS889	-33	0.0084 -		SFAM01.1	PPB	8/10/2022 8:20	SMPL
N3869-01	YBB61	949	0.2385 -		SFAM01.1	PPB	8/10/2022 8:23	SMPL
N3869-02	YBB62	541	0.1429 -		SFAM01.1	PPB	8/10/2022 8:25	SMPL
N3869-03	YBB63	1399	0.344 -		SFAM01.1	PPB	8/10/2022 8:27	SMPL
N3869-04	YBB79	1133	0.2816 -		SFAM01.1	PPB	8/10/2022 8:30	SMPL
N3869-05	YBB80	1924	0.467 -		SFAM01.1	PPB	8/10/2022 8:32	SMPL
N3869-06	YBB81	1383	0.3402 -		SFAM01.1	PPB	8/10/2022 8:34	SMPL
N3869-07	YBB82	953	0.2394 -		SFAM01.1	PPB	8/10/2022 8:36	SMPL
N3869-08	YBB83	1086	0.2706 -		SFAM01.1	PPB	8/10/2022 8:39	SMPL
N3869-09	YBB84	1189	0.2947 -		SFAM01.1	PPB	8/10/2022 8:41	SMPL
N3869-10	YBB88	1588	0.3883 -		SFAM01.1	PPB	8/10/2022 8:43	SMPL
N3869-11	YBB89	1043	0.2605 -		SFAM01.1	PPB	8/10/2022 8:45	SMPL

LB121429

INSTRUMENT ID : CV1

N3869-12	YBB90	995	0.2493 -	SFAM01.1	PPB	8/10/2022 8:48 SMPL
N3869-13	YBB91	686	0.1769 -	SFAM01.1	PPB	8/10/2022 8:50 SMPL
N3869-14	YBB92	864	0.2186 -	SFAM01.1	PPB	8/10/2022 8:52 SMPL
N3869-15	YBB93	726	0.1862 -	SFAM01.1	PPB	8/10/2022 8:54 SMPL
N3869-16	YBB94	1135	0.2821 -	SFAM01.1	PPB	8/10/2022 8:57 SMPL
CCV097	CCV097	19303	4.5396 -	SFAM01.1	PPB	8/10/2022 8:59 SMPL
CCB097	CCB097	-53	0.0037 -	SFAM01.1	PPB	8/10/2022 9:01 SMPL
N3869-17	YBB95	800	0.2036 -	SFAM01.1	PPB	8/10/2022 9:04 SMPL
N3869-18	YBB96	931	0.2343 -	SFAM01.1	PPB	8/10/2022 9:06 SMPL
N3869-19	YBB49	924	0.2326 -	SFAM01.1	PPB	8/10/2022 9:08 SMPL
N3869-20	YBB56	1051	0.2624 -	SFAM01.1	PPB	8/10/2022 9:10 SMPL
N3869-21	YBB56D	1081	0.2694 -	SFAM01.1	PPB	8/10/2022 9:13 SMPL
N3869-22	YBB56S	11594	2.7331 -	SFAM01.1	PPB	8/10/2022 9:15 SMPL
PB146890BL	PBS890	121	0.0445 -	SFAM01.1	PPB	8/10/2022 9:17 SMPL
N3915-01	YBB50	919	0.2315 -	SFAM01.1	PPB	8/10/2022 9:19 SMPL
N3915-02	YBB52	369	0.1026 -	SFAM01.1	PPB	8/10/2022 9:22 SMPL
N3915-03	YBB54	1318	0.325 -	SFAM01.1	PPB	8/10/2022 9:24 SMPL
N3915-04	YBB57	628	0.1633 -	SFAM01.1	PPB	8/10/2022 9:26 SMPL
N3915-05	YBB58	447	0.1209 -	SFAM01.1	PPB	8/10/2022 9:29 SMPL
N3915-06	YBB59	942	0.2369 -	SFAM01.1	PPB	8/10/2022 9:31 SMPL
N3915-07	YBB64	396	0.1089 -	SFAM01.1	PPB	8/10/2022 9:33 SMPL
N3915-08	YBBC6	615	0.1602 -	SFAM01.1	PPB	8/10/2022 9:35 SMPL
N3915-09	YBBC7	1200	0.2973 -	SFAM01.1	PPB	8/10/2022 9:38 SMPL
N3915-10	YBBC1	3848	0.9179 -	SFAM01.1	PPB	8/10/2022 9:40 SMPL
N3915-11	YBBC2	24497	5.7568 -	SFAM01.1	PPB	8/10/2022 9:42 SMPL
N3915-12	YBBC3	994	0.2491 -	SFAM01.1	PPB	8/10/2022 9:44 SMPL
N3915-13	YBBC4	1338	0.3297 -	SFAM01.1	PPB	8/10/2022 9:47 SMPL
CCV098	CCV098	19454	4.575 -	SFAM01.1	PPB	8/10/2022 9:52 SMPL
CCB098	CCB098	-73	-0.001 -	SFAM01.1	PPB	8/10/2022 9:54 SMPL
N3915-14	YBBC5	1341	0.3304 -	SFAM01.1	PPB	8/10/2022 9:56 SMPL
N3915-15	YBBC5D	1215	0.3008 -	SFAM01.1	PPB	8/10/2022 9:59 SMPL
N3915-16	YBBC5S	10909	2.5725 -	SFAM01.1	PPB	8/10/2022 10:01 SMPL
PB146891BL	PBS891	-33	0.0084 -	SFAM01.1	PPB	8/10/2022 10:03 SMPL
N3917-01	YBB97	1888	0.4586 -	SFAM01.1	PPB	8/10/2022 10:05 SMPL
N3917-02	YBB98	2161	0.5225 -	SFAM01.1	PPB	8/10/2022 10:08 SMPL
N3917-03	YBB99	1800	0.4379 -	SFAM01.1	PPB	8/10/2022 10:10 SMPL
N3917-04	YBB99D	1701	0.4147 -	SFAM01.1	PPB	8/10/2022 10:12 SMPL
N3917-05	YBB99S	11955	2.8177 -	SFAM01.1	PPB	8/10/2022 10:14 SMPL
N3917-06	YBBA0	1015	0.254 -	SFAM01.1	PPB	8/10/2022 10:17 SMPL
N3917-07	YBBA1	1018	0.2547 -	SFAM01.1	PPB	8/10/2022 10:19 SMPL
N3917-08	YBBA2	399	0.1096 -	SFAM01.1	PPB	8/10/2022 10:21 SMPL
N3917-09	YBBA3	1453	0.3566 -	SFAM01.1	PPB	8/10/2022 10:24 SMPL
N3917-10	YBBA4	1475	0.3618 -	SFAM01.1	PPB	8/10/2022 10:26 SMPL
N3917-11	YBBA5	1096	0.273 -	SFAM01.1	PPB	8/10/2022 10:28 SMPL
PB146888BL	PBS888	-12	0.0133 -	SFAM01.1	PPB	8/10/2022 10:30 SMPL
CCV099	CCV099	20750	4.8787 -	SFAM01.1	PPB	8/10/2022 10:35 SMPL
CCB099	CCB099	-62	0.0016 -	SFAM01.1	PPB	8/10/2022 10:37 SMPL