



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Iwona

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: Sohil

Run Number: LB121440

Reagent/Standard	Lot/Log #
Residual chlorine LOD, 0.05PPM	WP98307
Residual chlorine ICV-LCS, 0.4PPM	WP98306
Chlorine Calibration std, 0.1ppm	WP98301
Chlorine Calibration std, 0.2ppm	WP98302
Chlorine Calibration std, 0.8ppm	WP98304
Chlorine Calibration std, 0.0ppm	WP98300
Chlorine Calibration std, 1.6ppm	WP98305
Residual Chlorine LOQ 0.1ppm	WP98308
Residual Chlorine Calibration and CCV std, 0	WP98303
Total Chlorine Powder Pillows	W2759

Intercept: 0.0074

Slope: 1.0082

Regression: 0.999956

Seq	Lab ID	True Val (mg/l)	DF	Initial Reading	Final Reading	Difference	Result (mg/l)	%D	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	-0.01		08/10/2022	13:40
2	CAL2	0.1	1	0.000	0.110	0.110	0.10	-2	08/10/2022	13:43
3	CAL3	0.2	1	0.000	0.210	0.210	0.20	-0.5	08/10/2022	13:46
4	CAL4	0.4	1	0.000	0.420	0.420	0.41	-2.2	08/10/2022	13:49
5	CAL5	0.8	1	0.000	0.810	0.810	0.80	0.5	08/10/2022	13:52
6	CAL6	1.6	1	0.000	1.620	1.620	1.60	0.1	08/10/2022	13:55



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Iwona

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: Sohil

Run Number: LB121440

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.4100	0.4100	0.4010	08/10/2022	13:58
2	ICB				1	0.0000	0.0000	0.0000	-0.0050	08/10/2022	14:01
3	CCV1			0.4	1	0.0000	0.3900	0.3900	0.3810	08/10/2022	14:04
4	CCB1				1	0.0000	0.0100	0.0100	0.0050	08/10/2022	14:07
5	LB121440BL	50	50		1	0.0000	0.0100	0.0100	0.0050	08/10/2022	14:10
6	LB121440BS	50	50	0.4	1	0.0000	0.4000	0.4000	0.3910	08/10/2022	14:13
7	N3980-07	50	50		1	0.0000	0.0500	0.0500	0.0450	08/10/2022	14:16
8	N3980-08	50	50		1	0.0000	0.1100	0.1100	0.1040	08/10/2022	14:19
9	CCV2			0.4	1	0.0000	0.4000	0.4000	0.3910	08/10/2022	14:22
10	CCB2				1	0.0000	0.0000	0.0000	-0.0050	08/10/2022	14:25