



TOTAL VOLATILE SOLIDS SM2540 G

TEMP1 IN: 104 °C 08/10/2022 17:30 TEMP1 OUT: 103 °C 08/11/2022 08:00
TEMP2 IN: 104 °C 08/11/2022 08:30 TEMP2 OUT: 104 °C 08/11/2022 10:00
TEMP3 IN: 540 °C 08/11/2022 10:30 TEMP3 OUT: 540 °C 08/11/2022 12:00
TEMP4 IN: 560 °C 08/11/2022 12:30 TEMP4 OUT: 560 °C 08/11/2022 14:00

Run Number: LB121447

SUPERVISOR: Sohil

ANALYST: JIGNESH

BalanceID: WC SC-6

OvenID: WC OVEN-1

Dish #	Lab ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish + SampleWt Drying @103-@105°C (g)	Final Dish + SampleWt Drying @103-@105°C (g)	Dish + Samplewt Drying @550 (±50) °C (g)	Final Dish + Samplewt Drying @550 (±50) °C (g)	Weight Diff (g)	Result (mg/L)
1	LB121447BL	95.6331	95.6331	100	95.6331	95.6331	95.6331	95.6331	0.0000	0
2	N4154-01	89.1405	89.1405	100	89.2191	89.2191	89.1975	89.1975	0.0216	216
3	N4154-01DUP	97.2583	97.2583	100	97.3370	97.337	97.3147	97.3147	0.0223	223
4	N4154-03	89.6474	89.6474	100	89.7265	89.7265	89.7089	89.7089	0.0176	176

A = Sample Volume (ml)

B = Final Dish + Samplewt Drying @550 (±50) °C (g)

C = Final Dish + SampleWt Drying @103-@105 °C (g)

D = Weight (g)

E = Final Empty Dish Weight (g)

F = Final Dish + SampleWt Drying @103-@105 °C (g)

Weight D = C - B

Result % = $\frac{D}{A} \star 1000000$

WORKLIST(Hardcopy Internal Chain)

VB 2447 VB 121647

WorkList Name : tvs n4154

WorkList ID : 162181

Department : Wet-Chemistry

Date : 08-10-2022 16:53:40

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/15/2022	Water	N4154-01	TVS	Cool 4 deg C	PSEG03	J11	WV-TOTECOMP-1	08/10/2022	SM2540 G
08/15/2022	Water	N4154-03	TVS	Cool 4 deg C	PSEG03	J11	WV-TOTECOMP-2	08/10/2022	SM2540 G

Date/Time 08-10-22 17:00
 Raw Sample Received by: JTC(m)
 Raw Sample Relinquished by: JTC(m)

Date/Time 08-10-22 18:30
 Raw Sample Received by: JTC(m)
 Raw Sample Relinquished by: JTC(m)