



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB121448

Analyst By : JIGNESH

Supervisor Review By : Sohil

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	08/10/2022	17:15
2	CAL2	1	Water	NA	NA	20.6	7.00	08/10/2022	17:16
3	CAL3	1	Water	NA	NA	20.7	10.03	08/10/2022	17:19
4	ICV	1	Water	NA	NA	20.7	6.99	08/10/2022	17:22
5	CCV1	1	Water	NA	NA	20.8	2.01	08/10/2022	17:25
6	N4154-01	1	Water	NA	NA	24.3	8.66	08/10/2022	17:26
7	N4154-01DUP	1	Water	NA	NA	24.3	8.67	08/10/2022	17:27
8	N4154-03	1	Water	NA	NA	24.3	7.11	08/10/2022	17:30
9	CCV2	1	Water	NA	NA	20.7	12.03	08/10/2022	17:32

UB 121448

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph n4154 WorkList ID : 162182 Department : Wet-Chemistry Date : 08-10-2022 17:00:01

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/10/2022	Water	N4154-01	pH	Cool 4 deg C	PSEG03	J11	WW-TOTECOMP-1	08/10/2022	9040C
08/10/2022	Water	N4154-03	pH	Cool 4 deg C	PSEG03	J11	WW-TOTECOMP-2	08/10/2022	9040C

Date/Time 08.10.22 17:05
Raw Sample Received by: JN (QCC)
Raw Sample Relinquished by: JT (SM)

Date/Time 08.10.22 18:30
Raw Sample Received by: JT (SM)
Raw Sample Relinquished by: M (a2c)