



Analytical Summary Report

Analysis Method: 9045D

Parameter: Corrosivity

Run Number: LB121450

BalanceID: WC SC-4

Analyst By : JIGNESH

Supervisor Review By : Sohil

Slope : 99.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.9	4.01	08/10/2022	13:30
2	CAL2	1	Water	NA	NA	20.9	7.00	08/10/2022	13:31
3	CAL3	1	Water	NA	NA	20.9	10.03	08/10/2022	13:33
4	ICV	1	Water	NA	NA	20.9	6.98	08/10/2022	13:35
5	CCV1	1	Water	NA	NA	20.9	2.01	08/10/2022	13:36
6	N4120-07	1	Solid	20.02	20	23.8	8.86	08/10/2022	13:37
7	N4120-07DUP	1	Solid	20.03	20	23.8	8.87	08/10/2022	13:38
8	N4130-02	1	Solid	20.01	20	23.5	7.16	08/10/2022	13:40
9	N4130-04	1	Solid	20.03	20	23.1	6.76	08/10/2022	13:42
10	N4130-06	1	Solid	20.05	20	23.1	8.25	08/10/2022	13:45
11	N4130-08	1	Solid	20.02	20	22.9	7.10	08/10/2022	13:46
12	N4130-10	1	Solid	20.04	20	22.2	7.38	08/10/2022	13:49
13	N4130-12	1	Solid	20.01	20	23.4	7.36	08/10/2022	13:50
14	N4130-14	1	Solid	20.03	20	21.9	7.67	08/10/2022	14:00
15	N4130-16	1	Solid	20.02	20	23.7	6.67	08/10/2022	14:02
16	CCV2	1	Water	NA	NA	20.8	12.03	08/10/2022	14:10
17	N4136-02	1	Solid	20.02	20	23.4	12.36	08/10/2022	14:15
18	N4136-04	1	Solid	20.01	20	23.7	8.69	08/10/2022	14:20
19	N4141-01	1	Solid	20.01	20	23.9	5.87	08/10/2022	14:24
20	CCV3	1	Water	NA	NA	20.9	2.01	08/10/2022	14:25

121450

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity n4141

WorkList ID : 162168

Department : Wet-Chemistry

Date : 08-10-2022 12:36:07

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/09/2022	Solid	N4120-07	Corrosivity	Cool 4 deg C	GUAR01	J11	BLDG-19C	08/09/2022	9045D
08/08/2022	Solid	N4130-02	Corrosivity	Cool 4 deg C	loui01	J11	SB02	08/08/2022	9045D
08/08/2022	Solid	N4130-04	Corrosivity	Cool 4 deg C	loui01	J11	SB03	08/08/2022	9045D
08/08/2022	Solid	N4130-06	Corrosivity	Cool 4 deg C	loui01	J11	SB05	08/08/2022	9045D
08/09/2022	Solid	N4130-08	Corrosivity	Cool 4 deg C	loui01	J11	SB06	08/09/2022	9045D
08/09/2022	Solid	N4130-10	Corrosivity	Cool 4 deg C	loui01	J11	SB07	08/09/2022	9045D
08/09/2022	Solid	N4130-12	Corrosivity	Cool 4 deg C	loui01	J11	SB09	08/09/2022	9045D
08/09/2022	Solid	N4130-14	Corrosivity	Cool 4 deg C	loui01	J11	SB04	08/09/2022	9045D
08/09/2022	Solid	N4130-16	Corrosivity	Cool 4 deg C	loui01	J11	SB01	08/09/2022	9045D
08/09/2022	Solid	N4136-02	Corrosivity	Cool 4 deg C	DVIR01	J11	WC-4	08/09/2022	9045D
08/09/2022	Solid	N4136-04	Corrosivity	Cool 4 deg C	DVIR01	J11	WC-5	08/09/2022	9045D
08/10/2022	Solid	N4141-01	Corrosivity	Cool 4 deg C	ARAM01	J11	SOILD-WASTE	08/10/2022	9045D

Date/Time 08.10.22 13:00
 Raw Sample Received by: JT (Sam)
 Raw Sample Relinquished by: JT (Sam)

Date/Time 08.10.22 16:00
 Raw Sample Received by: JT (Sam)
 Raw Sample Relinquished by: JT (Sam)