

LB121456


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

40775



A= 0.0000e+000

B= 2.4202e-004 slope

C= -2.6374e-002 $\gamma = 1.01 \text{ intercept}$

Rho= 0.9995799

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	A/D
0.00	0.000	0.008	0.008	141	0.000	141					-
0.05	0.050					185	0				-13
0.20	0.200	0.175	-0.025	831	0.0 %	831					-3
2.50	2.500	2.427	-0.073	10138	0.0 %	10138					1
5.00	5.000	5.066	0.066	21039	0.0 %	21039					2
7.50	7.500	7.683	0.183	31853	0.0 %	31853					-2
10.0	10.000	9.842	-0.158	40775	0.0 %	40775					-2

LB121456

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	141	-		0 SFAM01.1	PPB	8/11/2022 10:53	Std
	0.2 S01	831	-		0.2 SFAM01.1	PPB	8/11/2022 10:55	Std
	2.5 S02	10138	-		2.5 SFAM01.1	PPB	8/11/2022 10:58	Std
	5 S03	21039	-		5 SFAM01.1	PPB	8/11/2022 11:00	Std
	7.5 S04	31853	-		7.5 SFAM01.1	PPB	8/11/2022 11:02	Std
	10 S05	40775	-		10 SFAM01.1	PPB	8/11/2022 11:05	Std
ICV080	ICV080	17691	4.2552	-	SFAM01.1	PPB	8/11/2022 11:07	SMPL
ICB080	ICB080	349	0.0581	-	SFAM01.1	PPB	8/11/2022 11:10	SMPL
CCV009	CCV009	20781	5.0031	-	SFAM01.1	PPB	8/11/2022 11:12	SMPL
CCB009	CCB009	347	0.0576	-	SFAM01.1	PPB	8/11/2022 11:14	SMPL
PB146923BL	PBW923	428	0.0772	-	SFAM01.1	PPB	8/11/2022 11:16	SMPL
N3787-01	YBAY3	460	0.085	-	SFAM01.1	PPB	8/11/2022 11:19	SMPL
N3787-02	YBAY4	462	0.0854	-	SFAM01.1	PPB	8/11/2022 11:21	SMPL
N3787-03	YBAY5	465	0.0862	-	SFAM01.1	PPB	8/11/2022 11:23	SMPL
N3787-04	YBAY6	466	0.0864	-	SFAM01.1	PPB	8/11/2022 11:26	SMPL
N3787-05	YBAY7	443	0.0808	-	SFAM01.1	PPB	8/11/2022 11:28	SMPL
N3787-06	YBB00	468	0.0869	-	SFAM01.1	PPB	8/11/2022 11:30	SMPL
N3787-07	YBAX9	457	0.0842	-	SFAM01.1	PPB	8/11/2022 11:32	SMPL
N3787-08	YBAY0	461	0.0852	-	SFAM01.1	PPB	8/11/2022 11:35	SMPL
N3787-09	YBAY1	450	0.0825	-	SFAM01.1	PPB	8/11/2022 11:37	SMPL
N3787-10	YBAY2	577	0.1133	-	SFAM01.1	PPB	8/11/2022 11:39	SMPL
N3787-11	YBAY8	1426	0.3187	-	SFAM01.1	PPB	8/11/2022 11:41	SMPL
N3787-12	YBAZ6	1561	0.3514	-	SFAM01.1	PPB	8/11/2022 11:44	SMPL
N3787-13	YBAZ7	1286	0.2849	-	SFAM01.1	PPB	8/11/2022 11:46	SMPL
N3787-14	YBAX5	560	0.1092	-	SFAM01.1	PPB	8/11/2022 11:48	SMPL
N3787-15	YBAX6	469	0.0871	-	SFAM01.1	PPB	8/11/2022 11:50	SMPL
N3787-16	YBAX7	466	0.0864	-	SFAM01.1	PPB	8/11/2022 11:53	SMPL
N3787-17	YBAX8	485	0.091	-	SFAM01.1	PPB	8/11/2022 11:55	SMPL
N3787-18	YBAZ4	480	0.0898	-	SFAM01.1	PPB	8/11/2022 11:57	SMPL
N3787-19	YBAZ5	462	0.0854	-	SFAM01.1	PPB	8/11/2022 12:00	SMPL
CCV010	CCV010	20420	4.9157	-	SFAM01.1	PPB	8/11/2022 12:02	SMPL
CCB010	CCB010	325	0.0523	-	SFAM01.1	PPB	8/11/2022 12:04	SMPL
N3787-20	YBAZ8	297	0.0455	-	SFAM01.1	PPB	8/11/2022 12:06	SMPL
N3787-21	YBAZ8D	304	0.0472	-	SFAM01.1	PPB	8/11/2022 12:09	SMPL
N3787-22	YBAZ8S	4764	1.1266	-	SFAM01.1	PPB	8/11/2022 12:11	SMPL
PB146924BL	PBW924	400	0.0704	-	SFAM01.1	PPB	8/11/2022 12:13	SMPL
N3788-01	YBB27	454	0.0835	-	SFAM01.1	PPB	8/11/2022 12:15	SMPL
N3788-02	YBB28	89	-0.0048	-	SFAM01.1	PPB	8/11/2022 12:18	SMPL
N3788-03	YBB29	104	-0.0012	-	SFAM01.1	PPB	8/11/2022 12:20	SMPL
N3788-04	YBB30	224	0.0278	-	SFAM01.1	PPB	8/11/2022 12:22	SMPL
N3788-05	YBB31	398	0.0699	-	SFAM01.1	PPB	8/11/2022 12:24	SMPL
N3788-06	YBB44	394	0.069	-	SFAM01.1	PPB	8/11/2022 12:27	SMPL
N3788-07	YBB23	406	0.0719	-	SFAM01.1	PPB	8/11/2022 12:29	SMPL
N3788-08	YBB24	401	0.0707	-	SFAM01.1	PPB	8/11/2022 12:31	SMPL
N3788-09	YBB25	402	0.0709	-	SFAM01.1	PPB	8/11/2022 12:34	SMPL
N3788-10	YBB26	421	0.0755	-	SFAM01.1	PPB	8/11/2022 12:36	SMPL

LB121456

INSTRUMENT ID : CV1

N3788-11	YBB32	439	0.0799 -	SFAM01.1	PPB	8/11/2022 12:38 SMPL
N3788-12	YBB40	399	0.0702 -	SFAM01.1	PPB	8/11/2022 12:40 SMPL
N3788-13	YBB41	393	0.0687 -	SFAM01.1	PPB	8/11/2022 12:43 SMPL
N3788-14	YBB19	402	0.0709 -	SFAM01.1	PPB	8/11/2022 12:45 SMPL
N3788-15	YBB20	403	0.0712 -	SFAM01.1	PPB	8/11/2022 12:47 SMPL
N3788-16	YBB21	398	0.0699 -	SFAM01.1	PPB	8/11/2022 12:49 SMPL
CCV011	CCV011	19787	4.7625 -	SFAM01.1	PPB	8/11/2022 12:52 SMPL
CCB011	CCB011	-38	-0.0356 -	SFAM01.1	PPB	8/11/2022 12:54 SMPL
N3788-17	YBB22	308	0.0482 -	SFAM01.1	PPB	8/11/2022 12:56 SMPL
N3788-18	YBB38	233	0.03 -	SFAM01.1	PPB	8/11/2022 12:59 SMPL
N3788-19	YBB39	294	0.0448 -	SFAM01.1	PPB	8/11/2022 13:01 SMPL
N3788-20	YBB42	310	0.0487 -	SFAM01.1	PPB	8/11/2022 13:03 SMPL
N3788-21	YBB42D	291	0.0441 -	SFAM01.1	PPB	8/11/2022 13:05 SMPL
N3788-22	YBB42S	4182	0.9858 -	SFAM01.1	PPB	8/11/2022 13:08 SMPL
PB146925BL	PBW925	307	0.0479 -	SFAM01.1	PPB	8/11/2022 13:10 SMPL
N3852-01	YBBA7	323	0.0518 -	SFAM01.1	PPB	8/11/2022 13:12 SMPL
N3852-02	YBBB1	403	0.0712 -	SFAM01.1	PPB	8/11/2022 13:14 SMPL
N3852-03	YBB17	308	0.0482 -	SFAM01.1	PPB	8/11/2022 13:17 SMPL
N3852-04	YBB33	299	0.046 -	SFAM01.1	PPB	8/11/2022 13:19 SMPL
N3852-05	YBB35	298	0.0457 -	SFAM01.1	PPB	8/11/2022 13:21 SMPL
N3852-06	YBB36	296	0.0453 -	SFAM01.1	PPB	8/11/2022 13:24 SMPL
N3852-07	YBB37	288	0.0433 -	SFAM01.1	PPB	8/11/2022 13:26 SMPL
N3852-08	YBB45	297	0.0455 -	SFAM01.1	PPB	8/11/2022 13:28 SMPL
N3852-09	YBB46	330	0.0535 -	SFAM01.1	PPB	8/11/2022 13:30 SMPL
N3852-10	YBB43	356	0.0598 -	SFAM01.1	PPB	8/11/2022 13:33 SMPL
N3852-11	YBBB5	337	0.0552 -	SFAM01.1	PPB	8/11/2022 13:35 SMPL
N3852-12	YBBB8	338	0.0554 -	SFAM01.1	PPB	8/11/2022 13:37 SMPL
N3852-13	YBB04	348	0.0578 -	SFAM01.1	PPB	8/11/2022 13:39 SMPL
CCV012	CCV012	18661	4.49 -	SFAM01.1	PPB	8/11/2022 13:42 SMPL
CCB012	CCB012	655	0.1321 -	SFAM01.1	PPB	8/11/2022 13:44 SMPL
N3852-14	YBB05	358	0.0603 -	SFAM01.1	PPB	8/11/2022 13:46 SMPL
N3852-15	YBB10	334	0.0545 -	SFAM01.1	PPB	8/11/2022 13:49 SMPL
N3852-16	YBB13	350	0.0583 -	SFAM01.1	PPB	8/11/2022 13:51 SMPL
N3852-17	YBB13D	370	0.0632 -	SFAM01.1	PPB	8/11/2022 13:53 SMPL
N3852-18	YBB13S	4096	0.9649 -	SFAM01.1	PPB	8/11/2022 13:55 SMPL
N3852-19	YBB14	380	0.0656 -	SFAM01.1	PPB	8/11/2022 13:58 SMPL
N3852-20	YBB06	463	0.0857 -	SFAM01.1	PPB	8/11/2022 14:00 SMPL
N3852-21	YBB07	426	0.0767 -	SFAM01.1	PPB	8/11/2022 14:02 SMPL
CCV013	CCV013	21921	5.279 -	SFAM01.1	PPB	8/11/2022 14:08 SMPL
CCB013	CCB013	291	0.0441 -	SFAM01.1	PPB	8/11/2022 14:11 SMPL