

LB121464


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

48584



A= 0.0000e+000

B= 2.0732e-004

C= -2.6090e-002

Rho= 0.9996973

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	910 D
0.0	0.000	-0.018	-0.018	37	0.000	37					-
0.2	0.200	0.168	-0.032	938	0.0 %	938					-16
2.5	2.500	2.543	0.043	12393	0.0 %	12393					2
5.0	5.000	5.127	0.127	24857	0.0 %	24857					3
7.5	7.500	7.333	-0.167	35496	0.0 %	35496					-2
10.0	10.000	10.046	0.046	48584	0.0 %	48584					0

LB121464 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 11 Aug 2022 15:54:39

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	37	-	0.0000	7471B	PPB	11 Aug 2022 16:00:59	S	Std
0.2 - 1	50.2	938	-	0.2000	7471B	PPB	11 Aug 2022 16:03:16	S	Std
2.5 - 1	52.5	12393	-	2.5000	7471B	PPB	11 Aug 2022 16:05:34	S	Std
5.0 - 1	55	24857	-	5.0000	7471B	PPB	11 Aug 2022 16:07:51	S	Std
7.5 - 1	57.5	35496	-	7.5000	7471B	PPB	11 Aug 2022 16:10:08	S	Std
10.0 - 1	59	48584	-	10.0000	7471B	PPB	11 Aug 2022 16:12:25	S	Std
ICV81 - 1	ICV81	19003	3.9137	-7471B	PPB	PPB	11 Aug 2022 16:15:17	U	SMPL
ICB81 - 1	ICB81	-32	-0.0327	-7471B	PPB	PPB	11 Aug 2022 16:17:32	U	SMPL
CCV14 - 1	CCV14	24902	5.1366	-7471B	PPB	PPB	11 Aug 2022 16:19:49	U	SMPL
CCB14 - 1	CCB14	-42	-0.0348	-7471B	PPB	PPB	11 Aug 2022 16:22:04	U	SMPL
CRA - 1	CRA	919	0.1644	-7471B	PPB	PPB	11 Aug 2022 16:24:21	U	SMPL
HighStd - 1	HighStd	51624	10.6767	-7471B	PPB	PPB	11 Aug 2022 16:26:36	U	SMPL
ChkStd - 1	ChkStd	37040	7.6531	-7471B	PPB	PPB	11 Aug 2022 16:30:44	U	SMPL
PB146941BL - 1	PBS	-71	-0.0408	-7471B	PPB	PPB	11 Aug 2022 16:32:59	U	SMPL
PB146941BS - 1	LCSS	20265	4.1753	-7471B	PPB	PPB	11 Aug 2022 16:35:17	U	SMPL
N4120-06 - 1	BLDG-19C	56948	11.7805	-7471B	PPB	PPB	11 Aug 2022 16:37:33	U	SMPL
N4136-01 - 1	WC-4	1498	0.2845	-7471B	PPB	PPB	11 Aug 2022 16:39:51	U	SMPL
N4136-03 - 1	WC-5	242	0.0241	-7471B	PPB	PPB	11 Aug 2022 16:42:10	U	SMPL
N4143-01 - 1	RB-21204	20082	4.1374	-7471B	PPB	PPB	11 Aug 2022 16:44:27	U	SMPL
N4143-03 - 1	CHRT-24051	1400	0.2642	-7471B	PPB	PPB	11 Aug 2022 16:46:42	U	SMPL
CCV15 - 1	CCV15	23078	4.7585	-7471B	PPB	PPB	11 Aug 2022 16:48:59	U	SMPL
CCB15 - 1	CCB15	-142	-0.0555	-7471B	PPB	PPB	11 Aug 2022 16:51:14	U	SMPL
N4143-05 - 1	CHRT-26414	4103	0.8246	-7471B	PPB	PPB	11 Aug 2022 16:53:32	U	SMPL
N4148-01 - 1	290533	603	0.0989	-7471B	PPB	PPB	11 Aug 2022 16:55:48	U	SMPL
N4150-01 - 1	CL-01-081022	5129	1.0373	-7471B	PPB	PPB	11 Aug 2022 16:58:04	U	SMPL
N4152-01 - 1	EO-02-081022	1580	0.3015	-7471B	PPB	PPB	11 Aug 2022 17:00:20	U	SMPL
N4152-03 - 1	EO-03-081022	2971	0.5899	-7471B	PPB	PPB	11 Aug 2022 17:02:35	U	SMPL
N4153-01 - 1	TR-04-081022	2508	0.4939	-7471B	PPB	PPB	11 Aug 2022 17:04:52	U	SMPL
N4153-01DUP - 1	TR-04-081022DUP	2462	0.4843	-7471B	PPB	PPB	11 Aug 2022 17:07:09	U	SMPL
N4153-01MS - 1	TR-04-081022MS	21401	4.4108	-7471B	PPB	PPB	11 Aug 2022 17:09:25	U	SMPL
N4153-01MSD - 1	TR-04-081022MSD	20862	4.2991	-7471B	PPB	PPB	11 Aug 2022 17:11:42	U	SMPL
CCV16 - 1	CCV16	24457	5.0444	-7471B	PPB	PPB	11 Aug 2022 17:14:00	U	SMPL
CCB16 - 1	CCB16	-58	-0.0381	-7471B	PPB	PPB	11 Aug 2022 17:16:17	U	SMPL
N4153-01LX5 - 1		509	0.0794	-7471B	PPB	PPB	11 Aug 2022 17:18:34	U	SMPL
N4153-01A - 1		19699	4.0579	-7471B	PPB	PPB	11 Aug 2022 17:20:49	U	SMPL
N4119-01DLX5 - 1		14020	2.8806	-7471B	PPB	PPB	11 Aug 2022 17:23:05	U	SMPL
N4120-06DLX2 - 1		25326	5.2246	-7471B	PPB	PPB	11 Aug 2022 17:25:22	U	SMPL
CCV16 - 1	CCV16	24036	4.9571	-7471B	PPB	PPB	11 Aug 2022 17:27:40	U	SMPL
CCB16 - 1	CCB16	-60	-0.0385	-7471B	PPB	PPB	11 Aug 2022 17:29:58	U	SMPL