



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Sohil

Run Number: LB121469

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	08/11/2022	16:00
2	CAL2	1	Water	NA	NA	20.8	7.00	08/11/2022	16:01
3	CAL3	1	Water	NA	NA	20.8	10.03	08/11/2022	16:02
4	ICV	1	Water	NA	NA	20.7	6.99	08/11/2022	16:05
5	CCV1	1	Water	NA	NA	20.9	2.01	08/11/2022	16:07
6	N4167-01	1	Solid	20.03	20	24.4	5.26	08/11/2022	16:20
7	N4167-01DUP	1	Solid	20.04	20	24.5	5.27	08/11/2022	16:21
8	CCV2	1	Water	NA	NA	20.8	12.03	08/11/2022	16:25

VB 121469

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph n4167

WorkList ID : 162214

Department : Wet-Chemistry

Date : 08-11-2022 13:43:27

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/10/2022	Solid	N4167-01	pH	Cool 4 deg C	TETR06	J11	RE137-CARBON-20220810	08/10/2022	9045D

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

08.11.22 13:50

JD (WC)

CD San

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

08.11.22 15:30

CP San

JD (WC)