

LB1214177


7470A

INSTRUMENT ID: EV3

Linear

 μ Abs.:

38904



A= 0.0000e+000

B= 2.5603e-004

C= 3.0747e-002

Rho= 0.9999451

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.021	0.021	-38	0.000	-38					
0.2	0.200	0.229	0.029	776	0.0 %	776					
2.5	2.500	2.440	-0.060	9412	0.0 %	9412					
5.0	5.000	4.965	-0.035	19272	0.0 %	19272					
7.5	7.500	7.553	0.053	29381	0.0 %	29381					
10.0	10.000	9.991	-0.009	38904	0.0 %	38904					

15
-2
-1
1
0

LB121477 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 12 Aug 2022 10:32:43

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-38	-	0.0000	7470A	PPB	12 Aug 2022 10:38:26	S	Std
0.2 - 1	50.2	776	-	0.2000	7470A	PPB	12 Aug 2022 10:40:44	S	Std
2.5 - 1	52.5	9412	-	2.5000	7470A	PPB	12 Aug 2022 10:43:00	S	Std
5.0 - 1	55	19272	-	5.0000	7470A	PPB	12 Aug 2022 10:45:17	S	Std
7.5 - 1	53.5	29381	-	7.5000	7470A	PPB	12 Aug 2022 10:47:33	S	Std
10.0 - 1	50	38904	-	10.0000	7470A	PPB	12 Aug 2022 10:56:39	S	Std
ICV82 - 1	ICV82	14746	3.8061	-	7470A	PPB	12 Aug 2022 10:59:30	U	SMPL
ICB82 - 1	ICB82	2	0.0313	-	7470A	PPB	12 Aug 2022 11:01:44	U	SMPL
CCV17 - 1	CCV17	18773	4.8371	-	7470A	PPB	12 Aug 2022 11:06:29	U	SMPL
CCB17 - 1	CCB17	-29	0.0233	-	7470A	PPB	12 Aug 2022 11:08:44	U	SMPL
CRA - 1	CRA	686	0.2064	-	7470A	PPB	12 Aug 2022 11:11:00	U	SMPL
HighStd - 1	HighStd	38382	9.8575	-	7470A	PPB	12 Aug 2022 11:13:14	U	SMPL
ChkStd - 1	ChkStd	25582	6.5804	-	7470A	PPB	12 Aug 2022 11:15:30	U	SMPL
PB146942BL - 1	PBW	-43	0.0197	-	7470A	PPB	12 Aug 2022 11:17:48	U	SMPL
PB146942BS - 1	LCSW	16069	4.1448	-	7470A	PPB	12 Aug 2022 11:22:45	U	SMPL
N4120-07 - 1	BLDG-19C	-19	0.0259	-	7470A	PPB	12 Aug 2022 11:25:00	U	SMPL
N4130-02 - 1	SB02	12	0.0338	-	7470A	PPB	12 Aug 2022 11:27:17	U	SMPL
N4130-04 - 1	SB03	26	0.0374	-	7470A	PPB	12 Aug 2022 11:29:33	U	SMPL
N4130-06 - 1	SB05	49	0.0433	-	7470A	PPB	12 Aug 2022 11:31:50	U	SMPL
N4130-08 - 1	SB06	74	0.0497	-	7470A	PPB	12 Aug 2022 11:34:06	U	SMPL
CCV18 - 1	CCV18	18185	4.6866	-	7470A	PPB	12 Aug 2022 11:36:22	U	SMPL
CCB18 - 1	CCB18	-13	0.0274	-	7470A	PPB	12 Aug 2022 11:38:37	U	SMPL
N4130-10 - 1	SB07	45	0.0423	-	7470A	PPB	12 Aug 2022 11:40:53	U	SMPL
N4130-12 - 1	SB09	30	0.0384	-	7470A	PPB	12 Aug 2022 11:43:08	U	SMPL
N4130-14 - 1	SB04	29	0.0382	-	7470A	PPB	12 Aug 2022 11:45:24	U	SMPL
N4130-16 - 1	SB01	41	0.0412	-	7470A	PPB	12 Aug 2022 11:47:40	U	SMPL
N4136-02 - 1	WC-4	36	0.0400	-	7470A	PPB	12 Aug 2022 11:49:56	U	SMPL
N4136-04 - 1	WC-5	32	0.0389	-	7470A	PPB	12 Aug 2022 11:52:12	U	SMPL
N4139-02 - 1	PIG-SWAB-1-7	42	0.0415	-	7470A	PPB	12 Aug 2022 11:54:29	U	SMPL
N4141-01 - 1	SOILD-WASTE	37	0.0402	-	7470A	PPB	12 Aug 2022 11:56:47	U	SMPL
N4141-01DUP - 1	SOILD-WASTEDUP	32	0.0389	-	7470A	PPB	12 Aug 2022 11:59:03	U	SMPL
N4141-01MS - 1	SOILD-WASTEMS	13855	3.5780	-	7470A	PPB	12 Aug 2022 12:01:19	U	SMPL
CCV19 - 1	CCV19	19941	5.1362	-	7470A	PPB	12 Aug 2022 12:03:35	U	SMPL
CCB19 - 1	CCB19	-17	0.0264	-	7470A	PPB	12 Aug 2022 12:05:51	U	SMPL
N4141-01MSD - 1	SOILD-WASTEMSD	12915	3.3373	-	7470A	PPB	12 Aug 2022 12:08:07	U	SMPL
PB146972BL - 1	PBW	3	0.0315	-	7470A	PPB	12 Aug 2022 12:10:23	U	SMPL
PB146972BS - 1	LCSW	15482	3.9945	-	7470A	PPB	12 Aug 2022 12:12:38	U	SMPL
N4154-01 - 1	WV-TOTECOMP-1	240	0.0922	-	7470A	PPB	12 Aug 2022 12:14:54	U	SMPL
N4154-03 - 1	WV-TOTECOMP-2	655	0.1984	-	7470A	PPB	12 Aug 2022 12:17:09	U	SMPL
N4154-03DUP - 1	WV-TOTECOMP-2DUP	635	0.1933	-	7470A	PPB	12 Aug 2022 12:19:26	U	SMPL
N4154-03MS - 1	WV-TOTECOMP-2MS	14232	3.6745	-	7470A	PPB	12 Aug 2022 12:21:42	U	SMPL
N4154-03MSD - 1	WV-TOTECOMP-2MSD	13385	3.4577	-	7470A	PPB	12 Aug 2022 12:23:59	U	SMPL
CCV20 - 1	CCV20	19423	5.0035	-	7470A	PPB	12 Aug 2022 12:26:01	U	SMPL
CCB20 - 1	CCB20	-15	0.0269	-	7470A	PPB	12 Aug 2022 12:30:16	U	SMPL
PB146911TB - 1	PB146911TB	50	0.0435	-	7470A	PPB	12 Aug 2022 12:32:32	U	SMPL
N4141-01LX5 - 1		30	0.0384	-	7470A	PPB	12 Aug 2022 12:34:48	U	SMPL
N4141-01A - 1		16408	4.2316	-	7470A	PPB	12 Aug 2022 12:46:47	U	SMPL
N4154-03LX5 - 1		63	0.0469	-	7470A	PPB	12 Aug 2022 12:49:03	U	SMPL
N4154-03A - 1		15207	3.9241	-	7470A	PPB	12 Aug 2022 12:51:19	U	SMPL
CCV21 - 1	CCV21	18669	4.8105	-	7470A	PPB	12 Aug 2022 12:53:35	U	SMPL
CCB21 - 1	CCB21	-48	0.0185	-	7470A	PPB	12 Aug 2022 12:55:51	U	SMPL