



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB121486

Slope : 99.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

True Value of CCV4 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	08/12/2022	11:15
2	CAL2	1	Water	NA	NA	20.8	7.00	08/12/2022	11:16
3	CAL3	1	Water	NA	NA	20.8	10.03	08/12/2022	11:19
4	ICV	1	Water	NA	NA	20.8	7.01	08/12/2022	11:20
5	CCV1	1	Water	NA	NA	20.9	2.01	08/12/2022	11:24
6	N4139-02	1	Solid	20.03	20	23.1	7.21	08/12/2022	11:25
7	N4159-01	1	Solid	20.03	20	22.7	4.58	08/12/2022	11:26
8	N4159-02	1	Solid	20.03	20	23.2	6.15	08/12/2022	11:29
9	N4179-01	1	Solid	20.04	20	22.1	7.40	08/12/2022	11:33
10	N4188-02	1	Solid	20.04	20	20.8	7.58	08/12/2022	11:37
11	N4188-02DUP	1	Solid	20.03	20	20.9	7.60	08/12/2022	11:38
12	N4188-05	1	Solid	20.02	20	20.8	8.57	08/12/2022	11:40
13	N4188-08	1	Solid	20.03	20	20.9	6.37	08/12/2022	11:45
14	N4188-11	1	Solid	20.01	20	20.8	5.45	08/12/2022	11:47
15	N4188-14	1	Solid	20.03	20	20.9	5.88	08/12/2022	11:50
16	CCV2	1	Water	NA	NA	20.8	12.03	08/12/2022	11:52
17	N4188-17	1	Solid	20.04	20	21.0	6.14	08/12/2022	12:00
18	N4188-20	1	Solid	20.05	20	21.2	8.25	08/12/2022	12:02
19	N4188-23	1	Solid	20.05	20	21.2	5.80	08/12/2022	12:05
20	N4188-26	1	Solid	20.03	20	20.9	8.19	08/12/2022	12:10

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	N4189-02	1	Solid	20.04	20	20.8	7.39	08/12/2022	12:15
22	N4189-04	1	Solid	20.02	20	21.1	8.10	08/12/2022	12:19
23	N4189-06	1	Solid	20.03	20	20.8	7.49	08/12/2022	12:22
24	N4189-08	1	Solid	20.04	20	20.6	7.90	08/12/2022	12:25
25	N4189-10	1	Solid	20.03	20	21.0	7.84	08/12/2022	12:30
26	N4189-12	1	Solid	20.04	20	20.2	7.23	08/12/2022	12:33
27	CCV3	1	Water	NA	NA	20.8	2.01	08/12/2022	12:35
28	N4189-14	1	Solid	20.03	20	20.5	7.48	08/12/2022	12:40
29	N4189-16	1	Solid	20.02	20	21.0	8.02	08/12/2022	12:44
30	N4189-18	1	Solid	20.03	20	21.2	7.34	08/12/2022	12:46
31	N4189-20	1	Solid	20.04	20	21.8	8.34	08/12/2022	12:50
32	N4189-20DUP	1	Solid	20.03	20	21.9	8.35	08/12/2022	12:51
33	CCV4	1	Water	NA	NA	20.8	12.03	08/12/2022	13:00