

LB121489


245.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

33455



A= 0.0000e+000

B= 2.9513e-004

C= -3.8204e-002

Rho= 0.9996771

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.012	-0.012	88	0.000	88					-28
0.05	0.050	0.036	-0.014	253	0.0 %	253	0				-20
0.2	0.200	0.161	-0.039	674	0.0 %	674					-1
2.5	2.500	2.487	-0.013	8555	0.0 %	8555					2
5.0	5.000	5.079	0.079	17338	0.0 %	17338					2
7.5	7.500	7.664	0.164	26098	0.0 %	26098					-2
10.0	10.000	9.835	-0.165	33455	0.0 %	33455					

LB121489 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 12 Aug 2022 14:42:18

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	88	-	0.0000	245.1	PPB	12 Aug 2022 14:45:37	S	Std
0.05 - 1	50.45	253	-	0.0500	245.1	PPB	12 Aug 2022 14:47:54	S	Std
0.2 - 1	51.2	674	-	0.2000	245.1	PPB	12 Aug 2022 14:50:11	S	Std
2.5 - 1	52.5	8555	-	2.5000	245.1	PPB	12 Aug 2022 14:52:28	S	Std
5.0 - 1	55	17338	-	5.0000	245.1	PPB	12 Aug 2022 14:54:45	S	Std
7.5 - 1	57.5	26098	-	7.5000	245.1	PPB	12 Aug 2022 14:57:00	S	Std
10.0 - 1	510	33455	-	10.0000	245.1	PPB	12 Aug 2022 15:01:54	S	Std
ICV83 - 1	ICV83	13827	4.0426	-	245.1	PPB	12 Aug 2022 15:04:45	U	SMPL
ICB83 - 1	ICB83	78	-0.0152	-	245.1	PPB	12 Aug 2022 15:07:00	U	SMPL
CCV22 - 1	CCV22	17406	5.0989	-	245.1	PPB	12 Aug 2022 15:09:16	U	SMPL
CCB22 - 1	CCB22	85	-0.0131	-	245.1	PPB	12 Aug 2022 15:11:31	U	SMPL
CRA - 1	CRA	254	0.0368	-	245.1	PPB	12 Aug 2022 15:13:52	U	SMPL
HighStd - 1	HighStd	33783	9.9322	-	245.1	PPB	12 Aug 2022 15:16:06	U	SMPL
ChkStd - 1	ChkStd	24862	7.2994	-	245.1	PPB	12 Aug 2022 15:18:22	U	SMPL
PB146981BL - 1	PBW	27	-0.0302	-	245.1	PPB	12 Aug 2022 15:20:39	U	SMPL
PB146981BS - 1	LCSW	14190	4.1497	-	245.1	PPB	12 Aug 2022 15:25:33	U	SMPL
N4170-02 - 1	EFFLUENT-METALS	660	0.1566	-	245.1	PPB	12 Aug 2022 15:32:43	U	SMPL
N4170-02DUP - 1	EFFLUENT-METALSDUP	655	0.1551	-	245.1	PPB	12 Aug 2022 15:34:59	U	SMPL
N4170-02MS - 1	EFFLUENT-METALSMS	11246	3.2809	-	245.1	PPB	12 Aug 2022 15:37:16	U	SMPL
N4170-02MSD - 1	EFFLUENT-METALSMSD	11166	3.2572	-	245.1	PPB	12 Aug 2022 15:39:33	U	SMPL
CCV23 - 1	CCV23	16186	4.7388	-	245.1	PPB	12 Aug 2022 15:41:49	U	SMPL
CCB23 - 1	CCB23	19	-0.0326	-	245.1	PPB	12 Aug 2022 15:44:06	U	SMPL
N4170-02 4x5		264	0.0397	-	245.1	PPB	12 Aug 2022 15:46:22	U	SMPL
N4170-02 A		13015	3.8029	-	245.1	PPB	12 Aug 2022 15:48:37	U	SMPL
CCV24 - 1	CCV24	16118	4.7187	-	245.1	PPB	12 Aug 2022 15:50:52	U	SMPL
CCB24 - 1	CCB24	38	-0.0270	-	245.1	PPB	12 Aug 2022 15:53:09	U	SMPL