

L3121492


7470A

INSTRUMENT ID: CVL

Linear

μ Abs.:

37253



A= 0.0000e+000

B= 2.6772e-004

C= 3.2957e-002

Rho= 0.9999216

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.035	0.035	8	0.000	8					18
0.2	0.200	0.235	0.035	756	0.0 %	756					3
2.5	2.500	2.428	-0.072	8946	0.0 %	8946					1
5.0	5.000	4.947	-0.053	18355	0.0 %	18355					1
7.5	7.500	7.548	0.048	28070	0.0 %	28070					0
10.0	10.000	10.006	0.006	37253	0.0 %	37253					

LB121492 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 12 Aug 2022 16:10:28

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc.	Method	Units	Date	Type	Type
0.0 - 1	50	8	-	0.0000	7470A	PPB	12 Aug 2022 16:11:53	S	Std
0.2 - 1	50.2	756	-	0.2000	7470A	PPB	12 Aug 2022 16:14:10	S	Std
2.5 - 1	52.5	8946	-	2.5000	7470A	PPB	12 Aug 2022 16:16:27	S	Std
5.0 - 1	55	18355	-	5.0000	7470A	PPB	12 Aug 2022 16:18:44	S	Std
7.5 - 1	57.5	28070	-	7.5000	7470A	PPB	12 Aug 2022 16:21:00	S	Std
10.0 - 1	510	37253	-	10.0000	7470A	PPB	12 Aug 2022 16:23:17	S	Std
ICV84 - 1	ICV84	15242	4.1136	-	7470A	PPB	12 Aug 2022 16:28:43	U	SMPL
ICB84 - 1	ICB84	-120	0.0008	-	7470A	PPB	12 Aug 2022 16:30:58	U	SMPL
CCV25 - 1	CCV25	17313	4.6681	-	7470A	PPB	12 Aug 2022 16:33:14	U	SMPL
CCB25 - 1	CCB25	-116	0.0019	-	7470A	PPB	12 Aug 2022 16:35:29	U	SMPL
CRA - 1	CRA	710	0.2230	-	7470A	PPB	12 Aug 2022 16:37:46	U	SMPL
HighStd - 1	HighStd	38063	10.2234	-	7470A	PPB	12 Aug 2022 16:42:37	U	SMPL
ChkStd - 1	ChkStd	23868	6.4230	-	7470A	PPB	12 Aug 2022 16:44:53	U	SMPL
PB146980BL - 1	PBW	123	0.0659	-	7470A	PPB	12 Aug 2022 16:47:11	U	SMPL
PB146980BS - 1	LCSW	14966	4.0397	-	7470A	PPB	12 Aug 2022 16:49:28	U	SMPL
N4156-01 - 1	FMC-0071	212	0.0897	-	7470A	PPB	12 Aug 2022 16:51:44	U	SMPL
N4156-02 - 1	FMC-00072	307	0.1151	-	7470A	PPB	12 Aug 2022 16:54:01	U	SMPL
N4156-03 - 1	FMC-0073	332	0.1218	-	7470A	PPB	12 Aug 2022 16:56:18	U	SMPL
N4156-04 - 1	FMC-0074	426	0.1470	-	7470A	PPB	12 Aug 2022 16:58:34	U	SMPL
N4156-05 - 1	31476	363	0.1301	-	7470A	PPB	12 Aug 2022 17:00:49	U	SMPL
CCV26 - 1	CCV26	17817	4.8030	-	7470A	PPB	12 Aug 2022 17:03:05	U	SMPL
CCB26 - 1	CCB26	-90	0.0089	-	7470A	PPB	12 Aug 2022 17:05:20	U	SMPL
N4156-08 - 1	11261	462	0.1566	-	7470A	PPB	12 Aug 2022 17:07:36	U	SMPL
N4159-01 - 1	D-2617	2265	0.6394	-	7470A	PPB	12 Aug 2022 17:09:52	U	SMPL
N4159-02 - 1	D-2619	-341	-0.0583	-	7470A	PPB	12 Aug 2022 17:12:07	U	SMPL
N4167-02 - 1	RE137-CARBON-20220810	-20	0.0276	-	7470A	PPB	12 Aug 2022 17:14:25	U	SMPL
N4179-01 - 1	D-167	338	0.1234	-	7470A	PPB	12 Aug 2022 17:16:40	U	SMPL
N4179-01DUP - 1	D-167DUP	387	0.1366	-	7470A	PPB	12 Aug 2022 17:18:56	U	SMPL
N4179-01MS - 1	D-167MS	13891	3.7519	-	7470A	PPB	12 Aug 2022 17:21:13	U	SMPL
N4179-01MSD - 1	D-167MSD	13461	3.6368	-	7470A	PPB	12 Aug 2022 17:23:30	U	SMPL
N4191-01 - 1	2362	33	0.0418	-	7470A	PPB	12 Aug 2022 17:25:47	U	SMPL
N4191-02 - 1	2664	53	0.0471	-	7470A	PPB	12 Aug 2022 17:28:03	U	SMPL
CCV27 - 1	CCV27	18280	4.9270	-	7470A	PPB	12 Aug 2022 17:30:19	U	SMPL
CCB27 - 1	CCB27	-23	0.0268	-	7470A	PPB	12 Aug 2022 17:32:35	U	SMPL
N4191-03 - 1	2665	1387	0.4043	-	7470A	PPB	12 Aug 2022 17:34:51	U	SMPL
PB146944TB - 1	PB146944TB	-48	0.0201	-	7470A	PPB	12 Aug 2022 17:37:07	U	SMPL
N4179-01LX5 - 1		88	0.0565	-	7470A	PPB	12 Aug 2022 17:39:24	U	SMPL
N4179-01A - 1		14876	4.0156	-	7470A	PPB	12 Aug 2022 17:41:40	U	SMPL
CCV28 - 1	CCV28	17952	4.8391	-	7470A	PPB	12 Aug 2022 17:43:56	U	SMPL
CCB28 - 1	CCB28	-10	0.0279	-	7470A	PPB	12 Aug 2022 17:47:48	U	SMPL