

LB121505


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

41619



A= 0.0000e+000
 B= 2.4038e-004 *Slope*
 C= 6.2014e-002 *Intercept*
 Rho= 0.9998719
 Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>1/1</i>
0.00	0.000	0.067	0.067	22	0.000	22	0				15
0.05	0.050					185					-4
0.20	0.200	0.230	0.030	699	0.0 %	699					11
2.50	2.500	2.408	-0.092	9760	0.0 %	9760					11
5.00	5.000	4.969	-0.031	20413	0.0 %	20413					11
7.50	7.500	7.459	-0.041	30772	0.0 %	30772					11
10.0	10.000	10.066	0.066	41619	0.0 %	41619					11

LB121505
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type
	0 S0	22	-		0 SFAM01.1	PPB	8/13/2022 10:36	Std
	0.2 S01	699	-		0.2 SFAM01.1	PPB	8/13/2022 10:38	Std
	2.5 S02	9760	-		2.5 SFAM01.1	PPB	8/13/2022 10:41	Std
	5 S03	20413	-		5 SFAM01.1	PPB	8/13/2022 10:43	Std
	7.5 S04	30772	-		7.5 SFAM01.1	PPB	8/13/2022 10:45	Std
	10 S05	41619	-		10 SFAM01.1	PPB	8/13/2022 10:48	Std
ICV085	ICV085	16551	4.0406 -		SFAM01.1	PPB	8/13/2022 10:50	SMPL
ICB085	ICB085	-38	0.0529 -		SFAM01.1	PPB	8/13/2022 10:53	SMPL
CCV029	CCV029	20955	5.0992 -		SFAM01.1	PPB	8/13/2022 10:55	SMPL
CCB029	CCB029	-43	0.0517 -		SFAM01.1	PPB	8/13/2022 10:57	SMPL
PB146990BL	PBS990	27	0.0685 -		SFAM01.1	PPB	8/13/2022 11:00	SMPL
N3915-01	YBB50	835	0.2627 -		SFAM01.1	PPB	8/13/2022 11:02	SMPL
N3915-02	YBB52	397	0.1574 -		SFAM01.1	PPB	8/13/2022 11:04	SMPL
N3915-03	YBB54	1416	0.4024 -		SFAM01.1	PPB	8/13/2022 11:06	SMPL
N3915-04	YBB57	86	0.0827 -		SFAM01.1	PPB	8/13/2022 11:09	SMPL
N3915-05	YBB58	2344	0.6255 -		SFAM01.1	PPB	8/13/2022 11:11	SMPL
N3915-06	YBB59	1065	0.318 -		SFAM01.1	PPB	8/13/2022 11:13	SMPL
N3915-07	YBB64	418	0.1625 -		SFAM01.1	PPB	8/13/2022 11:15	SMPL
N3915-08	YBBC6	691	0.2281 -		SFAM01.1	PPB	8/13/2022 11:18	SMPL
N3915-09	YBBC7	1363	0.3897 -		SFAM01.1	PPB	8/13/2022 11:20	SMPL
N3915-10	YBBC1	4313	1.0988 -		SFAM01.1	PPB	8/13/2022 11:22	SMPL
N3915-11	YBBC2	28062	6.8076 -		SFAM01.1	PPB	8/13/2022 11:24	SMPL
N3915-12	YBBC3	1097	0.3257 -		SFAM01.1	PPB	8/13/2022 11:27	SMPL
N3915-13	YBBC4	1529	0.4296 -		SFAM01.1	PPB	8/13/2022 11:29	SMPL
N3915-14	YBBC5	1591	0.4445 -		SFAM01.1	PPB	8/13/2022 11:31	SMPL
N3915-15	YBBC5D	1405	0.3998 -		SFAM01.1	PPB	8/13/2022 11:33	SMPL
N3915-16	YBBC5S	12780	3.1341 -		SFAM01.1	PPB	8/13/2022 11:36	SMPL
PB146991BL	PBS991	-21	0.057 -		SFAM01.1	PPB	8/13/2022 11:38	SMPL
N3901-01	GBJ42	104	0.087 -		SFAM01.1	PPB	8/13/2022 11:40	SMPL
N3901-02	GBJ42D	161	0.1007 -		SFAM01.1	PPB	8/13/2022 11:43	SMPL
CCV030	CCV030	20851	5.0742 -		SFAM01.1	PPB	8/13/2022 11:45	SMPL
CCB030	CCB030	-23	0.0565 -		SFAM01.1	PPB	8/13/2022 11:47	SMPL
N3901-03	GBJ42S	11283	2.7743 -		SFAM01.1	PPB	8/13/2022 11:49	SMPL
CCV031	CCV031	20844	5.0725 -		SFAM01.1	PPB	8/13/2022 11:52	SMPL
CCB031	CCB031	-29	0.055 -		SFAM01.1	PPB	8/13/2022 11:54	SMPL