

LB121506


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

39375



A= 0.0000e+000

B= 2.5574e-004 *slope*C= 1.4031e-002 *y-intercept*

Rho= 0.9998884

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>%D</i>
0.00	0.000	0.019	0.019	21	0.000	21					
0.05	0.050					185	0				
0.20	0.200	0.223	0.023	818	0.0 %	818					
2.50	2.500	2.498	-0.002	9711	0.0 %	9711					
5.00	5.000	4.971	-0.029	19383	0.0 %	19383					
7.50	7.500	7.404	-0.096	28895	0.0 %	28895					
10.0	10.000	10.085	0.085	39379	0.0 %	39379					

Handwritten notes:
 0.00
 12
 0
 1
 1
 1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	21	-		0 SFAM01.1	PPB	8/13/2022 12:57	Std
	0.2 S01	818	-		0.2 SFAM01.1	PPB	8/13/2022 12:59	Std
	2.5 S02	9711	-		2.5 SFAM01.1	PPB	8/13/2022 13:01	Std
	5 S03	19383	-		5 SFAM01.1	PPB	8/13/2022 13:04	Std
	7.5 S04	28895	-		7.5 SFAM01.1	PPB	8/13/2022 13:06	Std
	10 S05	39379	-		10 SFAM01.1	PPB	8/13/2022 13:08	Std
ICV086	ICV086	16098	4.131 -		SFAM01.1	PPB	8/13/2022 13:11	SMPL
ICB086	ICB086	-50	0.0012 -		SFAM01.1	PPB	8/13/2022 13:13	SMPL
CCV032	CCV032	19838	5.0875 -		SFAM01.1	PPB	8/13/2022 13:16	SMPL
CCB032	CCB032	-62	-0.0018 -		SFAM01.1	PPB	8/13/2022 13:18	SMPL
N3787-01	YBAY3	207	0.067 -		SFAM01.1	PPB	8/13/2022 13:27	SMPL
N3787-02	YBAY4	202	0.0657 -		SFAM01.1	PPB	8/13/2022 13:29	SMPL
N3787-03	YBAY5	189	0.0624 -		SFAM01.1	PPB	8/13/2022 13:31	SMPL
N3787-04	YBAY6	181	0.0603 -		SFAM01.1	PPB	8/13/2022 13:34	SMPL
N3787-05	YBAY7	184	0.0611 -		SFAM01.1	PPB	8/13/2022 13:36	SMPL
N3787-06	YBB00	245	0.0767 -		SFAM01.1	PPB	8/13/2022 13:38	SMPL
N3787-07	YBAX9	215	0.069 -		SFAM01.1	PPB	8/13/2022 13:40	SMPL
N3787-08	YBAY0	184	0.0611 -		SFAM01.1	PPB	8/13/2022 13:43	SMPL
N3787-09	YBAY1	250	0.078 -		SFAM01.1	PPB	8/13/2022 13:45	SMPL
N3787-10	YBAY2	365	0.1074 -		SFAM01.1	PPB	8/13/2022 13:47	SMPL
N3787-11	YBAY8	510	0.1445 -		SFAM01.1	PPB	8/13/2022 13:50	SMPL
N3787-12	YBAZ6	586	0.1639 -		SFAM01.1	PPB	8/13/2022 13:52	SMPL
N3787-13	YBAZ7	506	0.1434 -		SFAM01.1	PPB	8/13/2022 13:54	SMPL
N3787-14	YBAX5	251	0.0782 -		SFAM01.1	PPB	8/13/2022 13:56	SMPL
N3787-15	YBAX6	186	0.0616 -		SFAM01.1	PPB	8/13/2022 13:59	SMPL
N3787-16	YBAX7	202	0.0657 -		SFAM01.1	PPB	8/13/2022 14:01	SMPL
N3787-17	YBAX8	204	0.0662 -		SFAM01.1	PPB	8/13/2022 14:03	SMPL
N3787-18	YBAZ4	137	0.0491 -		SFAM01.1	PPB	8/13/2022 14:05	SMPL
N3787-19	YBAZ5	269	0.0828 -		SFAM01.1	PPB	8/13/2022 14:08	SMPL
CCV033	CCV033	19808	5.0798 -		SFAM01.1	PPB	8/13/2022 14:10	SMPL
CCB033	CCB033	-47	0.002 -		SFAM01.1	PPB	8/13/2022 14:12	SMPL
N3787-20	YBAZ8	154	0.0534 -		SFAM01.1	PPB	8/13/2022 14:15	SMPL
N3787-21	YBAZ8D	121	0.045 -		SFAM01.1	PPB	8/13/2022 14:17	SMPL
N3787-22	YBAZ8S	4629	1.1979 -		SFAM01.1	PPB	8/13/2022 14:19	SMPL
PB147002BL	PBW002	-38	0.0043 -		SFAM01.1	PPB	8/13/2022 14:24	SMPL
N3851-01	YBBA8	322	0.0964 -		SFAM01.1	PPB	8/13/2022 14:26	SMPL
N3851-02	YBBA9	420	0.1214 -		SFAM01.1	PPB	8/13/2022 14:28	SMPL
N3851-03	YBBB0	207	0.067 -		SFAM01.1	PPB	8/13/2022 14:31	SMPL
N3851-04	YBAX3	153	0.0532 -		SFAM01.1	PPB	8/13/2022 14:33	SMPL
N3851-05	YBAY9	283	0.0864 -		SFAM01.1	PPB	8/13/2022 14:35	SMPL
N3851-06	YBAZ1	251	0.0782 -		SFAM01.1	PPB	8/13/2022 14:37	SMPL
N3851-07	YBAZ2	306	0.0923 -		SFAM01.1	PPB	8/13/2022 14:40	SMPL
N3851-08	YBAZ3	292	0.0887 -		SFAM01.1	PPB	8/13/2022 14:42	SMPL
N3851-09	YBB01	275	0.0844 -		SFAM01.1	PPB	8/13/2022 14:44	SMPL
N3851-10	YBB02	41	0.0245 -		SFAM01.1	PPB	8/13/2022 14:47	SMPL
N3851-11	YBAZ9	21	0.0194 -		SFAM01.1	PPB	8/13/2022 14:49	SMPL

LB121506
INSTRUMENT ID: CV1

N3851-12	YBB03	26	0.0207 -	SFAM01.1	PPB	8/13/2022 14:51 SMPL
N3851-13	YBBB6	10	0.0166 -	SFAM01.1	PPB	8/13/2022 14:53 SMPL
N3851-14	YBBB7	21	0.0194 -	SFAM01.1	PPB	8/13/2022 14:56 SMPL
N3851-15	YBAW0	53	0.0276 -	SFAM01.1	PPB	8/13/2022 14:58 SMPL
N3851-16	YBAW1	45	0.0255 -	SFAM01.1	PPB	8/13/2022 15:00 SMPL
CCV034	CCV034	19582	5.022 -	SFAM01.1	PPB	8/13/2022 15:02 SMPL
CCB034	CCB034	-61	-0.0016 -	SFAM01.1	PPB	8/13/2022 15:05 SMPL
N3851-17	YBAW6	44	0.0253 -	SFAM01.1	PPB	8/13/2022 15:07 SMPL
N3851-18	YBAW9	30	0.0217 -	SFAM01.1	PPB	8/13/2022 15:09 SMPL
N3851-19	YBAW9D	39	0.024 -	SFAM01.1	PPB	8/13/2022 15:11 SMPL
N3851-20	YBAW9S	3899	1.0112 -	SFAM01.1	PPB	8/13/2022 15:14 SMPL
N3851-21	YBAX0	-4	0.013 -	SFAM01.1	PPB	8/13/2022 15:16 SMPL
PB147003BL	PBW003	42	0.0248 -	SFAM01.1	PPB	8/13/2022 15:18 SMPL
N3901-04	GBJ43	53	0.0276 -	SFAM01.1	PPB	8/13/2022 15:21 SMPL
N3901-05	GBJ43D	39	0.024 -	SFAM01.1	PPB	8/13/2022 15:23 SMPL
N3901-06	GBJ43S	4238	1.0979 -	SFAM01.1	PPB	8/13/2022 15:25 SMPL
PB147004BL	PBW004	2	0.0145 -	SFAM01.1	PPB	8/13/2022 15:27 SMPL
N3911-01	YBAW2	15	0.0179 -	SFAM01.1	PPB	8/13/2022 15:30 SMPL
N3911-02	YBAW3	11	0.0168 -	SFAM01.1	PPB	8/13/2022 15:32 SMPL
N3911-03	YBAW5	26	0.0207 -	SFAM01.1	PPB	8/13/2022 15:34 SMPL
N3911-04	YBAW5D	28	0.0212 -	SFAM01.1	PPB	8/13/2022 15:39 SMPL
N3911-05	YBAW5S	4425	1.1457 -	SFAM01.1	PPB	8/13/2022 15:42 SMPL
N3911-06	YBAW7	386	0.1127 -	SFAM01.1	PPB	8/13/2022 15:44 SMPL
N3911-07	YBAX1	379	0.111 -	SFAM01.1	PPB	8/13/2022 15:46 SMPL
N3911-08	YBAX2	307	0.0925 -	SFAM01.1	PPB	8/13/2022 15:49 SMPL
N3911-09	YBAX4	359	0.1058 -	SFAM01.1	PPB	8/13/2022 15:51 SMPL
N3911-10	YBBC8	352	0.1041 -	SFAM01.1	PPB	8/13/2022 15:53 SMPL
CCV035	CCV035	20176	5.1739 -	SFAM01.1	PPB	8/13/2022 15:55 SMPL
CCB035	CCB035	-57	-0.0005 -	SFAM01.1	PPB	8/13/2022 15:58 SMPL
PB147005BL	PBW005	24	0.0202 -	SFAM01.1	PPB	8/13/2022 16:03 SMPL
N3913-01	YBBA6	338	0.1005 -	SFAM01.1	PPB	8/13/2022 16:05 SMPL
N3913-02	YBB09	329	0.0982 -	SFAM01.1	PPB	8/13/2022 16:07 SMPL
N3913-03	YBB09D	320	0.0959 -	SFAM01.1	PPB	8/13/2022 16:12 SMPL
N3913-04	YBB09S	4266	1.105 -	SFAM01.1	PPB	8/13/2022 16:17 SMPL
N3913-05	YBB11	561	0.1575 -	SFAM01.1	PPB	8/13/2022 16:23 SMPL
N3913-06	YBB15	490	0.1393 -	SFAM01.1	PPB	8/13/2022 16:25 SMPL
N3913-07	YBB16	445	0.1278 -	SFAM01.1	PPB	8/13/2022 16:27 SMPL
N3913-08	YBB18	574	0.1608 -	SFAM01.1	PPB	8/13/2022 16:29 SMPL
N3913-09	YBBC9	529	0.1493 -	SFAM01.1	PPB	8/13/2022 16:32 SMPL
PB147001BL	PBW001	19	0.0189 -	SFAM01.1	PPB	8/13/2022 16:34 SMPL
CCV036	CCV036	20301	5.2059 -	SFAM01.1	PPB	8/13/2022 16:36 SMPL
CCB036	CCB036	-68	-0.0034 -	SFAM01.1	PPB	8/13/2022 16:38 SMPL