



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB121508

Analyst By : JIGNESH

Supervisor Review By : Sohil

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	08/13/2022	13:00
2	CAL2	1	Water	NA	NA	20.8	7.00	08/13/2022	13:01
3	CAL3	1	Water	NA	NA	20.8	10.03	08/13/2022	13:02
4	ICV	1	Water	NA	NA	20.8	6.99	08/13/2022	13:05
5	CCV1	1	Water	NA	NA	20.7	2.01	08/13/2022	13:10
6	N4159-03	1	Water	NA	NA	23.3	6.57	08/13/2022	13:52
7	N4159-03DUP	1	Water	NA	NA	23.4	6.58	08/13/2022	13:53
8	CCV2	1	Water	NA	NA	20.7	12.02	08/13/2022	14:00

WORKLIST(Hardcopy Internal Chain)

W 1215 08

WorkList Name : corrosivity n4159 W

WorkList ID : 162275

Department : Wet-Chemistry

Date : 08-13-2022 12:45:15

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/11/2022	Water	N4159-03	Corrosivity	Cool 4 deg C	PSEG03	J11	T-2590	08/11/2022	9040C

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

08.13.22 12:15
J. C. (SUN)
J. C. (SUN)

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

08.13.22 15:00
J. C. (SUN)
J. C. (SUN)