



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB121515

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.3	08/15/2022	09:25
2	N4154-01		1	15	100.00	>212.0	>212.0	08/15/2022	09:45
3	N4154-01DUP		1	15	100.00	>212.0	>212.0	08/15/2022	10:05
4	N4154-03		1	17	100.00	>212.0	>212.0	08/15/2022	10:35
5	N4159-03		1	18	100.00	>212.0	>212.0	08/15/2022	10:58
6	N4198-01		1	16	100.00	>212.0	>212.0	08/15/2022	11:15
7	N4200-19		1	18	100.00	>212.0	>212.0	08/15/2022	11:40
8	N4203-03		1	17	100.00	>212.0	>212.0	08/15/2022	12:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB121515

WorkList Name : fp81422 WorkList ID : 162283 Department : Wet-Chemistry Date : 08-14-2022 14:09:45

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/15/2022	Water	N4154-01	Flash Point	Cool 4 deg C	PSEG03	J11	WV-TOTECOMP-1	08/10/2022	1010B
08/15/2022	Water	N4154-03	Flash Point	Cool 4 deg C	PSEG03	J11	WV-TOTECOMP-2	08/10/2022	1010B
08/18/2022	Water	N4159-03	Flash Point	Cool 4 deg C	PSEG03	J11	T-2590	08/11/2022	1010B
08/17/2022	Water	N4198-01	Flash Point	Cool 4 deg C	PSEG03	K11	GFD-TOTES	08/12/2022	1010B
08/19/2022	Water	N4200-19	Flash Point	Cool 4 deg C	PSEG03	K11	WC-GW-1	08/12/2022	1010B
08/17/2022	Water	N4203-03	Flash Point	Cool 4 deg C	PSEG03	K11	61023	08/12/2022	1010B

Date/Time 08/15/22 09:00
 Raw Sample Received by: KS(WC)
 Raw Sample Relinquished by: RJ CEXT-1001

Date/Time 08/15/22 12:10
 Raw Sample Received by: RJ CEXT-1001
 Raw Sample Relinquished by: KS(WC)