



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB121517

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.3	08/15/2022	09:25
2	N4156-06		1	15	100.00	>212.0	>212.0	08/15/2022	12:35
3	N4156-06DUP		1	15	100.00	>212.0	>212.0	08/15/2022	12:55

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB121577

WorkList Name : fp81422-1

WorkList ID : 162284

Department : Wet-Chemistry

Date : 08-14-2022 14:10:52

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/15/2022	Solid	✓	N4156-06	Flash Point	Cool 4 deg C	PSEG03	J11	FMC-0070	08/10/2022 1010B

Date/Time 08/15/22 09:00
 Raw Sample Received by: KS(wc)
 Raw Sample Relinquished by: RJ (EFL-100)

Date/Time 08/15/22 13:00
 Raw Sample Received by: RJ (EFL-100)
 Raw Sample Relinquished by: KS(wc)