

LB121524


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

42112



A= 0.0000e+000

B= 2.3682e-004

C= 1.5337e-002

Rho= 0.9999147

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.034	0.034	80	0.000	80					1
0.2	0.200	0.203	0.003	792	0.0 %	792					2
2.5	2.500	2.409	-0.091	10106	0.0 %	10106					-4
5.0	5.000	5.065	0.065	21324	0.0 %	21324					1
7.5	7.500	7.501	0.001	31608	0.0 %	31608					00
10.0	10.000	9.988	-0.012	42112	0.0 %	42112					

LB121524 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 15 Aug 2022 13:58:45

Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	80	-	0.00007470A	PPB	15 Aug 2022 14:12:52	S	Std
0.2 - 1	50.2	792	-	0.20007470A	PPB	15 Aug 2022 14:15:10	S	Std
2.5 - 1	52.5	10106	-	2.50007470A	PPB	15 Aug 2022 14:17:27	S	Std
5.0 - 1	55	21324	-	5.00007470A	PPB	15 Aug 2022 14:19:44	S	Std
7.5 - 1	57.5	31608	-	7.50007470A	PPB	15 Aug 2022 14:22:00	S	Std
10.0 - 1	510	42112	-	10.00007470A	PPB	15 Aug 2022 14:24:16	S	Std
ICV87 - 1	ICV87	16077	3.8227	-7470A	PPB	15 Aug 2022 14:27:04	U	SMPL
ICB87 - 1	ICB87	-77	-0.0029	-7470A	PPB	15 Aug 2022 14:29:19	U	SMPL
CCV37 - 1	CCV37	20125	4.7813	-7470A	PPB	15 Aug 2022 14:31:34	U	SMPL
CCB37 - 1	CCB37	-79	-0.0034	-7470A	PPB	15 Aug 2022 14:33:49	U	SMPL
CRA - 1	CRA	734	0.1892	-7470A	PPB	15 Aug 2022 14:39:31	U	SMPL
HighStd - 1	HighStd	40443	9.5930	-7470A	PPB	15 Aug 2022 14:41:46	U	SMPL
ChkStd - 1	ChkStd	30741	7.2954	-7470A	PPB	15 Aug 2022 14:44:02	U	SMPL
PB147013BL - 1	PBW	-150	-0.0202	-7470A	PPB	15 Aug 2022 14:46:19	U	SMPL
PB147013BS - 1	LCSW	16284	3.8717	-7470A	PPB	15 Aug 2022 14:48:36	U	SMPL
N4188-02 - 1	S4	-88	-0.0055	-7470A	PPB	15 Aug 2022 14:50:52	U	SMPL
N4188-05 - 1	S3	32	0.0229	-7470A	PPB	15 Aug 2022 14:53:09	U	SMPL
N4188-08 - 1	S8	-5	0.0142	-7470A	PPB	15 Aug 2022 14:55:26	U	SMPL
N4188-11 - 1	S15	9	0.0175	-7470A	PPB	15 Aug 2022 14:57:43	U	SMPL
N4188-14 - 1	S14	0	0.0153	-7470A	PPB	15 Aug 2022 14:59:58	U	SMPL
CCV38 - 1	CCV38	20887	4.9618	-7470A	PPB	15 Aug 2022 15:02:14	U	SMPL
CCB38 - 1	CCB38	-83	-0.0043	-7470A	PPB	15 Aug 2022 15:04:29	U	SMPL
N4188-17 - 1	S10	-3	0.0146	-7470A	PPB	15 Aug 2022 15:06:45	U	SMPL
N4188-20 - 1	S7	-5	0.0142	-7470A	PPB	15 Aug 2022 15:09:00	U	SMPL
N4188-23 - 1	S11	14	0.0187	-7470A	PPB	15 Aug 2022 15:11:16	U	SMPL
N4188-26 - 1	S6	9	0.0175	-7470A	PPB	15 Aug 2022 15:13:32	U	SMPL
N4189-02 - 1	SB18	-8	0.0134	-7470A	PPB	15 Aug 2022 15:15:47	U	SMPL
N4189-04 - 1	SB19	1	0.0156	-7470A	PPB	15 Aug 2022 15:18:04	U	SMPL
N4189-06 - 1	SB20	11	0.0179	-7470A	PPB	15 Aug 2022 15:20:20	U	SMPL
N4189-08 - 1	SB15	5	0.0165	-7470A	PPB	15 Aug 2022 15:22:37	U	SMPL
N4189-10 - 1	SB16	5	0.0165	-7470A	PPB	15 Aug 2022 15:24:54	U	SMPL
N4189-12 - 1	SB17	-2	0.0149	-7470A	PPB	15 Aug 2022 15:27:09	U	SMPL
CCV39 - 1	CCV39	21221	5.0409	-7470A	PPB	15 Aug 2022 15:29:25	U	SMPL
CCB39 - 1	CCB39	-79	-0.0034	-7470A	PPB	15 Aug 2022 15:31:40	U	SMPL
N4189-14 - 1	SB13	-13	0.0123	-7470A	PPB	15 Aug 2022 15:33:56	U	SMPL
N4189-16 - 1	SB12	-4	0.0144	-7470A	PPB	15 Aug 2022 15:36:12	U	SMPL
N4189-18 - 1	SB14	-8	0.0134	-7470A	PPB	15 Aug 2022 15:38:27	U	SMPL
N4189-20 - 1	SB11	-5	0.0142	-7470A	PPB	15 Aug 2022 15:40:43	U	SMPL
N4189-20DUP - 1	SB11DUP	5	0.0165	-7470A	PPB	15 Aug 2022 15:42:58	U	SMPL
N4189-20MS - 1	SB11MS	17982	4.2738	-7470A	PPB	15 Aug 2022 15:45:15	U	SMPL
N4189-20MSD - 1	SB11MSD	14816	3.5240	-7470A	PPB	15 Aug 2022 15:47:31	U	SMPL
PB147014BL - 1	PBW	-76	-0.0027	-7470A	PPB	15 Aug 2022 15:49:49	U	SMPL
PB147014BS - 1	LCSW	15259	3.6290	-7470A	PPB	15 Aug 2022 15:52:05	U	SMPL
N4159-03 - 1	T-2590	-74	-0.0022	-7470A	PPB	15 Aug 2022 15:54:21	U	SMPL
CCV40 - 1	CCV40	20378	4.8412	-7470A	PPB	15 Aug 2022 15:56:37	U	SMPL
CCB40 - 1	CCB40	-90	-0.0060	-7470A	PPB	15 Aug 2022 15:58:52	U	SMPL
N4203-03 - 1	61023	-13	0.0123	-7470A	PPB	15 Aug 2022 16:01:09	U	SMPL
N4203-03DUP - 1	61023DUP	2	0.0158	-7470A	PPB	15 Aug 2022 16:03:24	U	SMPL
N4203-03MS - 1	61023MS	17335	4.1206	-7470A	PPB	15 Aug 2022 16:05:40	U	SMPL
N4203-03MSD - 1	61023MSD	15570	3.7026	-7470A	PPB	15 Aug 2022 16:07:56	U	SMPL
PB147015BL - 1	PBW	-85	-0.0048	-7470A	PPB	15 Aug 2022 16:10:11	U	SMPL
PB147015BS - 1	LCSW	14977	3.5622	-7470A	PPB	15 Aug 2022 16:12:28	U	SMPL
N4198-01 - 1	GFD-TOTES	291	0.0843	-7470A	PPB	15 Aug 2022 16:14:44	U	SMPL
N4198-01DUP - 1	GFD-TOTESDUP	201	0.0629	-7470A	PPB	15 Aug 2022 16:17:01	U	SMPL
N4198-01MS - 1	GFD-TOTESMS	18550	4.4083	-7470A	PPB	15 Aug 2022 16:19:19	U	SMPL
N4198-01MSD - 1	GFD-TOTESMSD	17076	4.0593	-7470A	PPB	15 Aug 2022 16:21:34	U	SMPL
CCV41 - 1	CCV41	21070	5.0051	-7470A	PPB	15 Aug 2022 16:23:50	U	SMPL
CCB41 - 1	CCB41	-108	-0.0102	-7470A	PPB	15 Aug 2022 16:26:06	U	SMPL
N4200-19 - 1	WC-GW-1	8895	2.1218	-7470A	PPB	15 Aug 2022 16:28:22	U	SMPL
PB146975TB - 1	PB146975TB	-56	0.0021	-7470A	PPB	15 Aug 2022 16:30:38	U	SMPL
PB146998TB - 1	PB146998TB	1	0.0156	-7470A	PPB	15 Aug 2022 16:32:53	U	SMPL
N4189-20LX5 - 1		-30	0.0082	-7470A	PPB	15 Aug 2022 16:35:09	U	SMPL
N4189-20A - 1		16028	3.8111	-7470A	PPB	15 Aug 2022 16:37:25	U	SMPL
N4203-03LX5 - 1		-69	-0.0010	-7470A	PPB	15 Aug 2022 16:39:42	U	SMPL
N4203-03A - 1		16142	3.8381	-7470A	PPB	15 Aug 2022 16:41:58	U	SMPL
N4198-01LX5 - 1		-78	-0.0031	-7470A	PPB	15 Aug 2022 16:44:15	U	SMPL
N4198-01A - 1		18094	4.3003	-7470A	PPB	15 Aug 2022 16:46:32	U	SMPL
CCV42 - 1	CCV42	20979	4.9836	-7470A	PPB	15 Aug 2022 16:48:48	U	SMPL
CCB42 - 1	CCB42	-101	-0.0086	-7470A	PPB	15 Aug 2022 16:51:04	U	SMPL