

LB121525

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

8/15/2022 14:55 Reviewed by : 12 Instrument ID : Konelab

Test: Total CN

Sample Id	^{mg/L} Result	Dil. 1 +	Response	Errors
ICV1	94.635	0.0	0.075	
ICB1	0.015	0.0	0.001	
CCV1	246.938	0.0	0.193	
CCB1	0.143	0.0	0.001	
PB147040BL	-0.055	0.0	0.001	
N4051-02	-0.261	0.0	0.001	
N4051-02DUP	-0.232	0.0	0.001	
CCV2	260.530	0.0	0.204	
CCB2	-0.044	0.0	0.001	

N	9
Mean	66.852
SD	110.4466
CV%	165.21

Aquakem v. 7.2AQ1

Results from time period:

Mon Aug 15 14:43:04 2022

Mon Aug 15 14:46:57 2022

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	-0.5758	µg/l	8/15/2022 11:25:04	
5.0PPBCN	A	Total CN	P	4.3633	µg/l	8/15/2022 11:25:05	
10PPBCN	A	Total CN	P	9.2898	µg/l	8/15/2022 11:25:06	
50PPBCN	A	Total CN	P	48.4601	µg/l	8/15/2022 11:25:07	
100PPBCN	A	Total CN	P	98.8152	µg/l	8/15/2022 11:25:08	
250PPBCN	A	Total CN	P	258.4717	µg/l	8/15/2022 11:25:09	
500PPBCN	A	Total CN	P	496.1757	µg/l	8/15/2022 11:25:10	
ICV1	S	Total CN	P	94.6349	µg/l	8/15/2022 14:43:05	
ICB1	S	Total CN	P	0.0148	µg/l	8/15/2022 14:43:07	
CCV1	S	Total CN	P	246.9384	µg/l	8/15/2022 14:43:08	
CCB1	S	Total CN	P	0.1432	µg/l	8/15/2022 14:43:10	
PB147040BL	S	Total CN	P	-0.0552	µg/l	8/15/2022 14:43:12	
N4051-02	S	Total CN	P	-0.2607	µg/l	8/15/2022 14:43:14	
N4051-02DUP	S	Total CN	P	-0.2325	µg/l	8/15/2022 14:46:55	
CCV2	S	Total CN	P	260.5303	µg/l	8/15/2022 14:46:56	
CCB2	S	Total CN	P	-0.0436	µg/l	8/15/2022 14:46:57	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12

Instrument ID : Konelab

8/15/2022 11:28


Test Total CN

Accepted 8/15/2022 11:28

Factor 1285
Bias 0.001

Coeff. of det. 0.999555

Errors



	Calibrator	Response	Calc. con.	Conc.	Re Errors
1	0.0PPBCN	0.001	-0.5758	0.0000	-
2	5.0PPBCN	0.005	4.3633	5.0000	-12.7
3	10PPBCN	0.009	9.2898	10.0000	-7.1
4	50PPBCN	0.039	48.4601	50.0000	-3.1
5	100PPBCN	0.078	98.8152	100.0000	-1.2
6	250PPBCN	0.202	258.4717	250.0000	3.4
7	500PPBCN	0.387	496.1757	500.0000	-0.8

12
8/15/22