

LB121531


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

54030



A= 0.0000e+000

B= 1.8465e-004

C= -2.9973e-002

Rho= 0.9999467

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/00
0.0	0.000	-0.006	-0.006	132	0.000	132					-
0.2	0.200	0.168	-0.032	1071	0.0 %	1071					-16
2.5	2.500	2.499	-0.001	13698	0.0 %	13698					-0
5.0	5.000	5.062	0.062	27578	0.0 %	27578					-0
7.5	7.500	7.530	0.030	40943	0.0 %	40943					-0
10.0	10.000	9.946	-0.054	54030	0.0 %	54030					-1

LB121531 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 16 Aug 2022 07:12:40

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	132	-	0.0000	7471B	PPB	16 Aug 2022 07:22:29	S	Std
0.2 - 1	50.2	1071	-	0.2000	7471B	PPB	16 Aug 2022 07:24:47	S	Std
2.5 - 1	52.5	13698	-	2.5000	7471B	PPB	16 Aug 2022 07:27:04	S	Std
5.0 - 1	55	27578	-	5.0000	7471B	PPB	16 Aug 2022 07:29:21	S	Std
7.5 - 1	57.5	40943	-	7.5000	7471B	PPB	16 Aug 2022 07:31:37	S	Std
10.0 - 1	511	54030	-	10.0000	7471B	PPB	16 Aug 2022 07:36:29	S	Std
ICV88 - 1	ICV88	22724	4.1659	-	7471B	PPB	16 Aug 2022 07:39:14	U	SMPL
ICB88 - 1	ICB88	-89	-0.0464	-	7471B	PPB	16 Aug 2022 07:41:29	U	SMPL
CCV43 - 1	CCV43	28538	5.2394	-	7471B	PPB	16 Aug 2022 07:43:45	U	SMPL
CCB43 - 1	CCB43	-104	-0.0492	-	7471B	PPB	16 Aug 2022 07:46:00	U	SMPL
CRA - 1	CRA	1128	0.1783	-	7471B	PPB	16 Aug 2022 07:50:53	U	SMPL
HighStd - 1	HighStd	54909	10.1087	-	7471B	PPB	16 Aug 2022 07:53:08	U	SMPL
ChkStd - 1	ChkStd	39740	7.3078	-	7471B	PPB	16 Aug 2022 07:55:24	U	SMPL
PB147036BL - 1	PBS	-156	-0.0588	-	7471B	PPB	16 Aug 2022 07:57:43	U	SMPL
PB147036BS - 1	LCSS	23148	4.2442	-	7471B	PPB	16 Aug 2022 08:00:07	U	SMPL
N4139-01 - 1	PIG-SWAB-1-7	16704	3.0543	-	7471B	PPB	16 Aug 2022 08:02:19	U	SMPL
N4179-02 - 1	RO-7211930	1518	0.2503	-	7471B	PPB	16 Aug 2022 08:04:36	U	SMPL
N4183-02 - 1	STOCKPILE	3699	0.6530	-	7471B	PPB	16 Aug 2022 08:06:54	U	SMPL
N4188-01 - 1	S4	2805	0.4880	-	7471B	PPB	16 Aug 2022 08:09:12	U	SMPL
N4188-04 - 1	S3	5739	1.0297	-	7471B	PPB	16 Aug 2022 08:11:27	U	SMPL
CCV44 - 1	CCV44	28081	5.1550	-	7471B	PPB	16 Aug 2022 08:13:43	U	SMPL
CCB44 - 1	CCB44	-94	-0.0473	-	7471B	PPB	16 Aug 2022 08:15:58	U	SMPL
N4188-07 - 1	S8	5670	1.0170	-	7471B	PPB	16 Aug 2022 08:18:14	U	SMPL
N4188-10 - 1	S15	5629	1.0094	-	7471B	PPB	16 Aug 2022 08:20:30	U	SMPL
N4188-13 - 1	S14	1368	0.2226	-	7471B	PPB	16 Aug 2022 08:22:45	U	SMPL
N4188-16 - 1	S10	3665	0.6468	-	7471B	PPB	16 Aug 2022 08:25:01	U	SMPL
N4188-19 - 1	S7	1414	0.2311	-	7471B	PPB	16 Aug 2022 08:27:16	U	SMPL
N4188-22 - 1	S11	2017	0.3425	-	7471B	PPB	16 Aug 2022 08:29:33	U	SMPL
N4188-25 - 1	S6	2360	0.4058	-	7471B	PPB	16 Aug 2022 08:31:50	U	SMPL
N4191-04 - 1	RO-367	3187	0.5585	-	7471B	PPB	16 Aug 2022 08:34:07	U	SMPL
N4191-06 - 1	RO-211	11040	2.0085	-	7471B	PPB	16 Aug 2022 08:36:23	U	SMPL
N4197-02 - 1	TT-203-IDWSO-20220811-1	1157	0.1837	-	7471B	PPB	16 Aug 2022 08:38:39	U	SMPL
CCV45 - 1	CCV45	28354	5.2055	-	7471B	PPB	16 Aug 2022 08:40:54	U	SMPL
CCB45 - 1	CCB45	-100	-0.0484	-	7471B	PPB	16 Aug 2022 08:43:10	U	SMPL
N4197-03 - 1	TT-203-IDWSO-20220811-2	978	0.1506	-	7471B	PPB	16 Aug 2022 08:45:27	U	SMPL
N4199-01 - 1	HD-02-081222	1648	0.2743	-	7471B	PPB	16 Aug 2022 08:47:43	U	SMPL
N4201-02 - 1	331	2710	0.4704	-	7471B	PPB	16 Aug 2022 08:49:58	U	SMPL
N4208-01 - 1	W-1	1655	0.2756	-	7471B	PPB	16 Aug 2022 08:52:14	U	SMPL
N4208-02 - 1	W-2	1406	0.2296	-	7471B	PPB	16 Aug 2022 08:54:30	U	SMPL
N4208-02DUP - 1	W-2DUP	1280	0.2064	-	7471B	PPB	16 Aug 2022 08:56:46	U	SMPL
N4208-02MS - 1	W-2MS	22923	4.2026	-	7471B	PPB	16 Aug 2022 08:59:03	U	SMPL
N4208-02MSD - 1	W-2MSD	22231	4.0749	-	7471B	PPB	16 Aug 2022 09:01:20	U	SMPL
PB147038BL - 1	PBS	-54	-0.0399	-	7471B	PPB	16 Aug 2022 09:03:38	U	SMPL
PB147038BS - 1	LCSS	22421	4.1100	-	7471B	PPB	16 Aug 2022 09:05:55	U	SMPL
CCV46 - 1	CCV46	28322	5.1995	-	7471B	PPB	16 Aug 2022 09:08:11	U	SMPL
CCB46 - 1	CCB46	-109	-0.0501	-	7471B	PPB	16 Aug 2022 09:10:28	U	SMPL
N4140-01 - 1	LINDEN-MR-B1B-SP	129264	23.8380	-	7471B	PPB	16 Aug 2022 09:12:45	U	SMPL
N4200-01 - 1	S13	1615	0.2682	-	7471B	PPB	16 Aug 2022 09:15:01	U	SMPL
N4200-01DUP - 1	S13DUP	1561	0.2583	-	7471B	PPB	16 Aug 2022 09:20:21	U	SMPL
N4200-01MS - 1	S13MS	24356	4.4672	-	7471B	PPB	16 Aug 2022 09:33:00	U	SMPL
N4200-01MSD - 1	S13MSD	23996	4.4008	-	7471B	PPB	16 Aug 2022 09:35:15	U	SMPL
N4200-04 - 1	S9	3705	0.6541	-	7471B	PPB	16 Aug 2022 09:37:33	U	SMPL
N4200-07 - 1	S12	2354	0.4047	-	7471B	PPB	16 Aug 2022 09:39:51	U	SMPL
N4200-10 - 1	S5	1222	0.1957	-	7471B	PPB	16 Aug 2022 09:42:08	U	SMPL
N4200-13 - 1	S1	318	0.0287	-	7471B	PPB	16 Aug 2022 09:44:25	U	SMPL
N4200-16 - 1	S2	452	0.0535	-	7471B	PPB	16 Aug 2022 09:46:41	U	SMPL
CCV47 - 1	CCV47	28073	5.1536	-	7471B	PPB	16 Aug 2022 09:48:57	U	SMPL
CCB47 - 1	CCB47	-104	-0.0492	-	7471B	PPB	16 Aug 2022 09:51:13	U	SMPL
N4205-01 - 1	SS-1	9859	1.7904	-	7471B	PPB	16 Aug 2022 09:53:30	U	SMPL
N4205-02 - 1	SS-2	8923	1.6176	-	7471B	PPB	16 Aug 2022 09:55:46	U	SMPL
N4205-03 - 1	SS-3	11447	2.0837	-	7471B	PPB	16 Aug 2022 09:58:01	U	SMPL
N4205-04 - 1	SS-4	13320	2.4295	-	7471B	PPB	16 Aug 2022 10:00:18	U	SMPL
N4205-05 - 1	SS-5	11221	2.0419	-	7471B	PPB	16 Aug 2022 10:02:33	U	SMPL
N4205-06 - 1	SS-6	6496	1.1695	-	7471B	PPB	16 Aug 2022 10:04:50	U	SMPL
N4205-07 - 1	SS-7	8428	1.5258	-	7471B	PPB	16 Aug 2022 10:07:07	U	SMPL
N4211-01 - 1	WC-1	929	0.1416	-	7471B	PPB	16 Aug 2022 10:09:23	U	SMPL
N4211-02 - 1	WC-2	441	0.0515	-	7471B	PPB	16 Aug 2022 10:11:40	U	SMPL
CCV48 - 1	CCV48	28089	5.1565	-	7471B	PPB	16 Aug 2022 10:13:56	U	SMPL
CCB48 - 1	CCB48	-102	-0.0488	-	7471B	PPB	16 Aug 2022 10:16:13	U	SMPL
N4208-02LX5 - 1		256	0.0173	-	7471B	PPB	16 Aug 2022 10:28:03	U	SMPL
N4208-02A - 1		22309	4.0893	-	7471B	PPB	16 Aug 2022 10:30:19	U	SMPL
N4200-01LX5 - 1		283	0.0223	-	7471B	PPB	16 Aug 2022 10:32:35	U	SMPL
N4200-01A - 1		24052	4.4111	-	7471B	PPB	16 Aug 2022 10:34:52	U	SMPL
N4140-01DLX5 - 1		21984	4.0293	-	7471B	PPB	16 Aug 2022 10:37:08	U	SMPL
CCV49 - 1	CCV49	27944	5.1297	-	7471B	PPB	16 Aug 2022 10:39:26	U	SMPL
CCB49 - 1	CCB49	-136	-0.0551	-	7471B	PPB	16 Aug 2022 10:41:46	U	SMPL