

LB121541


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

43002



A= 0.0000e+000

B= 2.3347e-004 slope

C= 2.3331e-002 *intercept*

Rho= 0.9999023

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.060	0.060	158	0.000	158					=
0.05	0.050					185	0				-9
0.20	0.200	0.183	-0.017	684	0.0 %	684					-1
2.50	2.500	2.463	-0.037	10451	0.0 %	10451					0
5.00	5.000	5.012	0.012	21366	0.0 %	21366					-1
7.50	7.500	7.419	-0.081	31676	0.0 %	31676					1
10.0	10.000	10.063	0.063	43002	0.0 %	43002					

LB121541

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	158	-		0 SFAM01.1	PPB	8/16/2022 13:50	Std
	0.2 S01	684	-		0.2 SFAM01.1	PPB	8/16/2022 13:52	Std
	2.5 S02	10451	-		2.5 SFAM01.1	PPB	8/16/2022 13:55	Std
	5 S03	21366	-		5 SFAM01.1	PPB	8/16/2022 13:57	Std
	7.5 S04	31676	-		7.5 SFAM01.1	PPB	8/16/2022 13:59	Std
	10 S05	43002	-		10 SFAM01.1	PPB	8/16/2022 14:02	Std
ICV089	ICV089	17465	4.1009 -		SFAM01.1	PPB	8/16/2022 14:05	SMPL
ICB089	ICB089	86	0.0434 -		SFAM01.1	PPB	8/16/2022 14:07	SMPL
CCV050	CCV050	21242	4.9827 -		SFAM01.1	PPB	8/16/2022 14:09	SMPL
CCB050	CCB050	94	0.0453 -		SFAM01.1	PPB	8/16/2022 14:12	SMPL
PB147066BL	PBW066	110	0.049 -		SFAM01.1	PPB	8/16/2022 14:14	SMPL
N4137-01	GBJN8	105	0.0478 -		SFAM01.1	PPB	8/16/2022 14:16	SMPL
N4137-02	GBJN8D	86	0.0434 -		SFAM01.1	PPB	8/16/2022 14:18	SMPL
N4137-03	GBJN8S	4517	1.0779 -		SFAM01.1	PPB	8/16/2022 14:21	SMPL
N4137-06	GBJN9	10	0.0257 -		SFAM01.1	PPB	8/16/2022 14:25	SMPL
N4137-07	GBJP0	130	0.0537 -		SFAM01.1	PPB	8/16/2022 14:27	SMPL
PB147067BL	PBW067	110	0.049 -		SFAM01.1	PPB	8/16/2022 14:30	SMPL
N4138-01	GBJP1	110	0.049 -		SFAM01.1	PPB	8/16/2022 14:32	SMPL
N4138-02	GBJP1D	101	0.0469 -		SFAM01.1	PPB	8/16/2022 14:34	SMPL
N4138-03	GBJP1S	4554	1.0866 -		SFAM01.1	PPB	8/16/2022 14:36	SMPL
N4138-04	GBJP2	75	0.0408 -		SFAM01.1	PPB	8/16/2022 14:39	SMPL
N4138-05	GBJP3	114	0.0499 -		SFAM01.1	PPB	8/16/2022 14:41	SMPL
N4137-05	GBJP5	22199	5.2061 -		SFAM01.1	PPB	8/16/2022 14:43	SMPL
CCV051	CCV051	22253	5.2187 -		SFAM01.1	PPB	8/16/2022 14:55	SMPL
CCB051	CCB051	66	0.0387 -		SFAM01.1	PPB	8/16/2022 14:58	SMPL