



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB121552

Analyst By : JIGNESH

Supervisor Review By : Sohil

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	08/17/2022	08:25
2	CAL2	1	Water	NA	NA	20.8	7.00	08/17/2022	08:26
3	CAL3	1	Water	NA	NA	20.8	10.03	08/17/2022	08:29
4	ICV	1	Water	NA	NA	20.7	6.98	08/17/2022	08:30
5	CCV1	1	Water	NA	NA	20.7	2.01	08/17/2022	08:31
6	N4203-03	1	Water	NA	NA	24.4	9.64	08/17/2022	08:35
7	N4203-03DUP	1	Water	NA	NA	24.5	9.65	08/17/2022	08:36
8	CCV2	1	Water	NA	NA	20.7	12.03	08/17/2022	08:37

WORKLIST(Hardcopy Internal Chain)

VB 121552

WorkList Name : PH N4203

WorkList ID : 162367

Department : Wet-Chemistry

Date : 08-17-2022 08:18:56

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/12/2022	Water	N4203-03	pH	Cool 4 deg C	PSEG03	K11	61023	08/12/2022	9040C

Date/Time 08/17/22 08:22

Raw Sample Received by: JH/OC

Raw Sample Relinquished by: JH/OC

Date/Time 08/17/22 09:30

Raw Sample Received by: JH/OC

Raw Sample Relinquished by: JH/OC