

LB 121 854


7470A

INSTRUMENT ID: CV4

Linear

 μ Abs.:

40204



A= 0.0000e+000

B= 2.4718e-004

C= 1.5402e-002

Rho= 0.9997843

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	1/2
0.0	0.000	0.038	0.038	92	0.000	92					1
0.2	0.200	0.203	0.003	760	0.0 %	760					2
2.5	2.500	2.365	-0.135	9507	0.0 %	9507					-5
5.0	5.000	5.112	0.112	20621	0.0 %	20621					2
7.5	7.500	7.528	0.028	30393	0.0 %	30393					0
10.0	10.000	9.953	-0.047	40204	0.0 %	40204					0

LB121554 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 17 Aug 2022 10:36:45

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	92	-	0.0000	7470A	PPB	17 Aug 2022 10:41:41	S	Std
0.2 - 1	S0.2	760	-	0.2000	7470A	PPB	17 Aug 2022 10:43:59	S	Std
2.5 - 1	S2.5	9507	-	2.5000	7470A	PPB	17 Aug 2022 10:46:15	S	Std
5.0 - 1	S5	20621	-	5.0000	7470A	PPB	17 Aug 2022 10:48:32	S	Std
7.5 - 1	S7.5	30393	-	7.5000	7470A	PPB	17 Aug 2022 10:50:49	S	Std
10.0 - 1	S10	40204	-	10.0000	7470A	PPB	17 Aug 2022 10:55:42	S	Std
ICV90 - 1	ICV90	16373	4.0624	-	7470A	PPB	17 Aug 2022 10:58:29	U	SMPL
ICB90 - 1	ICB90	-49	0.0033	-	7470A	PPB	17 Aug 2022 11:00:44	U	SMPL
CCV52 - 1	CCV52	21983	5.4491	-	7470A	PPB	17 Aug 2022 11:03:00	U	SMPL
CCB52 - 1	CCB52	-50	0.0030	-	7470A	PPB	17 Aug 2022 11:05:15	U	SMPL
CRA - 1	CRA	726	0.1949	-	7470A	PPB	17 Aug 2022 11:07:31	U	SMPL
HighStd - 1	HighStd	40541	10.0362	-	7470A	PPB	17 Aug 2022 11:09:46	U	SMPL
ChkStd - 1	ChkStd	29632	7.3398	-	7470A	PPB	17 Aug 2022 11:12:01	U	SMPL
PB147039BL - 1	PBW	-165	-0.0254	-	7470A	PPB	17 Aug 2022 11:14:18	U	SMPL
PB147039BS - 1	LCSW	15301	3.7975	-	7470A	PPB	17 Aug 2022 11:16:35	U	SMPL
N4200-02 - 1	S13	-82	-0.0049	-	7470A	PPB	17 Aug 2022 11:18:52	U	SMPL
N4200-02DUP - 1	S13DUP	-12	0.0124	-	7470A	PPB	17 Aug 2022 11:21:08	U	SMPL
N4200-02MS - 1	S13MS	15497	3.8459	-	7470A	PPB	17 Aug 2022 11:23:25	U	SMPL
N4200-02MSD - 1	S13MSD	14910	3.7008	-	7470A	PPB	17 Aug 2022 11:25:42	U	SMPL
N4200-05 - 1	S9	-97	-0.0086	-	7470A	PPB	17 Aug 2022 11:27:58	U	SMPL
CCV53 - 1	CCV53	18927	4.6937	-	7470A	PPB	17 Aug 2022 11:38:31	U	SMPL
CCB53 - 1	CCB53	-127	-0.0160	-	7470A	PPB	17 Aug 2022 11:40:46	U	SMPL
N4200-08 - 1	S12	-27	0.0087	-	7470A	PPB	17 Aug 2022 11:43:02	U	SMPL
N4200-11 - 1	S5	-19	0.0107	-	7470A	PPB	17 Aug 2022 11:45:17	U	SMPL
N4200-14 - 1	S1	-1	0.0152	-	7470A	PPB	17 Aug 2022 11:47:32	U	SMPL
N4200-17 - 1	S2	-3	0.0147	-	7470A	PPB	17 Aug 2022 11:49:48	U	SMPL
N4202-02 - 1	SB10	-2	0.0149	-	7470A	PPB	17 Aug 2022 11:52:04	U	SMPL
N4202-04 - 1	SB08	-2	0.0149	-	7470A	PPB	17 Aug 2022 11:54:21	U	SMPL
N4203-01 - 1	33105	186	0.0614	-	7470A	PPB	17 Aug 2022 11:56:38	U	SMPL
N4203-02 - 1	61023	296	0.0886	-	7470A	PPB	17 Aug 2022 11:58:55	U	SMPL
PB147065BL - 1	PBW	27	0.0221	-	7470A	PPB	17 Aug 2022 12:01:12	U	SMPL
PB147065BS - 1	LCSW	13297	3.3021	-	7470A	PPB	17 Aug 2022 12:03:28	U	SMPL
CCV54 - 1	CCV54	19689	4.8821	-	7470A	PPB	17 Aug 2022 12:05:43	U	SMPL
CCB54 - 1	CCB54	-83	-0.0051	-	7470A	PPB	17 Aug 2022 12:07:59	U	SMPL
N4046-02 - 1	BEACONPOINT-BACKFILL-01	203	0.0656	-	7470A	PPB	17 Aug 2022 12:10:15	U	SMPL
N4046-04 - 1	BEACONPOINT-BACKFILL-02	399	0.1140	-	7470A	PPB	17 Aug 2022 12:12:31	U	SMPL
N4046-06 - 1	BEACONPOINT-BACKFILL-03	301	0.0898	-	7470A	PPB	17 Aug 2022 12:14:47	U	SMPL
N4046-08 - 1	BEACONPOINT-BACKFILL-04	134	0.0485	-	7470A	PPB	17 Aug 2022 12:17:03	U	SMPL
N4046-10 - 1	BEACONPOINT-BACKFILL-05	95	0.0389	-	7470A	PPB	17 Aug 2022 12:19:19	U	SMPL
N4046-12 - 1	BEACONPOINT-BACKFILL-06	77	0.0344	-	7470A	PPB	17 Aug 2022 12:21:35	U	SMPL
N4046-14 - 1	BEACONPOINT-BACKFILL-07	58	0.0297	-	7470A	PPB	17 Aug 2022 12:23:52	U	SMPL
N4046-16 - 1	BEACONPOINT-BACKFILL-08	100	0.0401	-	7470A	PPB	17 Aug 2022 12:26:09	U	SMPL
N4046-18 - 1	BEACONPOINT-BACKFILL-09	100	0.0401	-	7470A	PPB	17 Aug 2022 12:28:26	U	SMPL
N4046-20 - 1	BEACONPOINT-BACKFILL-10	85	0.0364	-	7470A	PPB	17 Aug 2022 12:30:42	U	SMPL
CCV55 - 1	CCV55	20435	5.0665	-	7470A	PPB	17 Aug 2022 12:32:58	U	SMPL
CCB55 - 1	CCB55	27	0.0221	-	7470A	PPB	17 Aug 2022 12:35:14	U	SMPL
N4046-22 - 1	BEACONPOINT-BACKFILL-11	200	0.0648	-	7470A	PPB	17 Aug 2022 12:37:30	U	SMPL
N4046-22DUP - 1	BEACONPOINT-BACKFILL-11DUP	120	0.0451	-	7470A	PPB	17 Aug 2022 12:39:45	U	SMPL
N4046-22MS - 1	BEACONPOINT-BACKFILL-11MS	15639	3.8810	-	7470A	PPB	17 Aug 2022 12:42:01	U	SMPL
N4046-22MSD - 1	BEACONPOINT-BACKFILL-11MSD	14355	3.5636	-	7470A	PPB	17 Aug 2022 12:44:16	U	SMPL
PB146999TB - 1	PB146999TB	-43	0.0048	-	7470A	PPB	17 Aug 2022 12:46:33	U	SMPL
PB147065TB - 1	PB147041TB	49	0.0275	-	7470A	PPB	17 Aug 2022 12:48:49	U	SMPL
N4200-02LX5 - 1		-52	0.0025	-	7470A	PPB	17 Aug 2022 12:51:05	U	SMPL
N4200-02A - 1		17793	4.4134	-	7470A	PPB	17 Aug 2022 12:53:22	U	SMPL
N4046-22LX5 - 1		-52	0.0025	-	7470A	PPB	17 Aug 2022 12:55:40	U	SMPL
N4046-22A - 1		18981	4.7071	-	7470A	PPB	17 Aug 2022 12:57:55	U	SMPL
CCV56 - 1	CCV56	20870	5.1740	-	7470A	PPB	17 Aug 2022 13:00:11	U	SMPL
CCB56 - 1	CCB56	-65	-0.0007	-	7470A	PPB	17 Aug 2022 13:02:27	U	SMPL