



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Iwona

Parameter: COD

SUPERVISOR REVIEW BY: jignesh

Run Number: LB121558

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP96871
COD calibration std. 100 ppm	WP96868
COD calibration std. 10 ppm	WP96866
COD calibration std. 150 ppm	WP96869
COD calibration std. 50 ppm	WP96867
COD calibration std. 0 ppm	WP96865
COD Digestion Vials Low Level 0-150Mg/L	W2935
COD ICV-LCS std, 50ppm	WP98367
COD CCV std, 50ppm	WP98366
COD LOD std, 5ppm	WP98368
COD LOQ std, 10.0PPM	WP98462

Temp In (C): 149	Date In: 08/16/2022	Time In: 13:00
Temp Out (C): 151	Date Out: 08/16/2022	Time Out: 15:00

Intercept: 0.5982

Slope: 0.9936

Regression: 0.9999

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	-0.602		05/19/2022	13:35
2	CAL2	10	1	Water	11.000	10.469	-4.7	05/19/2022	13:35
3	CAL3	50	1	Water	50.000	49.72	0.6	05/19/2022	13:36
4	CAL4	100	1	Water	101.000	101.049	-1	05/19/2022	13:36
5	CAL5	150	1	Water	149.000	149.358	0.4	05/19/2022	13:37



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Iwona

Parameter: COD

SUPERVISOR REVIEW BY: jignesh

Run Number: LB121558

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50	NA	NA	1	Water	51.000	50.726	05/19/2022	13:37
2	ICB		NA	NA	1	Water	0.000	-0.602	05/19/2022	13:38
3	CCV1	50	NA	NA	1	Water	49.000	48.714	08/17/2022	09:10
4	CCB1		NA	NA	1	Water	0.000	-0.602	08/17/2022	09:10
5	LB121558BL		1	50	1	Solid	1.000	0.404	08/17/2022	09:11
6	LB121558BS	50	1	50	1	Solid	51.000	50.726	08/17/2022	09:11
7	N3980-01		1	50	1	Solid	6.000	5.437	08/17/2022	09:12
8	N3980-02		1	50	1	Solid	11.000	10.469	08/17/2022	09:12
9	CCV2	50	NA	NA	1	Water	50.000	49.720	08/17/2022	09:13
10	CCB2		NA	NA	1	Water	1.000	0.404	08/17/2022	09:13