



PERCENT SOLID

Supervisor: Sohil
Analyst: jignesh
Date: 8/18/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 17:10
In Date: 08/17/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:20
Out Date: 08/18/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB121567

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N4226-01	B-1	1	0.92	8.60	9.52	9.17	95.9	
N4232-01	BU-01-081722	8	0.91	8.71	9.62	8.78	90.4	
N4232-02	BU-01-081722-EPH-2	9	0.96	8.76	9.72	8.64	87.7	
N4233-01	40053	10	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
N4233-02	40063	11	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
N4239-03	MDL-SOIL-QT3-2022-05	2	1.00	1.00	2.00	2.00	100.0	
N4239-04	MDL-SOIL-QT3-2022-06	3	1.00	1.00	2.00	2.00	100.0	
N4239-05	MDL-MED-SOIL-QT3-2022-05	4	1.00	1.00	2.00	2.00	100.0	
N4239-06	MDL-MED-SOIL-QT3-2022-06	5	1.00	1.00	2.00	2.00	100.0	
N4240-01	TR-06-081722	12	0.93	8.93	9.86	9.34	94.2	
N4240-02	TR-06-081722-EPH-2	13	0.92	8.92	9.84	9.00	90.6	
N4242-01	CL-02-081722	6	0.95	8.61	9.56	8.95	92.9	
N4242-02	CL-02-081722-EPH-2	7	0.95	8.61	9.56	8.95	92.9	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

VB 121567

WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-081722			WorkList ID : 162366			Department : Wet-Chemistry		Date : 08-17-2022 08:11:30	
Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/26/2022	Solid	N4226-01	Percent Solids	Cool 4 deg C	TECT01	N11	B-1	08/16/2022	Chemtech -SO
08/22/2022	Solid	N4232-01	Percent Solids	Cool 4 deg C	PSEG05	K11	BU-01-081722	08/17/2022	Chemtech -SO
08/22/2022	Solid	N4232-02	Percent Solids	Cool 4 deg C	PSEG05	K11	BU-01-081722-EPH-2	08/17/2022	Chemtech -SO
08/22/2022	Solid	N4233-01	Percent Solids	Cool 4 deg C	PSEG03	K11	40053	08/17/2022	Chemtech -SO
08/22/2022	Solid	N4233-02	Percent Solids	Cool 4 deg C	PSEG03	K11	40063	08/17/2022	Chemtech -SO
08/27/2022	Solid	N4239-03	Percent Solids	Cool 4 deg C	CHEM02	QA Of	MDL-SOIL-QT3-2022-05	08/17/2022	Chemtech -SO
08/27/2022	Solid	N4239-04	Percent Solids	Cool 4 deg C	CHEM02	QA Of	MDL-SOIL-QT3-2022-06	08/17/2022	Chemtech -SO
08/27/2022	Solid	N4239-05	Percent Solids	Cool 4 deg C	CHEM02	QA Of	MDL-MED-SOIL-QT3-2022-05	08/17/2022	Chemtech -SO
08/27/2022	Solid	N4239-06	Percent Solids	Cool 4 deg C	CHEM02	QA Of	MDL-MED-SOIL-QT3-2022-06	08/17/2022	Chemtech -SO
08/22/2022	Solid	N4240-01	Percent Solids	Cool 4 deg C	PSEG05	K11	TR-06-081722	08/17/2022	Chemtech -SO
08/22/2022	Solid	N4240-02	Percent Solids	Cool 4 deg C	PSEG05	K11	TR-06-081722-EPH-2	08/17/2022	Chemtech -SO
08/22/2022	Solid	N4242-01	Percent Solids	Cool 4 deg C	PSEG05	K11	CL-02-081722	08/17/2022	Chemtech -SO
08/22/2022	Solid	N4242-02	Percent Solids	Cool 4 deg C	PSEG05	K11	CL-02-081722-EPH-2	08/17/2022	Chemtech -SO

Date/Time 08-17-22 16:25
Raw Sample Received by: JD (WC)
Raw Sample Relinquished by: KS (WC)

Date/Time 08-17-22 17:20
Raw Sample Received by: KS (WC)
Raw Sample Relinquished by: JD (WC)