

LB121562


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

47749



A= 0.0000e+000

B= 2.0930e-004 slope

C= -1.7458e-002 -1.7458×10^{-2}

Rho= 0.9999440

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0 D
0.00	0.000	-0.009	-0.009	39	0.000	39					1
0.05	0.050					185	0				1
0.20	0.200	0.201	0.001	1045	0.0 %	1045					1
2.50	2.500	2.533	0.033	12185	0.0 %	12185					1
5.00	5.000	4.939	-0.061	23679	0.0 %	23679					1
7.50	7.500	7.560	0.060	36206	0.0 %	36206					1
10.0	10.000	9.976	-0.024	47749	0.0 %	47749					0

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	39	-		0 SFAM01.1	PPB	8/17/2022 16:06	Std
	0.2 S01	1045	-		0.2 SFAM01.1	PPB	8/17/2022 16:09	Std
	2.5 S02	12185	-		2.5 SFAM01.1	PPB	8/17/2022 16:13	Std
	5 S03	23679	-		5 SFAM01.1	PPB	8/17/2022 16:15	Std
	7.5 S04	36206	-		7.5 SFAM01.1	PPB	8/17/2022 16:17	Std
	10 S05	47749	-		10 SFAM01.1	PPB	8/17/2022 16:19	Std
ICV092	ICV092	18085	3.7677	-	SFAM01.1	PPB	8/17/2022 16:22	SMPL
ICB092	ICB092	-3	-0.0181	-	SFAM01.1	PPB	8/17/2022 16:24	SMPL
CCV061	CCV061	23249	4.8485	-	SFAM01.1	PPB	8/17/2022 16:27	SMPL
CCB061	CCB061	-23	-0.0223	-	SFAM01.1	PPB	8/17/2022 16:29	SMPL
PB147094BL	PBS094	-9	-0.0193	-	SFAM01.1	PPB	8/17/2022 16:31	SMPL
N4006-04	JRMN0	16479	3.4316	-	SFAM01.1	PPB	8/17/2022 16:40	SMPL
N4006-05	JRMN1	39729	8.2977	-	SFAM01.1	PPB	8/17/2022 16:43	SMPL
N4006-06	JRMN3	248817	52.0594	-	SFAM01.1	PPB	8/17/2022 16:45	SMPL
PB147095BL	PBS095	-29	-0.0235	-	SFAM01.1	PPB	8/17/2022 16:52	SMPL
N4007-01	JRMM9	59268	12.3872	-	SFAM01.1	PPB	8/17/2022 16:54	SMPL
N4007-03	JRMM9S	67466	14.103	-	SFAM01.1	PPB	8/17/2022 16:59	SMPL
N4007-04	JRMN2	25527	5.3253	-	SFAM01.1	PPB	8/17/2022 17:01	SMPL
PB147096BL	PBS096	455	0.0778	-	SFAM01.1	PPB	8/17/2022 17:03	SMPL
N4135-01	GBJN0	771	0.1439	-	SFAM01.1	PPB	8/17/2022 17:06	SMPL
N4135-02	GBJN1	1701	0.3386	-	SFAM01.1	PPB	8/17/2022 17:08	SMPL
N4135-03	GBJN2	1747	0.3482	-	SFAM01.1	PPB	8/17/2022 17:10	SMPL
N4135-04	GBJN3	2844	0.5778	-	SFAM01.1	PPB	8/17/2022 17:12	SMPL
N4135-05	GBJN4	1333	0.2615	-	SFAM01.1	PPB	8/17/2022 17:15	SMPL
N4135-06	GBJN5	1281	0.2507	-	SFAM01.1	PPB	8/17/2022 17:17	SMPL
N4135-07	GBJN6	1773	0.3536	-	SFAM01.1	PPB	8/17/2022 17:19	SMPL
CCV062	CCV062	23467	4.8941	-	SFAM01.1	PPB	8/17/2022 17:21	SMPL
CCB062	CCB062	-12	-0.02	-	SFAM01.1	PPB	8/17/2022 17:24	SMPL
N4135-08	GBJN7	386	0.0633	-	SFAM01.1	PPB	8/17/2022 17:26	SMPL
N4135-09	GBJN7D	383	0.0627	-	SFAM01.1	PPB	8/17/2022 17:28	SMPL
N4135-10	GBJN7S	12581	2.6157	-	SFAM01.1	PPB	8/17/2022 17:30	SMPL
PB147097BL	PBS097	-88	-0.0359	-	SFAM01.1	PPB	8/17/2022 17:35	SMPL
N4212-01	BGC46	485	0.0841	-	SFAM01.1	PPB	8/17/2022 17:37	SMPL
N4212-02	BGC46D	464	0.0797	-	SFAM01.1	PPB	8/17/2022 17:42	SMPL
N4212-03	BGC46S	13076	2.7193	-	SFAM01.1	PPB	8/17/2022 17:44	SMPL
N4212-04	BGC47	452	0.0771	-	SFAM01.1	PPB	8/17/2022 17:47	SMPL
N4212-05	BGC48	1141	0.2214	-	SFAM01.1	PPB	8/17/2022 17:49	SMPL
N4212-06	BGC49	581	0.1041	-	SFAM01.1	PPB	8/17/2022 17:51	SMPL
N4212-07	BGC50	420	0.0704	-	SFAM01.1	PPB	8/17/2022 17:54	SMPL
N4212-08	BGC51	178	0.0198	-	SFAM01.1	PPB	8/17/2022 17:56	SMPL
N4212-09	BGC52	210	0.0265	-	SFAM01.1	PPB	8/17/2022 17:58	SMPL
N4212-10	BGC53	975	0.1866	-	SFAM01.1	PPB	8/17/2022 18:00	SMPL
N4135-11	GBJP6	24964	5.2075	-	SFAM01.1	PPB	8/17/2022 18:03	SMPL
CCV063	CCV063	24150	5.0371	-	SFAM01.1	PPB	8/17/2022 18:09	SMPL
CCB063	CCB063	-8	-0.0191	-	SFAM01.1	PPB	8/17/2022 18:12	SMPL
N4006-01	JRMM8	19435	4.0502	-	SFAM01.1	PPB	8/17/2022 18:14	SMPL

LB121568

INSTRUMENT ID : CV1

N4006-02	JRMM8D	19218	4.0048 -	SFAM01.1	PPB	8/17/2022 18:16 SMPL
N4006-03	JRMM8S	33177	6.9264 -	SFAM01.1	PPB	8/17/2022 18:19 SMPL
N4006-06DLX10	JRMN3	28178	5.8801 -	SFAM01.1	PPB	8/17/2022 18:21 SMPL
N4007-02	JRMM9D	60191	12.5804 -	SFAM01.1	PPB	8/17/2022 18:25 SMPL
N4007-01DLX2	JRMM9	30438	6.3532 -	SFAM01.1	PPB	8/17/2022 18:38 SMPL
N4007-02DLX2	JRMM9D	29967	6.2546 -	SFAM01.1	PPB	8/17/2022 18:41 SMPL
N4007-03DLX2	JRMM9S	35697	7.4539 -	SFAM01.1	PPB	8/17/2022 18:43 SMPL
CCV064	CCV064	24282	5.0647 -	SFAM01.1	PPB	8/17/2022 18:45 SMPL
CCB064	CCB064	-8	-0.0191 -	SFAM01.1	PPB	8/17/2022 18:47 SMPL