



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB121576

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	08/18/2022	13:05
2	CAL2	1	Water	NA	NA	20.7	7.00	08/18/2022	13:06
3	CAL3	1	Water	NA	NA	20.7	10.03	08/18/2022	13:07
4	ICV	1	Water	NA	NA	20.8	6.98	08/18/2022	13:10
5	CCV1	1	Water	NA	NA	20.7	2.01	08/18/2022	13:15
6	N4241-07	1	Water	NA	NA	27.4	5.15	08/18/2022	13:20
7	N4241-08	1	Water	NA	NA	25.2	5.20	08/18/2022	13:24
8	N4241-09	1	Water	NA	NA	25.6	5.03	08/18/2022	13:26
9	N4241-10	1	Water	NA	NA	25.2	4.58	08/18/2022	13:30
10	N4247-12	1	Water	NA	NA	24.1	9.10	08/18/2022	13:35
11	N4247-14	1	Water	NA	NA	24.6	5.94	08/18/2022	13:40
12	N4247-15	1	Water	NA	NA	24.2	5.98	08/18/2022	13:44
13	N4247-16	1	Water	NA	NA	24.2	6.00	08/18/2022	13:45
14	N4247-16DUP	1	Water	NA	NA	24.3	6.01	08/18/2022	13:46
15	CCV2	1	Water	NA	NA	20.7	12.03	08/18/2022	13:50

WORKLIST(Hardcopy Internal Chain)

WB 121576

WorkList Name : ph n4247 w

WorkList ID : 162418

Department : Wet-Chemistry

Date : 08-18-2022 12:54:11

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/17/2022	Water	N4241-07	pH	Cool 4 deg C	PSEG03	J11	TRE-1630	08/17/2022	9040C
08/17/2022	Water	N4241-08	pH	Cool 4 deg C	PSEG03	J11	TRE-1631	08/17/2022	9040C
08/17/2022	Water	N4241-09	pH	Cool 4 deg C	PSEG03	J11	TRE-1632	08/17/2022	9040C
08/17/2022	Water	N4241-10	pH	Cool 4 deg C	PSEG03	J11	TRE-1633	08/17/2022	9040C
08/18/2022	Water	N4247-12	pH	Cool 4 deg C	PSEG03	K11	34088	08/18/2022	9040C
08/18/2022	Water	N4247-14	pH	Cool 4 deg C	PSEG03	K11	DL-1	08/18/2022	9040C
08/18/2022	Water	N4247-15	pH	Cool 4 deg C	PSEG03	K11	DL-2	08/18/2022	9040C
08/18/2022	Water	N4247-16	pH	Cool 4 deg C	PSEG03	K11	DL-3	08/18/2022	9040C

Date/Time 08/18/22 13:00
 Raw Sample Received by: J.C. Saml
 Raw Sample Relinquished by: J.C. Saml

Date/Time 08/18/22 14:20
 Raw Sample Received by: J.C. Saml
 Raw Sample Relinquished by: J.C. Saml