

LB121582


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

45453



A= 0.0000e+000

B= 2.2014e-004

C= 9.6926e-004

Rho= 0.9999948

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Y/D
0.0	0.000	-0.003	-0.003	-20	0.000	-20					
0.2	0.200	0.200	0.000	906	0.0 %	906					
2.5	2.500	2.499	-0.001	11348	0.0 %	11348					
5.0	5.000	5.018	0.018	22791	0.0 %	22791					
7.5	7.500	7.479	-0.021	33968	0.0 %	33968					
10.0	10.000	10.007	0.007	45453	0.0 %	45453					

4001
000001

LB121582 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 18 Aug 2022 15:53:11

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-20	-	0.0000	7470A	PPB	18 Aug 2022 15:54:49	S	Std
0.2 - 1	50.2	906	-	0.2000	7470A	PPB	18 Aug 2022 15:57:06	S	Std
2.5 - 1	52.5	11348	-	2.5000	7470A	PPB	18 Aug 2022 15:59:23	S	Std
5.0 - 1	55	22791	-	5.0000	7470A	PPB	18 Aug 2022 16:01:40	S	Std
7.5 - 1	57.5	33968	-	7.5000	7470A	PPB	18 Aug 2022 16:03:57	S	Std
10.0 - 1	511	45453	-	10.0000	7470A	PPB	18 Aug 2022 16:06:13	S	Std
ICV94 - 1	ICV94	18812	4.1422	-	7470A	PPB	18 Aug 2022 16:11:50	U	SMPL
ICB94 - 1	ICB94	-101	-0.0213	-	7470A	PPB	18 Aug 2022 16:14:04	U	SMPL
CCV67 - 1	CCV67	24529	5.4008	-	7470A	PPB	18 Aug 2022 16:21:44	U	SMPL
CCB67 - 1	CCB67	-130	-0.0276	-	7470A	PPB	18 Aug 2022 16:23:58	U	SMPL
CRA - 1	CRA	885	0.1958	-	7470A	PPB	18 Aug 2022 16:26:14	U	SMPL
HighStd - 1	HighStd	45714	10.0645	-	7470A	PPB	18 Aug 2022 16:28:29	U	SMPL
ChkStd - 1	ChkStd	32505	7.1566	-	7470A	PPB	18 Aug 2022 16:30:45	U	SMPL
PB147111BL - 1	PBW	-156	-0.0334	-	7470A	PPB	18 Aug 2022 16:33:03	U	SMPL
PB147111BS - 1	LCSW	16509	3.6353	-	7470A	PPB	18 Aug 2022 16:35:20	U	SMPL
N4224-02 - 1	TP1-0816	-101	-0.0213	-	7470A	PPB	18 Aug 2022 16:37:37	U	SMPL
N4224-05 - 1	TP2-0816	-25	-0.0045	-	7470A	PPB	18 Aug 2022 16:39:54	U	SMPL
N4224-12 - 1	TP3-0816	-18	-0.0030	-	7470A	PPB	18 Aug 2022 16:42:11	U	SMPL
N4227-02 - 1	SP-1	-17	-0.0028	-	7470A	PPB	18 Aug 2022 16:44:28	U	SMPL
N4227-12 - 1	SP-2	-7	-0.0006	-	7470A	PPB	18 Aug 2022 16:46:44	U	SMPL
CCV68 - 1	CCV68	20688	4.5552	-	7470A	PPB	18 Aug 2022 16:48:59	U	SMPL
CCB68 - 1	CCB68	-130	-0.0276	-	7470A	PPB	18 Aug 2022 16:51:14	U	SMPL
N4227-22 - 1	SP-3	-9	-0.0010	-	7470A	PPB	18 Aug 2022 16:53:30	U	SMPL
N4227-22DUP - 1	SP-3DUP	-10	-0.0012	-	7470A	PPB	18 Aug 2022 16:55:47	U	SMPL
N4227-22MS - 1	SP-3MS	15482	3.4092	-	7470A	PPB	18 Aug 2022 16:58:02	U	SMPL
N4227-22MSD - 1	SP-3MSD	14933	3.2883	-	7470A	PPB	18 Aug 2022 17:00:18	U	SMPL
PB147074TB - 1	PB147074TB	-94	-0.0197	-	7470A	PPB	18 Aug 2022 17:02:35	U	SMPL
N4227-22LX5 - 1		-49	-0.0098	-	7470A	PPB	18 Aug 2022 17:04:51	U	SMPL
N4227-22A - 1		18183	4.0038	-	7470A	PPB	18 Aug 2022 17:07:07	U	SMPL
CCV69 - 1	CCV69	23297	5.1296	-	7470A	PPB	18 Aug 2022 17:09:24	U	SMPL
CCB69 - 1	CCB69	-127	-0.0270	-	7470A	PPB	18 Aug 2022 17:11:41	U	SMPL