

LB121526


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

55224



A= 0.0000e+000

B= 1.8158e-004

C= 3.9298e-002

Rho= 0.9999078

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	1/10
0.0	0.000	0.032	0.032	-38	0.000	-38					
0.2	0.200	0.225	0.025	1021	0.0 %	1021					
2.5	2.500	2.496	-0.004	13531	0.0 %	13531					
5.0	5.000	4.911	-0.089	26830	0.0 %	26830					
7.5	7.500	7.468	-0.032	40912	0.0 %	40912					
10.0	10.000	10.067	0.067	55224	0.0 %	55224					

1/10
13
0
2
0
1

LB121586 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 19 Aug 2022 07:46:15

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-38	-	0.0000	7471B	PPB	19 Aug 2022 08:00:22	S	Std
0.2 - 1	50.2	1021	-	0.2000	7471B	PPB	19 Aug 2022 08:02:39	S	Std
2.5 - 1	52.5	13531	-	2.5000	7471B	PPB	19 Aug 2022 08:04:56	S	Std
5.0 - 1	55	26830	-	5.0000	7471B	PPB	19 Aug 2022 08:07:13	S	Std
7.5 - 1	57.5	40912	-	7.5000	7471B	PPB	19 Aug 2022 08:09:30	S	Std
10.0 - 1	511	55224	-	10.0000	7471B	PPB	19 Aug 2022 08:11:47	S	Std
ICV95 - 1	ICV95	21431	3.9308	-	7471B	PPB	19 Aug 2022 08:14:41	U	SMPL
ICB95 - 1	ICB95	-158	0.0106	-	7471B	PPB	19 Aug 2022 08:16:55	U	SMPL
CCV70 - 1	CCV70	27651	5.0603	-	7471B	PPB	19 Aug 2022 08:19:12	U	SMPL
CCB70 - 1	CCB70	-166	0.0092	-	7471B	PPB	19 Aug 2022 08:21:27	U	SMPL
CRA - 1	CRA	999	0.2207	-	7471B	PPB	19 Aug 2022 08:23:43	U	SMPL
HighStd - 1	HighStd	55535	10.1236	-	7471B	PPB	19 Aug 2022 08:25:59	U	SMPL
ChkStd - 1	ChkStd	38701	7.0668	-	7471B	PPB	19 Aug 2022 08:28:14	U	SMPL
PB147124BL - 1	PBS	-191	0.0046	-	7471B	PPB	19 Aug 2022 08:30:32	U	SMPL
PB147124BS - 1	LCSS	20237	3.7140	-	7471B	PPB	19 Aug 2022 08:32:49	U	SMPL
N4224-01 - 1	TP1-0816	3467	0.6689	-	7471B	PPB	19 Aug 2022 08:35:06	U	SMPL
N4224-04 - 1	TP2-0816	1266	0.2692	-	7471B	PPB	19 Aug 2022 08:37:23	U	SMPL
N4224-11 - 1	TP3-0816	1783	0.3631	-	7471B	PPB	19 Aug 2022 08:39:40	U	SMPL
N4225-01 - 1	R-70-1	1903	0.3849	-	7471B	PPB	19 Aug 2022 08:41:57	U	SMPL
N4226-01 - 1	B-1	12631	2.3329	-	7471B	PPB	19 Aug 2022 08:44:13	U	SMPL
CCV71 - 1	CCV71	27835	5.0937	-	7471B	PPB	19 Aug 2022 08:46:28	U	SMPL
CCB71 - 1	CCB71	-130	0.0157	-	7471B	PPB	19 Aug 2022 08:48:44	U	SMPL
N4227-01 - 1	SP-1	1126	0.2438	-	7471B	PPB	19 Aug 2022 08:51:01	U	SMPL
N4227-11 - 1	SP-2	1033	0.2269	-	7471B	PPB	19 Aug 2022 08:53:17	U	SMPL
N4227-11DUP - 1	SP-2DUP	1127	0.2439	-	7471B	PPB	19 Aug 2022 08:55:32	U	SMPL
N4227-11MS - 1	SP-2MS	21064	3.8642	-	7471B	PPB	19 Aug 2022 08:57:48	U	SMPL
N4227-11MSD - 1	SP-2MSD	21591	3.9599	-	7471B	PPB	19 Aug 2022 09:00:04	U	SMPL
N4227-21 - 1	SP-3	1240	0.2645	-	7471B	PPB	19 Aug 2022 09:02:22	U	SMPL
N4228-01 - 1	OK-03-081622	1954	0.3941	-	7471B	PPB	19 Aug 2022 09:04:40	U	SMPL
N4232-01 - 1	BU-01-081722	1371	0.2882	-	7471B	PPB	19 Aug 2022 09:06:57	U	SMPL
N4240-01 - 1	TR-06-081722	1476	0.3073	-	7471B	PPB	19 Aug 2022 09:09:14	U	SMPL
N4242-01 - 1	CL-02-081722	2558	0.5038	-	7471B	PPB	19 Aug 2022 09:11:30	U	SMPL
CCV72 - 1	CCV72	27560	5.0438	-	7471B	PPB	19 Aug 2022 09:13:45	U	SMPL
CCB72 - 1	CCB72	-116	0.0182	-	7471B	PPB	19 Aug 2022 09:16:01	U	SMPL
N4243-01 - 1	B-4	2222	0.4428	-	7471B	PPB	19 Aug 2022 09:18:18	U	SMPL
N4227-11LLX5 - 1		97	0.0569	-	7471B	PPB	19 Aug 2022 09:20:34	U	SMPL
N4227-11A - 1		23576	4.3203	-	7471B	PPB	19 Aug 2022 09:22:50	U	SMPL
CCV73 - 1	CCV73	25327	4.6383	-	7471B	PPB	19 Aug 2022 09:25:06	U	SMPL
CCB73 - 1	CCB73	-143	0.0133	-	7471B	PPB	19 Aug 2022 09:27:22	U	SMPL