



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB121587

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	08/18/2022	17:15
2	CAL2	1	Water	NA	NA	20.7	6.98	08/18/2022	17:16
3	CAL3	1	Water	NA	NA	20.7	10.03	08/18/2022	17:20
4	ICV	1	Water	NA	NA	20.8	7.02	08/18/2022	17:22
5	CCV1	1	Water	NA	NA	20.7	2.01	08/18/2022	17:25
6	N4259-04	1	Water	NA	NA	24.9	6.38	08/18/2022	17:26
7	N4259-04DUP	1	Water	NA	NA	24.8	6.39	08/18/2022	17:27
8	N4259-05	1	Water	NA	NA	23.3	6.97	08/18/2022	17:30
9	N4259-06	1	Water	NA	NA	25.7	7.57	08/18/2022	17:32
10	N4259-07	1	Water	NA	NA	24.7	4.52	08/18/2022	17:35
11	N4259-08	1	Water	NA	NA	25.6	6.02	08/18/2022	17:36
12	CCV2	1	Water	NA	NA	20.8	12.03	08/18/2022	17:40

WORKLIST(Hardcopy Internal Chain)

VB 121587

WorkList Name : CORROSIONITY N4261 V WorkList ID : 162428 Department : Wet-Chemistry Date : 08-18-2022 17:05:02

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/18/2022	Water	N4259-04	Corrosivity	Cool 4 deg C	PSEG03	K11	8221	08/18/2022	9040C
08/18/2022	Water	N4259-05	Corrosivity	Cool 4 deg C	PSEG03	K11	8188	08/18/2022	9040C
08/18/2022	Water	N4259-06	Corrosivity	Cool 4 deg C	PSEG03	K11	8217	08/18/2022	9040C
08/18/2022	Water	N4259-07	Corrosivity	Cool 4 deg C	PSEG03	K11	8223	08/18/2022	9040C
08/18/2022	Water	N4259-08	Corrosivity	Cool 4 deg C	PSEG03	K11	8184	08/18/2022	9040C

Date/Time 08.15.22 14:10
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 08.15.22 18:20
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]