



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB121588

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	08/18/2022	17:15
2	CAL2	1	Water	NA	NA	20.7	6.98	08/18/2022	17:16
3	CAL3	1	Water	NA	NA	20.7	10.03	08/18/2022	17:20
4	ICV	1	Water	NA	NA	20.8	7.02	08/18/2022	17:22
5	CCV1	1	Water	NA	NA	20.7	2.01	08/18/2022	17:25
6	N4255-01	1	Solid	20.01	20	21.8	6.05	08/18/2022	17:45
7	N4259-02	1	Solid	20.03	20	22.3	5.56	08/18/2022	17:46
8	N4259-03	1	Solid	20.04	20	21.7	6.87	08/18/2022	17:49
9	N4261-02	1	Solid	20.02	20	20.9	8.51	08/18/2022	17:50
10	N4261-02DUP	1	Solid	20.03	20	20.4	8.52	08/18/2022	17:51
11	CCV2	1	Water	NA	NA	20.8	12.03	08/18/2022	18:00

WORKLIST(Hardcopy Internal Chain)

VB121588

WorkList Name : CORROSIONITY N4257 WorkList ID : 162427 Department : Wet-Chemistry Date : 08-18-2022 17:04:09

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/18/2022	Solid	N4255-01	Corrosivity	Cool 4 deg C	PSEG03	K11	POLYLINER-COMP	08/18/2022	9045D
08/18/2022	Solid	N4259-02	Corrosivity	Cool 4 deg C	PSEG03	K11	8220	08/18/2022	9045D
08/18/2022	Solid	N4259-03	Corrosivity	Cool 4 deg C	PSEG03	K11	8186-8185-8216-8215	08/18/2022	9045D
08/17/2022	Solid	N4261-02	Corrosivity	Cool 4 deg C	PSEG03	K11	C5	08/17/2022	9045D

Date/Time 08-18-22 17:10
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 08-18-22 18:20
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]