

LB121610

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12 Instrument ID : Konelab

8/19/2022 14:44

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.001	0.0	0.113	
ICB1	0.002	0.0	0.009	
CCV1	0.999	0.0	0.113	
CCB1	-0.000	0.0	0.008	
PB147148BL	0.001	0.0	0.009	
PB147148BS	1.002	0.0	0.113	
N3980-12	0.028	0.0	0.011	
██████████	██████████	██████████	0.011	
N4264-02	-0.003	0.0	0.008	
N4264-02DUP	-0.002	0.0	0.008	
N4264-02MS	1.026	0.0	0.116	
N4264-02MSD	1.020	0.0	0.115	
CCV2	1.004	0.0	0.113	
CCB2	-0.000	0.0	0.008	

N 14
Mean 0.436
SD 0.5148
CV% 118.07

Aquakem v. 7.2AQ1

Results from time period:

Fri Aug 19 13:54:26 2022

Fri Aug 19 14:44:09 2022

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.00PPM	A	Phenolics-	P	0.001	mg/l	8/19/2022 13:54:26	
0.05PPM	A	Phenolics-	P	0.052	mg/l	8/19/2022 13:54:27	
0.1PPM	A	Phenolics-	P	0.1052	mg/l	8/19/2022 13:54:28	
0.25PPM	A	Phenolics-	P	0.2537	mg/l	8/19/2022 13:54:29	
0.50PPM	A	Phenolics-	P	0.4978	mg/l	8/19/2022 13:54:30	
1.0PPM	A	Phenolics-	P	0.9811	mg/l	8/19/2022 13:54:31	
2.0PPM	A	Phenolics-	P	2.0092	mg/l	8/19/2022 13:54:32	
ICV1	S	Phenolics-	P	1.0008	mg/l	8/19/2022 14:23:16	
ICB1	S	Phenolics-	P	0.0016	mg/l	8/19/2022 14:23:19	
CCV1	S	Phenolics-	P	0.9994	mg/l	8/19/2022 14:23:21	
CCB1	S	Phenolics-	P	-0.0004	mg/l	8/19/2022 14:23:22	
PB147148BL	S	Phenolics-	P	0.0013	mg/l	8/19/2022 14:23:25	
PB147148BS	S	Phenolics-	P	1.0021	mg/l	8/19/2022 14:23:27	
N3980-12	S	Phenolics-	P	0.0283	mg/l	8/19/2022 14:30:30	
N4264-02	S	Phenolics-	P	-0.0033	mg/l	8/19/2022 14:30:32	
N4264-02DUP	S	Phenolics-	P	-0.002	mg/l	8/19/2022 14:30:33	
N4264-02MS	S	Phenolics-	P	1.0264	mg/l	8/19/2022 14:30:34	
N4264-02MSD	S	Phenolics-	P	1.02	mg/l	8/19/2022 14:30:35	
CCV2	S	Phenolics-	P	1.0038	mg/l	8/19/2022 14:30:38	
CCB2	S	Phenolics-	P	-0.0003	mg/l	8/19/2022 14:44:09	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12 Instrument ID : Konelab

8/19/2022 13:55


Test Phenolics-

Accepted 8/19/2022 13:55

Factor 9.577
Bias 0.008

Coeff. of det. 0.999844

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.00PPM	0.009	0.0010	0.0000	
2	Phenol-2	0.014	0.0520	0.0500	
3	Phenol-2	0.019	0.1052	0.1000	
4	Phenol-2	0.035	0.2537	0.2500	
5	Phenol-2	0.060	0.4978	0.5000	
6	Phenol-2	0.111	0.9811	1.0000	
7	Phenol-2	0.218	2.0092	2.0000	

De

Errors

1
4.00
5.2
1.5
10.5
11.8
0.8

12
8/18/22