



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: Non-Polar Material
Run Number: LB121624
Analysis Date: 08/22/2022
BalanceID: WC SC-6
OvenID: EXT OVEN-3

ANALYST: jignesh
REVIEWED BY: Iwona
Extraction Date: 08/22/2022
Extraction IN Time: 11:00
Extraction OUT Time: 11:35
Thermometer ID: EXT OVEN#3

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB121624BL	LB121624BL	WATER	1.3	1000	100	2.9843	2.9843	3.02	2.9844	2.9844	0.0001	0.1
2	LB121624BS	LB121624BS	WATER	1.3	1000	100	2.7961	2.7961	3.01	2.8131	2.8131	0.0170	17
3	LB121624BSD	LB121624BSD	WATER	1.3	1000	100	2.9931	2.9931	3.03	3.0104	3.0104	0.0173	17.3
4	N4264-02	U-3	WATER	1.3	1000	100	2.9854	2.9854	3.02	2.9861	2.9861	0.0007	0.7
5	N4276-01	DSN-001-SEWER-OUTLET-0	WATER	1.6	1000	100	2.9819	2.9819	3.03	2.9829	2.9829	0.0010	1
6	N4276-02	DSN-002-SEWER-OUTLET-0	WATER	1.6	1000	100	3.0277	3.0277	3.04	3.0284	3.0284	0.0007	0.7
7	N4277-02	B-1	WATER	1.6	1000	100	3.0507	3.0507	3.04	3.0513	3.0513	0.0006	0.6
8	N4277-03	U-2	WATER	1.6	1000	100	2.9947	2.9947	3.05	2.9954	2.9954	0.0007	0.7

QC Batch# LB121624Test: Non-Polar MaterialAnalysis Date: 08/22/2022**Chemicals Used:**

Chemical Name	Chemical Lot #
HEXANE	W2938
pH Paper 0-14	M4909
Sodium Sulfate	EP2254
1:1 HCL	WP98461
Silica Gel	W2839
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	5.00 ML	WP98457
LCSWD	5.00 ML	WP98458
MS/MSD	NA	NA

BALANCE CALIBRATION / OVEN Dessicator Data**Analytical Balance ID # : WC SC-6****Before Analysis**

0.0020 gram Balance: 0.0021 (0.0018-0.0022) In OVEN TEMP1 : 70 °C Dessicator Time In1 : 12:46
1.0000 gram Balance: 1.0005 (0.9950-1.0050) In Time1: 12:10
Bal Check Time: 11:10 Out OVEN TEMP1: 70 °C Dessicator Time Out1: 13:15
Out Time1: 12:45

After Analysis

0.0020 gram Balance: 0.0019 (0.0018-0.0022) In OVEN TEMP2 : 71 °C Dessicator Time In2 : 14:31
1.0000 gram Balance: 1.0004 (0.9950-1.0050) In Time2: 14:00
Bal Check Time: 15:20 Out OVEN TEMP2: 71 °C Dessicator Time Out2: 15:10
Out Time2: 14:30

121623

WORKLIST(Hardcopy Internal Chain)

WorkList Name : NON POPLAR N4277 WorkList ID : 162488 Department : Wet-Chemistry Date : 08-22-2022 10:43:43

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method	
09/01/2022	Water	N4264-02	Non-Polar Material	L	Conc H2SO4 to	TECT01	K11	U-3	08/18/2022	1664A
09/02/2022	Water	N4276-01	Non-Polar Material	OK	Conc H2SO4 to	TRIS02	K11	DSN-001-SEWER-OUTLET-001	08/19/2022	1664A
09/02/2022	Water	N4276-02	Non-Polar Material	OK	Conc H2SO4 to	TRIS02	K11	DSN-002-SEWER-OUTLET-002	08/19/2022	1664A
09/02/2022	Water	N4277-02	Non-Polar Material	M	Conc H2SO4 to	TECT01	K11	B-1	08/19/2022	1664A
09/02/2022	Water	N4277-03	Non-Polar Material	M	Conc H2SO4 to	TECT01	K11	U-2	08/19/2022	1664A

Date/Time 08/22/22 10:50
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 08/22/22 13:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]